


Data Path : Z:\HPCHEM1\BNA_G\Data\BG033117\
 Data File : BG026518.D
 Acq On : 31 Mar 2017 15:45
 Operator : SJ/MA
 Sample : DFTPP23
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP23

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG032717MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Apr 03 01:47:03 2017



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.0	41602	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	24.4	46129	PASS
70	69	0.00	2	0.6	275	PASS
127	198	10	80	40.3	76152	PASS
197	198	0.00	2	0.3	484	PASS
198	198	100	100	100.0	188821	PASS
199	198	5	9	6.8	12826	PASS
275	198	10	60	26.5	50024	PASS
365	198	1	100	2.9	5512	PASS
441	442	0.01	100	15.5	24168	PASS
442	198	50	100	82.4	155640	PASS
443	442	15	24	19.9	30986	PASS

Scan 2499 (17.771 min): BG026518.D
DFTPP23

m/z	Abundance
44.10	929.0
134.20	161.0
148.20	161.0
207.10	693.0
460.80	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

m/z	Abundance
37.00	197.0
38.10	469.0
39.10	2643.0
40.10	254.0
44.10	1004.0
49.00	304.0
50.10	9380.0
51.10	39464.0
52.10	2531.0
53.10	232.0
55.10	225.0
56.00	1342.0
57.10	2436.0
58.10	319.0
61.00	502.0
62.20	557.0
63.10	1584.0
64.00	232.0
65.00	874.0
66.00	311.0
69.10	42056.0
70.10	252.0
73.10	552.0
74.10	4426.0
75.10	7320.0
76.10	2934.0
77.10	54432.0
78.10	3861.0
79.10	2830.0
80.10	2412.0
81.10	3252.0
82.10	885.0
83.00	958.0
84.20	240.0
85.10	598.0
86.00	998.0
87.10	888.0
88.00	301.0
91.10	724.0
92.00	1007.0
93.00	4938.0
94.00	432.0
98.10	5020.0
99.10	5005.0
100.10	641.0
101.10	2152.0
103.10	751.0
104.10	1895.0
105.10	1348.0
107.10	17648.0
108.10	2218.0
110.10	33400.0
111.10	5714.0
112.10	497.0
113.10	265.0
116.00	828.0
117.10	11332.0
118.00	1219.0
120.00	466.0
122.10	1124.0
123.00	2142.0
124.00	845.0
125.00	789.0
127.10	67112.0
128.10	5347.0
129.10	23904.0
130.00	2236.0
131.00	360.0
132.10	414.0
134.00	688.0
135.00	2719.0
136.10	841.0
137.00	932.0
138.00	225.0
139.10	168.0
140.00	412.0
141.00	2847.0
142.10	1148.0
143.10	819.0
144.10	278.0
145.00	227.0
146.10	626.0
147.00	1766.0
148.00	3110.0
149.00	877.0
150.90	218.0
151.50	178.0
152.10	171.0
153.00	796.0
154.00	824.0
155.10	1543.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

156.10	2951.0
157.20	632.0
158.00	811.0
159.00	440.0
160.00	1006.0
161.10	1557.0
162.10	558.0
163.90	287.0
165.10	1209.0
166.10	1090.0
167.10	6405.0
168.10	5814.0
169.00	831.0
170.00	463.0
171.10	365.0
171.90	441.0
173.00	650.0
174.00	1476.0
175.10	2254.0
176.10	935.0
177.10	1118.0
178.00	392.0
179.00	4350.0
180.10	3507.0
181.10	1911.0
182.00	282.0
184.10	496.0
185.00	2575.0
186.10	19464.0
187.10	6335.0
188.10	924.0
189.10	1005.0
190.10	289.0
191.10	365.0
192.00	1615.0
193.00	2064.0
194.10	446.0
196.00	4963.0
196.80	1454.0
198.00	150528.0
199.00	10906.0
200.00	946.0
200.90	239.0
201.50	908.0
203.00	934.0
204.10	4751.0
205.10	7545.0
206.10	35856.0
207.10	5274.0
208.10	1042.0
209.10	679.0
211.00	1277.0
212.30	272.0
213.10	162.0
214.90	207.0
216.10	681.0
217.00	8734.0
218.00	1347.0
221.10	6740.0
222.10	1337.0
223.00	2086.0
224.10	19408.0
225.10	5366.0
226.10	732.0
227.10	7372.0
228.10	1446.0
229.00	1626.0
231.10	831.0
231.90	219.0
234.00	547.0
235.10	615.0
236.00	290.0
237.00	754.0
238.20	267.0
240.00	255.0
241.10	336.0
242.10	937.0
243.10	894.0
244.10	15324.0
245.10	2556.0
246.00	2710.0
247.00	603.0
248.10	150.0
248.90	570.0
249.90	248.0
251.10	312.0
252.20	227.0
252.80	227.0
254.10	1019.0
255.10	80176.0
256.10	12574.0
257.00	1284.0
258.10	4118.0
259.10	673.0
261.10	263.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

264.10	166.0
265.10	1440.0
266.00	344.0
271.90	217.0
273.10	2466.0
274.10	6036.0
275.10	39320.0
276.10	5003.0
277.10	2621.0
278.10	573.0
279.10	180.0
281.00	174.0
283.10	444.0
284.20	178.0
285.10	668.0
290.00	175.0
292.00	261.0
293.10	801.0
294.10	307.0
294.80	185.0
296.00	7891.0
297.10	1464.0
302.00	292.0
303.20	937.0
304.10	335.0
314.00	598.0
315.00	1025.0
316.10	751.0
321.00	408.0
323.10	3070.0
324.10	525.0
326.00	227.0
327.00	584.0
328.00	354.0
332.10	188.0
333.10	222.0
334.10	1965.0
335.00	1085.0
341.10	450.0
342.00	194.0
345.90	457.0
346.90	298.0
352.10	596.0
353.10	504.0
354.20	742.0
355.10	290.0
365.00	3621.0
366.00	466.0
371.00	168.0
372.10	1500.0
373.20	332.0
383.10	409.0
390.10	174.0
391.20	198.0
402.10	714.0
403.00	523.0
404.00	325.0
421.00	825.0
422.00	851.0
423.10	4723.0
424.10	2259.0
425.20	331.0
441.10	14202.0
442.10	94440.0
443.10	18408.0
444.10	1790.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

m/z	Abundance
38.20	533.0
39.10	3739.0
40.20	245.0
41.10	153.0
44.10	914.0
45.10	157.0
49.10	489.0
50.10	12393.0
51.10	49752.0
52.10	2648.0
53.10	181.0
55.10	182.0
56.10	2086.0
57.10	3808.0
59.00	157.0
61.10	873.0
62.10	1167.0
63.10	2377.0
64.20	322.0
65.10	1319.0
69.00	57696.0
70.10	321.0
73.00	455.0
74.10	6723.0
75.10	9123.0
76.20	3013.0
77.10	73872.0
78.10	4372.0
79.10	4861.0
80.00	3167.0
81.10	4857.0
82.10	1317.0
83.10	1111.0
84.00	309.0
85.10	1081.0
86.10	1458.0
87.00	654.0
88.00	155.0
91.10	1319.0
92.10	1351.0
93.00	7791.0
94.00	414.0
96.20	342.0
97.10	224.0
98.10	5062.0
99.00	5488.0
100.00	580.0
101.00	3044.0
103.10	817.0
104.10	1712.0
105.10	2030.0
107.10	21552.0
108.10	3173.0
110.10	43560.0
111.00	5878.0
112.10	1058.0
112.90	171.0
113.20	169.0
114.10	254.0
116.10	1280.0
117.10	15289.0
118.00	1289.0
119.10	238.0
120.10	421.0
122.10	1320.0
123.00	1956.0
124.00	1056.0
125.10	1070.0
127.10	89976.0
128.10	6774.0
129.00	32736.0
130.00	2883.0
131.10	590.0
132.10	360.0
133.10	313.0
134.00	1444.0
135.00	2323.0
136.00	966.0
137.10	1480.0
138.00	441.0
139.10	185.0
140.00	558.0
141.00	4087.0
142.10	1361.0
143.00	786.0
144.00	379.0
145.10	212.0
146.00	610.0
147.10	2690.0
148.10	4279.0
149.10	1520.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

150.10	525.0
151.60	723.0
153.00	1248.0
154.10	1117.0
155.00	2471.0
156.10	3271.0
157.00	1081.0
158.00	745.0
159.10	648.0
160.00	1582.0
161.10	1846.0
162.00	579.0
163.10	264.0
164.10	652.0
165.00	1797.0
166.00	1412.0
167.10	8655.0
168.10	5329.0
169.00	1201.0
170.00	270.0
170.60	181.0
171.10	347.0
172.00	866.0
173.00	1100.0
174.10	2321.0
175.10	3991.0
176.10	1463.0
177.00	2292.0
178.00	419.0
179.00	6423.0
180.10	5161.0
181.00	2556.0
182.00	498.0
183.10	238.0
184.10	619.0
185.00	3086.0
186.10	26744.0
187.10	8552.0
188.10	698.0
189.00	1498.0
190.10	305.0
191.10	927.0
192.10	2057.0
193.00	2809.0
194.00	559.0
194.80	388.0
195.10	421.0
196.10	5915.0
198.00	224448.0
199.00	14597.0
200.00	1007.0
201.50	898.0
203.00	1170.0
204.10	8412.0
205.10	13190.0
206.10	54248.0
207.10	7854.0
208.10	1656.0
209.00	635.0
210.00	939.0
211.10	1894.0
215.00	724.0
216.10	1140.0
217.00	13020.0
218.00	1442.0
221.10	9865.0
222.00	1658.0
223.10	2711.0
224.10	31616.0
225.10	7640.0
226.10	1279.0
227.10	11535.0
228.10	1555.0
229.00	2881.0
230.00	369.0
231.10	1269.0
232.10	197.0
232.90	312.0
234.10	669.0
235.10	749.0
236.10	925.0
237.00	1513.0
239.20	387.0
240.00	478.0
241.10	762.0
242.10	1482.0
243.20	1935.0
244.10	26584.0
245.10	3602.0
246.10	3982.0
247.00	1040.0
248.00	261.0
249.10	989.0
249.80	156.0
251.00	342.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

252.10	420.0
253.00	885.0
255.10	126168.0
256.10	17920.0
257.00	1810.0
258.00	6617.0
259.10	943.0
260.10	386.0
261.10	190.0
264.20	288.0
265.00	2912.0
266.00	581.0
269.00	215.0
270.10	154.0
271.20	300.0
272.00	340.0
273.00	3898.0
274.10	9502.0
275.10	57464.0
276.10	7783.0
277.00	4335.0
278.10	796.0
282.10	202.0
283.20	548.0
284.00	285.0
285.10	907.0
289.00	316.0
290.00	204.0
291.10	211.0
291.90	166.0
293.00	1068.0
294.00	351.0
295.00	380.0
296.00	14317.0
297.10	2091.0
298.00	265.0
302.00	175.0
303.10	1821.0
304.10	628.0
308.10	308.0
309.00	254.0
310.10	286.0
311.00	160.0
312.00	186.0
314.00	386.0
315.10	1244.0
316.10	984.0
317.10	181.0
321.10	481.0
322.10	509.0
323.10	6161.0
324.10	953.0
326.10	176.0
326.90	792.0
328.10	306.0
329.00	209.0
332.00	397.0
333.10	431.0
334.10	3817.0
335.10	1003.0
340.20	154.0
341.20	604.0
342.10	275.0
346.00	1111.0
352.10	1406.0
353.10	1090.0
354.10	1689.0
355.00	432.0
359.00	177.0
365.00	6461.0
365.90	1230.0
370.10	169.0
371.10	361.0
372.10	2602.0
373.10	414.0
373.90	196.0
383.00	799.0
384.10	196.0
390.10	406.0
391.00	234.0
401.00	262.0
402.10	955.0
403.10	2056.0
404.00	442.0
421.00	1291.0
422.10	1257.0
423.10	9582.0
424.10	2837.0
425.10	510.0
438.00	178.0
441.10	28600.0
442.10	180800.0
443.10	37040.0
444.10	3395.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

m/z	Abundance
37.10	214.0
38.10	327.0
39.20	1753.0
44.10	888.0
49.10	328.0
50.10	8747.0
51.10	35592.0
52.10	2044.0
55.00	222.0
56.10	1704.0
57.00	2327.0
58.00	257.0
61.10	683.0
62.10	730.0
63.10	1585.0
64.10	352.0
65.10	780.0
66.00	208.0
69.10	39272.0
70.00	254.0
73.00	283.0
74.10	4165.0
75.10	7465.0
76.10	2771.0
77.10	55496.0
78.10	3682.0
79.00	3385.0
80.10	2603.0
81.10	3612.0
82.10	1195.0
83.00	818.0
84.00	188.0
85.10	898.0
86.00	1085.0
87.10	594.0
88.10	416.0
91.10	688.0
92.00	813.0
93.10	6030.0
94.10	506.0
96.00	577.0
98.10	4346.0
99.10	3991.0
100.10	305.0
101.00	2100.0
103.00	553.0
104.00	1637.0
105.00	1631.0
106.10	502.0
107.10	18104.0
108.10	3463.0
110.10	32696.0
111.00	5196.0
112.10	538.0
113.10	211.0
116.10	1571.0
117.10	12098.0
118.10	1128.0
120.20	218.0
121.00	188.0
122.10	982.0
123.10	1721.0
124.00	788.0
125.00	1012.0
127.10	71368.0
128.10	4521.0
129.10	23152.0
130.10	2275.0
131.00	291.0
132.90	176.0
134.10	573.0
135.10	2022.0
136.00	850.0
137.10	1175.0
138.10	387.0
139.10	331.0
140.10	408.0
141.00	3296.0
142.00	1080.0
143.00	907.0
144.00	371.0
146.00	977.0
147.00	1755.0
148.00	3522.0
149.00	655.0
150.10	223.0
151.20	409.0
152.10	284.0
153.00	900.0
154.10	998.0
155.10	2606.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

156.00	3494.0
157.20	382.0
158.00	805.0
159.10	806.0
160.10	1252.0
161.00	1503.0
162.10	410.0
163.00	180.0
164.10	169.0
165.00	1337.0
166.00	1567.0
167.10	6903.0
168.00	4292.0
169.10	948.0
170.80	402.0
172.00	757.0
173.10	1084.0
174.10	1763.0
175.10	3225.0
176.00	723.0
177.00	1465.0
178.20	777.0
179.00	6314.0
180.10	4466.0
181.10	1816.0
182.10	367.0
184.20	409.0
185.10	2966.0
186.10	22456.0
187.00	6890.0
188.00	568.0
189.00	1350.0
190.00	337.0
191.00	1046.0
192.10	1666.0
193.10	2025.0
194.10	804.0
195.30	571.0
196.10	5025.0
198.00	191488.0
199.00	12977.0
200.00	832.0
201.40	1070.0
203.00	1282.0
204.10	5993.0
205.10	10722.0
206.10	45544.0
207.10	6406.0
208.00	1400.0
209.10	584.0
210.20	828.0
211.10	1816.0
211.80	784.0
215.00	582.0
216.10	903.0
217.00	12253.0
218.10	1633.0
219.00	326.0
221.00	8521.0
223.10	2930.0
224.10	25432.0
225.10	7871.0
226.00	623.0
227.00	10514.0
228.00	1565.0
229.00	2778.0
230.00	234.0
231.10	1066.0
232.00	200.0
233.20	200.0
234.00	448.0
235.00	788.0
236.00	437.0
237.00	1107.0
239.00	426.0
240.00	382.0
241.00	642.0
242.10	1760.0
243.10	1445.0
244.10	21960.0
245.10	2735.0
246.00	3897.0
247.00	722.0
247.80	254.0
249.00	766.0
251.00	157.0
253.00	601.0
255.10	112744.0
256.10	15701.0
257.10	1402.0
258.10	5894.0
259.00	721.0
260.00	301.0
261.00	234.0
263.00	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23

265.00	2179.0
266.10	575.0
270.00	314.0
271.00	154.0
272.10	641.0
273.10	3587.0
274.00	9732.0
275.10	53288.0
276.10	7251.0
277.10	3915.0
278.00	870.0
279.10	219.0
281.30	184.0
283.10	800.0
284.10	364.0
285.10	914.0
286.00	276.0
289.00	231.0
290.00	221.0
291.10	155.0
292.10	394.0
293.10	860.0
294.00	203.0
295.10	414.0
296.10	13316.0
297.10	1497.0
301.00	237.0
302.10	457.0
303.10	1564.0
304.20	354.0
308.10	300.0
313.10	214.0
314.10	775.0
315.00	1437.0
316.10	1229.0
317.00	155.0
321.10	713.0
323.10	4776.0
324.10	943.0
326.20	177.0
327.10	687.0
328.00	452.0
329.00	152.0
332.00	472.0
333.00	530.0
334.10	3754.0
335.10	578.0
340.00	243.0
341.10	685.0
342.00	202.0
346.00	1109.0
350.90	191.0
352.00	1801.0
353.00	1187.0
354.10	1251.0
355.10	226.0
363.90	167.0
365.00	6454.0
366.00	999.0
371.10	297.0
372.10	2462.0
373.10	747.0
383.10	780.0
384.10	220.0
390.10	284.0
391.10	243.0
396.70	168.0
401.10	182.0
402.10	1282.0
403.10	1281.0
404.10	617.0
405.00	272.0
421.00	1164.0
422.00	1732.0
423.10	10135.0
424.10	2192.0
425.00	293.0
435.70	173.0
438.60	179.0
441.10	29704.0
442.10	191680.0
443.10	37512.0
444.10	3516.0
445.00	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP23