


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040119\
 Data File : BG040264.D
 Acq On : 1 Apr 2019 13:39
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG032719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Mar 28 05:47:00 2019



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2501

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.5	39426	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.0	35706	PASS
70	69	0.00	2	1.6	555	PASS
127	198	10	80	44.2	46525	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	105168	PASS
199	198	5	9	7.1	7456	PASS
275	198	10	60	27.5	28904	PASS
365	198	1	100	3.4	3625	PASS
441	198	0.01	100	9.9	10376	PASS
442	198	50	100	60.2	63346	PASS
443	442	15	24	19.1	12121	PASS