

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
 Data File : BG060710.D
 Acq On : 1 Apr 2024 11:08
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:45:57 2024
 Quant Title :
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.492	0.552	-12.2	130	0.00
3	Pyridine	1.485	1.815	-22.2	127	0.00
4	n-Nitrosodimethylamine	0.887	1.017	-14.7	123	0.00
5 S	2-Fluorophenol	1.201	1.284	-6.9	120	0.00
6	Aniline	2.250	2.556	-13.6	125	0.00
7 S	Phenol-d6	1.782	1.980	-11.1	123	0.00
8	2-Chlorophenol	1.361	1.510	-10.9	123	0.00
9	Benzaldehyde	0.983	1.052	-7.0	125	0.00
10 C	Phenol	1.819	2.050	-12.7	123	0.00
11	bis(2-Chloroethyl)ether	1.468	1.655	-12.7	127	0.00
12	1,3-Dichlorobenzene	1.415	1.579	-11.6	127	0.00
13 C	1,4-Dichlorobenzene	1.444	1.609	-11.4	127	0.00
14	1,2-Dichlorobenzene	1.415	1.588	-12.2	128	0.00
15	Benzyl Alcohol	1.444	1.633	-13.1	124	0.00
16	2,2'-oxybis(1-Chloropropane	3.314	3.666	-10.6	125	0.00
17	2-Methylphenol	1.373	1.580	-15.1	128	0.00
18	Hexachloroethane	0.513	0.539	-5.1	123	0.00
19 P	n-Nitroso-di-n-propylamine	1.359	1.534	-12.9	125	0.00
20	3+4-Methylphenols	1.880	2.151	-14.4	126	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
22	Acetophenone	0.551	0.620	-12.5	126	0.00
23 S	Nitrobenzene-d5	0.350	0.388	-10.9	125	0.00
24	Nitrobenzene	0.352	0.421	-19.6	127	0.00
25	Isophorone	0.807	0.920	-14.0	125	0.00
26 C	2-Nitrophenol	0.120	0.115	4.2	102	0.00
27	2,4-Dimethylphenol	0.324	0.352	-8.6	121	0.00
28	bis(2-Chloroethoxy)methane	0.477	0.537	-12.6	126	0.00
29 C	2,4-Dichlorophenol	0.298	0.335	-12.4	120	0.00
30	1,2,4-Trichlorobenzene	0.337	0.377	-11.9	125	0.00
31	Naphthalene	1.082	1.221	-12.8	124	0.00
32	Benzoic acid	0.191	0.184	3.7	102	0.00
33	4-Chloroaniline	0.503	0.567	-12.7	122	0.00
34 C	Hexachlorobutadiene	0.225	0.247	-9.8	124	0.00
35	Caprolactam	0.118	0.129	-9.3	117	0.00
36 C	4-Chloro-3-methylphenol	0.380	0.433	-13.9	121	0.00
37	2-Methylnaphthalene	0.795	0.888	-11.7	124	0.00
38	1-Methylnaphthalene	0.751	0.841	-12.0	127	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.575	0.645	-12.2	127	0.00
41 P	Hexachlorocyclopentadiene	0.292	0.301	-3.1	115	0.00
42 S	2,4,6-Tribromophenol	0.227	0.221	2.6	106	0.00
43 C	2,4,6-Trichlorophenol	0.370	0.405	-9.5	119	0.00
44	2,4,5-Trichlorophenol	0.442	0.496	-12.2	122	0.00
45 S	2-Fluorobiphenyl	1.363	1.516	-11.2	126	0.00
46	1,1'-Biphenyl	1.413	1.575	-11.5	126	0.00
47	2-Chloronaphthalene	1.073	1.194	-11.3	125	0.00
48	2-Nitroaniline	0.307	0.353	-15.0	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
 Data File : BG060710.D
 Acq On : 1 Apr 2024 11:08
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:45:57 2024
 Quant Title :
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	Acenaphthylene	1.803	2.058	-14.1	128	0.00
50	Dimethylphthalate	1.581	1.760	-11.3	123	0.00
51	2,6-Dinitrotoluene	0.247	0.301	-21.9	124	0.00
52 C	Acenaphthene	1.132	1.278	-12.9	126	0.00
53	3-Nitroaniline	0.277	0.350	-26.4#	127	0.00
54 P	2,4-Dinitrophenol	0.110	0.110	0.0	111	0.00
55	Dibenzofuran	1.786	2.009	-12.5	126	0.00
56 P	4-Nitrophenol	0.231	0.267	-15.6	123	0.00
57	2,4-Dinitrotoluene	0.320	0.405	-26.6#	128	0.00
58	Fluorene	1.433	1.600	-11.7	124	0.00
59	2,3,4,6-Tetrachlorophenol	0.385	0.432	-12.2	118	0.00
60	Diethylphthalate	1.590	1.760	-10.7	123	0.00
61	4-Chlorophenyl-phenylether	0.758	0.845	-11.5	125	0.00
62	4-Nitroaniline	0.289	0.359	-24.2	129	0.00
63	Azobenzene	1.550	1.769	-14.1	126	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	0.073	0.074	-1.4	110	0.00
66 c	n-Nitrosodiphenylamine	0.572	0.663	-15.9	125	0.00
67	4-Bromophenyl-phenylether	0.203	0.207	-2.0	110	0.00
68	Hexachlorobenzene	0.229	0.260	-13.5	123	0.00
69	Atrazine	0.200	0.222	-11.0	119	0.00
70 C	Pentachlorophenol	0.150	0.165	-10.0	116	0.00
71	Phenanthrene	1.040	1.173	-12.8	124	0.00
72	Anthracene	1.056	1.206	-14.2	125	0.00
73	Carbazole	0.931	1.050	-12.8	124	0.00
74	Di-n-butylphthalate	1.161	1.275	-9.8	118	0.00
75 C	Fluoranthene	1.262	1.374	-8.9	121	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
77	Benzidine	0.527	0.645	-22.4	114	0.00
78	Pyrene	1.399	1.592	-13.8	122	0.00
79 S	Terphenyl-d14	1.136	1.254	-10.4	119	0.00
80	Butylbenzylphthalate	0.510	0.542	-6.3	110	0.00
81	Benzo(a)anthracene	1.410	1.584	-12.3	119	0.00
82	3,3'-Dichlorobenzidine	0.476	0.537	-12.8	114	0.00
83	Chrysene	1.359	1.526	-12.3	119	0.00
84	Bis(2-ethylhexyl)phthalate	0.813	0.895	-10.1	113	0.00
85 c	Di-n-octyl phthalate	1.325	1.431	-8.0	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	1.408	1.567	-11.3	116	-0.01
88	Benzo(b)fluoranthene	1.149	1.271	-10.6	115	0.00
89	Benzo(k)fluoranthene	1.176	1.350	-14.8	118	0.00
90 C	Benzo(a)pyrene	1.117	1.264	-13.2	116	0.00
91	Dibenzo(a,h)anthracene	1.187	1.336	-12.6	117	-0.01
92	Benzo(g,h,i)perylene	1.155	1.292	-11.9	117	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_G
LabSampleId :
SSTDCCC040