

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
 Data File : BG060710.D
 Acq On : 1 Apr 2024 11:08
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:45:57 2024
 Quant Title :
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	112	0.00
2	1,4-Dioxane	40.000	44.831	-12.1	130	0.00
3	Pyridine	40.000	48.907	-22.3	127	0.00
4	n-Nitrosodimethylamine	40.000	45.831	-14.6	123	0.00
5 S	2-Fluorophenol	80.000	85.575	-7.0	120	0.00
6	Aniline	40.000	45.427	-13.6	125	0.00
7 S	Phenol-d6	80.000	88.890	-11.1	123	0.00
8	2-Chlorophenol	40.000	44.382	-11.0	123	0.00
9	Benzaldehyde	40.000	42.824	-7.1	125	0.00
10 C	Phenol	40.000	45.097	-12.7	123	0.00
11	bis(2-Chloroethyl)ether	40.000	45.096	-12.7	127	0.00
12	1,3-Dichlorobenzene	40.000	44.609	-11.5	127	0.00
13 C	1,4-Dichlorobenzene	40.000	44.568	-11.4	127	0.00
14	1,2-Dichlorobenzene	40.000	44.881	-12.2	128	0.00
15	Benzyl Alcohol	40.000	45.240	-13.1	124	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	44.252	-10.6	125	0.00
17	2-Methylphenol	40.000	46.036	-15.1	128	0.00
18	Hexachloroethane	40.000	42.035	-5.1	123	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	45.141	-12.9	125	0.00
20	3+4-Methylphenols	40.000	45.760	-14.4	126	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	44.961	-12.4	126	0.00
23 S	Nitrobenzene-d5	80.000	88.567	-10.7	125	0.00
24	Nitrobenzene	40.000	43.443	-8.6	127	0.00
25	Isophorone	40.000	45.583	-14.0	125	0.00
26 C	2-Nitrophenol	40.000	35.769	10.6	102	0.00
27	2,4-Dimethylphenol	40.000	43.419	-8.5	121	0.00
28	bis(2-Chloroethoxy)methane	40.000	45.027	-12.6	126	0.00
29 C	2,4-Dichlorophenol	40.000	44.936	-12.3	120	0.00
30	1,2,4-Trichlorobenzene	40.000	44.678	-11.7	125	0.00
31	Naphthalene	40.000	45.140	-12.9	124	0.00
32	Benzoic acid	40.000	36.111	9.7	102	0.00
33	4-Chloroaniline	40.000	45.114	-12.8	122	0.00
34 C	Hexachlorobutadiene	40.000	43.948	-9.9	124	0.00
35	Caprolactam	40.000	43.691	-9.2	117	0.00
36 C	4-Chloro-3-methylphenol	40.000	45.482	-13.7	121	0.00
37	2-Methylnaphthalene	40.000	44.642	-11.6	124	0.00
38	1-Methylnaphthalene	40.000	44.763	-11.9	127	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.897	-12.2	127	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.240	-3.1	115	0.00
42 S	2,4,6-Tribromophenol	80.000	77.769	2.8	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.803	-9.5	119	0.00
44	2,4,5-Trichlorophenol	40.000	44.873	-12.2	122	0.00
45 S	2-Fluorobiphenyl	80.000	89.005	-11.3	126	0.00
46	1,1'-Biphenyl	40.000	44.587	-11.5	126	0.00
47	2-Chloronaphthalene	40.000	44.496	-11.2	125	0.00
48	2-Nitroaniline	40.000	41.655	-4.1	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
 Data File : BG060710.D
 Acq On : 1 Apr 2024 11:08
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 01 12:45:57 2024
 Quant Title :
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	Acenaphthylene	40.000	45.636	-14.1	128	0.00
50	Dimethylphthalate	40.000	44.527	-11.3	123	0.00
51	2,6-Dinitrotoluene	40.000	42.894	-7.2	124	0.00
52 C	Acenaphthene	40.000	45.139	-12.8	126	0.00
53	3-Nitroaniline	40.000	45.237	-13.1	127	0.00
54 P	2,4-Dinitrophenol	40.000	39.965	0.1	111	0.00
55	Dibenzofuran	40.000	44.998	-12.5	126	0.00
56 P	4-Nitrophenol	40.000	46.284	-15.7	123	0.00
57	2,4-Dinitrotoluene	40.000	44.489	-11.2	128	0.00
58	Fluorene	40.000	44.644	-11.6	124	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.920	-12.3	118	0.00
60	Diethylphthalate	40.000	44.263	-10.7	123	0.00
61	4-Chlorophenyl-phenylether	40.000	44.586	-11.5	125	0.00
62	4-Nitroaniline	40.000	49.734	-24.3	129	0.00
63	Azobenzene	40.000	45.649	-14.1	126	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	37.798	5.5	110	0.00
66 c	n-Nitrosodiphenylamine	40.000	46.399	-16.0	125	0.00
67	4-Bromophenyl-phenylether	40.000	40.843	-2.1	110	0.00
68	Hexachlorobenzene	40.000	45.417	-13.5	123	0.00
69	Atrazine	40.000	44.417	-11.0	119	0.00
70 C	Pentachlorophenol	40.000	44.095	-10.2	116	0.00
71	Phenanthrene	40.000	45.097	-12.7	124	0.00
72	Anthracene	40.000	45.650	-14.1	125	0.00
73	Carbazole	40.000	45.101	-12.8	124	0.00
74	Di-n-butylphthalate	40.000	43.926	-9.8	118	0.00
75 C	Fluoranthene	40.000	43.559	-8.9	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzidine	40.000	48.978	-22.4	114	0.00
78	Pyrene	40.000	45.537	-13.8	122	0.00
79 S	Terphenyl-d14	80.000	88.337	-10.4	119	0.00
80	Butylbenzylphthalate	40.000	42.473	-6.2	110	0.00
81	Benzo(a)anthracene	40.000	44.943	-12.4	119	0.00
82	3,3'-Dichlorobenzidine	40.000	45.087	-12.7	114	0.00
83	Chrysene	40.000	44.925	-12.3	119	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.033	-10.1	113	0.00
85 c	Di-n-octyl phthalate	40.000	43.219	-8.0	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.503	-11.3	116	-0.01
88	Benzo(b)fluoranthene	40.000	44.249	-10.6	115	0.00
89	Benzo(k)fluoranthene	40.000	45.921	-14.8	118	0.00
90 C	Benzo(a)pyrene	40.000	45.276	-13.2	116	0.00
91	Dibenzo(a,h)anthracene	40.000	45.026	-12.6	117	-0.01
92	Benzo(g,h,i)perylene	40.000	44.749	-11.9	117	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_G
LabSampleId :
SSTDCCC040