



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
Data File : BG060716.D
Acq On : 1 Apr 2024 15:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Mar 30 03:12:26 2024






AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.0	89520	PASS
68	69	0.00	2	0.6	393	PASS
69	198	0.00	100	33.5	65269	PASS
70	69	0.00	2	0.4	245	PASS
127	198	10	80	40.9	79616	PASS
197	198	0.00	2	0.3	612	PASS
198	198	100	100	100.0	194603	PASS
199	198	5	9	6.5	12671	PASS
275	198	10	60	29.6	57568	PASS
365	198	1	100	4.6	8974	PASS
441	198	0.01	100	13.8	26933	PASS
442	442	50	100	100.0	175424	PASS
443	442	15	24	19.5	34157	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
Data File : BG060716.D
Acq On : 1 Apr 2024 15:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 01 16:37:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration



TIC: BG060716.D\data.ms

(70) Pentachlorophenol (C)

16.977min (-0.003) 411689.79 ng




response 200041

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.30	63.14
264.00	64.70	63.19
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040124\
Data File : BG060716.D
Acq On : 1 Apr 2024 15:24
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 01 16:37:24 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration



TIC: BG060716.D\data.ms

(77) Benzidine

19.568min (-0.003) 819732.87 ng

response 1273535

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.31
183.00	11.90	11.90
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/1/2024	BNA_G	<u>BG060716.D</u>
Compound Name	Response	Retention Time
DDT	497769	20.814
DDD	7546	20.426
DDE	1676	19.868
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
9222	506991	1.82

Instrument :
BNA_G
ClientSampleId :
DFTPP