

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040418\
 Data File : BG033573.D
 Acq On : 5 Apr 2018 00:37
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 05 04:03:02 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 04 18:35:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.542	0.529	2.4	126	0.00
3	Pyridine	1.497	1.681	-12.3	130	0.00
4	n-Nitrosodimethylamine	0.614	0.671	-9.3	139	0.00
5 S	2-Fluorophenol	1.067	1.139	-6.7	126	0.00
6	Aniline	2.155	2.412	-11.9	133	0.00
7 S	Phenol-d6	1.833	2.058	-12.3	139	0.00
8	2-Chlorophenol	1.219	1.352	-10.9	131	0.00
9	Benzaldehyde	1.165	0.964	17.3	108	0.00
10 C	Phenol	1.809	2.026	-12.0	136	0.00
11	bis(2-Chloroethyl)ether	1.477	1.574	-6.6	126	0.00
12	1,3-Dichlorobenzene	1.594	1.574	1.3	124	0.00
13 C	1,4-Dichlorobenzene	1.597	1.569	1.8	126	0.00
14	1,2-Dichlorobenzene	1.558	1.612	-3.5	127	0.00
15	Benzyl Alcohol	1.443	1.627	-12.8	135	0.00
16	2,2'-oxybis(1-Chloropropane	2.408	2.673	-11.0	139	0.00
17	2-Methylphenol	1.233	1.386	-12.4	141	0.00
18	Hexachloroethane	0.616	0.619	-0.5	128	0.00
19 P	n-Nitroso-di-n-propylamine	1.343	1.493	-11.2	131	0.00
20	3+4-Methylphenols	1.755	2.072	-18.1	141	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
22	Acetophenone	0.548	0.563	-2.7	125	0.00
23 S	Nitrobenzene-d5	0.435	0.448	-3.0	131	0.00
24	Nitrobenzene	0.466	0.465	0.2	126	0.00
25	Isophorone	0.845	0.875	-3.6	131	0.00
26 C	2-Nitrophenol	0.176	0.185	-5.1	127	0.00
27	2,4-Dimethylphenol	0.256	0.274	-7.0	142	0.00
28	bis(2-Chloroethoxy)methane	0.422	0.441	-4.5	132	0.00
29 C	2,4-Dichlorophenol	0.315	0.343	-8.9	136	0.00
30	1,2,4-Trichlorobenzene	0.409	0.399	2.4	126	0.00
31	Naphthalene	1.036	1.043	-0.7	130	0.00
32	Benzoic acid	0.238	0.219	8.0	135	0.00
33	4-Chloroaniline	0.464	0.489	-5.4	132	0.00
34 C	Hexachlorobutadiene	0.310	0.291	6.1	125	0.00
35	Caprolactam	0.123	0.136	-10.6	138	0.00
36 C	4-Chloro-3-methylphenol	0.411	0.460	-11.9	136	0.00
37	2-Methylnaphthalene	0.804	0.830	-3.2	137	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
39	1,2,4,5-Tetrachlorobenzene	0.724	0.684	5.5	125	0.00
40 P	Hexachlorocyclopentadiene	0.425	0.351	17.4	106	0.00
41 S	2,4,6-Tribromophenol	0.299	0.282	5.7	121	0.00
42 C	2,4,6-Trichlorophenol	0.421	0.433	-2.9	129	0.00
43	2,4,5-Trichlorophenol	0.498	0.515	-3.4	126	0.00
44 S	2-Fluorobiphenyl	1.385	1.343	3.0	127	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040418\
 Data File : BG033573.D
 Acq On : 5 Apr 2018 00:37
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 05 04:03:02 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 04 18:35:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.537	1.549	-0.8	132	0.00
46	2-Chloronaphthalene	1.185	1.178	0.6	129	0.00
47	2-Nitroaniline	0.444	0.491	-10.6	140	0.00
48	Acenaphthylene	1.978	1.995	-0.9	132	0.00
49	Dimethylphthalate	1.741	1.732	0.5	131	0.00
50	2,6-Dinitrotoluene	0.356	0.366	-2.8	129	0.00
51 C	Acenaphthene	1.275	1.280	-0.4	130	0.00
52	3-Nitroaniline	0.373	0.387	-3.8	134	0.00
53 P	2,4-Dinitrophenol	0.149	0.124	16.8	100	0.00
54	Dibenzofuran	1.883	1.881	0.1	131	0.00
55 P	4-Nitrophenol	0.262	0.246	6.1	119	0.00
56	2,4-Dinitrotoluene	0.524	0.532	-1.5	132	0.00
57	Fluorene	1.604	1.573	1.9	131	0.00
58	2,3,4,6-Tetrachlorophenol	0.468	0.459	1.9	124	0.00
59	Diethylphthalate	1.690	1.708	-1.1	133	0.00
60	4-Chlorophenyl-phenylether	0.922	0.885	4.0	129	0.00
61	4-Nitroaniline	0.378	0.388	-2.6	133	0.00
62	Azobenzene	1.637	1.686	-3.0	134	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
64	4,6-Dinitro-2-methylphenol	0.120	0.117	2.5	117	0.00
65 c	n-Nitrosodiphenylamine	0.571	0.600	-5.1	131	0.00
66	4-Bromophenyl-phenylether	0.235	0.238	-1.3	126	0.00
67	Hexachlorobenzene	0.248	0.249	-0.4	127	0.00
68	Atrazine	0.231	0.232	-0.4	124	0.00
69 C	Pentachlorophenol	0.170	0.170	0.0	124	0.00
70	Phenanthrene	1.042	1.065	-2.2	128	0.00
71	Anthracene	1.055	1.081	-2.5	128	0.00
72	Carbazole	0.935	0.973	-4.1	129	0.00
73	Di-n-butylphthalate	1.088	1.154	-6.1	131	0.00
74 C	Fluoranthene	1.289	1.292	-0.2	126	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
76	Benzidine	0.542	0.519	4.2	112	0.00
77	Pyrene	1.334	1.318	1.2	125	0.00
78 S	Terphenyl-d14	0.935	0.902	3.5	124	0.00
79	Butylbenzylphthalate	0.492	0.530	-7.7	136	0.00
80	Benzo(a)anthracene	1.274	1.276	-0.2	129	0.00
81	3,3'-Dichlorobenzidine	0.453	0.480	-6.0	129	0.00
82	Chrysene	1.233	1.246	-1.1	130	0.00
83	Bis(2-ethylhexyl)phthalate	0.679	0.746	-9.9	138	0.00
84 c	Di-n-octyl phthalate	1.039	1.194	-14.9	143	0.00
85	Indeno(1,2,3-cd)pyrene	1.333	1.332	0.1	125	0.00
86 I	Perylene-d12	1.000	1.000	0.0	135	0.01
87	Benzo(b)fluoranthene	1.155	1.118	3.2	129	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040418\
Data File : BG033573.D
Acq On : 5 Apr 2018 00:37
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 05 04:03:02 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 04 18:35:46 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.169	1.188	-1.6	135	0.00
89 C	Benzo(a)pyrene	1.110	1.121	-1.0	134	0.00
90	Dibenzo(a,h)anthracene	1.045	0.963	7.8	119	0.00
91	Benzo(a,h,i)perylene	1.035	0.976	5.7	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0