

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040422\
 Data File : BG053021.D
 Acq On : 4 Apr 2022 23:35
 Operator : CG/JU
 Sample : N2207-04
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 20220343-098-21-01

Quant Time: Apr 05 01:08:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 16:48:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.129	152	36070	20.000	ng	-0.02
21) Naphthalene-d8	10.937	136	137115	20.000	ng	-0.02
39) Acenaphthene-d10	14.744	164	97053	20.000	ng	-0.02
64) Phenanthrene-d10	17.487	188	229986	20.000	ng	-0.02
76) Chrysene-d12	21.770	240	246611	20.000	ng	-0.03
86) Perylene-d12	25.071	264	257859	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.691	112	204456	98.714	ng	0.00
7) Phenol-d6	7.278	99	295002	94.475	ng	-0.01
23) Nitrobenzene-d5	9.287	82	207363	75.904	ng	-0.02
42) 2,4,6-Tribromophenol	16.230	330	137098	105.126	ng	-0.02
45) 2-Fluorobiphenyl	13.369	172	423495	64.356	ng	-0.02
79) Terphenyl-d14	20.090	244	786728	56.748	ng	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040422\
Data File : BG053021.D
Acq On : 4 Apr 2022 23:35
Operator : CG/JU
Sample : N2207-04
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
20220343-098-21-01

Quant Time: Apr 05 01:08:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 16:48:57 2022
Response via : Initial Calibration

