

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG040615\
 Data File : BG016224.D
 Acq On : 6 Apr 2015 15:37
 Operator : TP/IZ
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

apatel
 4/7/2015 5:08:10 PM

Quant Time: Apr 07 04:32:48 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 03:04:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	53694	20.00	ng	0.00
21) Naphthalene-d8	10.48	136	257957	20.00	ng	0.00
38) Acenaphthene-d10	14.34	164	153773	20.00	ng	0.00
63) Phenanthrene-d10	17.08	188	332126	20.00	ng	0.00
75) Chrysene-d12	21.26	240	307477	20.00	ng	0.00
86) Perylene-d12	23.51	264	271608	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.27	112	17813	5.51	ng	0.00
7) Phenol-d6	6.86	99	26149	5.74	ng	0.00
23) Nitrobenzene-d5	8.85	82	21696	5.17	ng	0.00
41) 2,4,6-Tribromophenol	15.83	330	4242	2.56	ng	0.00
44) 2-Fluorobiphenyl	12.95	172	49726	5.46	ng	0.00
78) Terphenyl-d14	19.70	244	75746	6.18	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.17	88	4471	3.36	ng	98
3) Pyridine	3.57	79	11559	2.79	ng	# 85
4) n-Nitrosodimethylamine	3.49	42	3154	2.58	ng	# 84
6) Aniline	7.02	93	18805	2.89	ng	97
8) 2-Chlorophenol	7.25	128	9851	2.56	ng	91
9) Benzaldehyde	6.84	77	8764	4.39	ng	94
10) Phenol	6.89	94	13404	2.69	ng	94
11) bis(2-Chloroethyl)ether	7.12	93	11880	2.95	ng	97
12) 1,3-Dichlorobenzene	7.58	146	11082	2.70	ng	96
13) 1,4-Dichlorobenzene	7.73	146	12073	2.84	ng	88
14) 1,2-Dichlorobenzene	8.04	146	11156	2.76	ng	94
15) Benzyl Alcohol	7.93	79	7776	2.33	ng	95
16) 2,2'-oxybis(1-Chloropropan	8.23	45	13016	2.87	ng	99
17) 2-Methylphenol	8.14	107	8902	2.55	ng	96
18) Hexachloroethane	8.77	117	4713	3.06	ng	# 79
19) n-Nitroso-di-n-propylamine	8.49	70	8064	2.41	ng	# 94
20) 3+4-Methylphenols	8.46	107	11369	2.39	ng	98
22) Acetophenone	8.51	105	15177	2.61	ng	# 97
24) Nitrobenzene	8.89	77	11776	2.61	ng	90
25) Isophorone	9.41	82	23984	2.53	ng	97
27) 2,4-Dimethylphenol	9.66	122	10002	2.45	ng	93
28) bis(2-Chloroethoxy)methane	9.90	93	14703	2.70	ng	99
29) 2,4-Dichlorophenol	10.13	162	7082	2.00	ng	92
30) 1,2,4-Trichlorobenzene	10.35	180	8927	2.36	ng	95
31) Naphthalene	10.53	128	36771	2.68	ng	99
33) 4-Chloroaniline	10.64	127	15256	2.51	ng	98
36) 4-Chloro-3-methylphenol	11.77	107	10059	2.37	ng	98
37) 2-Methylnaphthalene	12.15	142	25630	2.61	ng	91
39) 1,2,4,5-Tetrachlorobenzene	12.52	216	8692	2.30	ng	98
45) 1,1'-Biphenyl	13.17	154	31645	2.81	ng	99
46) 2-Chloronaphthalene	13.21	162	24278	2.77	ng	96
47) 2-Nitroaniline	13.42	65	5580	2.16	ng	89
48) Acenaphthylene	14.06	152	41581	2.60	ng	98
49) Dimethylphthalate	13.79	163	29082	2.69	ng	98

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG040615\
 Data File : BG016224.D
 Acq On : 6 Apr 2015 15:37
 Operator : TP/IZ
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5


Manual Integrations
 APPROVED

apatel
 4/7/2015 5:08:10 PM

Quant Time: Apr 07 04:32:48 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 07 03:04:07 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
50) 2,6-Dinitrotoluene	13.91	165	6006	2.34	ng	95
51) Acenaphthene	14.40	154	25936	2.70	ng	97
52) 3-Nitroaniline	14.25	138	7505	2.36	ng	90
54) Dibenzofuran	14.73	168	36527	2.72	ng	99
56) 2,4-Dinitrotoluene	14.70	165	7298	2.07	ng	# 95
57) Fluorene	15.39	166	31523	2.64	ng	99
59) Diethylphthalate	15.16	149	30547	2.73	ng	98
60) 4-Chlorophenyl-phenylether	15.38	204	12727	2.44	ng	93
61) 4-Nitroaniline	15.40	138	8716	2.43	ng	99
62) Azobenzene	15.67	77	34486	2.99	ng	98
65) n-Nitrosodiphenylamine	15.60	169	28315	2.66	ng	95
68) Atrazine	16.54	200	7294	2.32	ng	94
70) Phenanthrene	17.12	178	53295	2.87	ng	95
71) Anthracene	17.21	178	49300	2.67	ng	96
72) Carbazole	17.48	167	49353	2.72	ng	97
73) Di-n-butylphthalate	18.05	149	57812	2.78	ng	96
74) Fluoranthene	19.14	202	52439	2.57	ng	94
76) Benzidine	19.33	184	28447	3.26	ng	98
77) Pyrene	19.50	202	57122	2.97	ng	95
79) Butylbenzylphthalate	20.40	149	21020	2.53	ng	97
80) Benzo(a)anthracene	21.24	228	49141	2.82	ng	92
81) 3,3'-Dichlorobenzidine	21.18	252	13637	2.26	ng	94
82) Chrysene	21.30	228	49726	3.09	ng	98
83) Bis(2-ethylhexyl)phthalate	21.17	149	30620	2.70	ng	98
84) Di-n-octyl phthalate	22.05	149	44570	2.33	ng	99
85) Indeno(1,2,3-cd)pyrene	25.79	276	42602	2.11	ng	96
87) Benzo(b)fluoranthene	22.82	252	42112	2.68	ng	99
88) Benzo(k)fluoranthene	22.87	252	40575	2.67	ng	99
89) Benzo(a)pyrene	23.41	252	37318	2.56	ng	97
90) Dibenzo(a,h)anthracene	25.80	278	36536	2.43	ng	96
91) Benzo(g,h,i)perylene	26.49	276	35782	2.39	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.70 min Scan# 852
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
152	100		
150	151.5	130.1	195.1
115	57.9	50.5	75.7

Manual Integrations
APPROVED




apatel
4/7/2015 5:08:10 PM




#2

1,4-Dioxane
Concen: 3.36 ng
RT: 3.17 min Scan# 82
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
88	100		
58	64.6	50.6	76.0
43	21.7	17.0	25.4





#3

Pyridine

Concen: 2.79 ng

RT: 3.57 min Scan# 150

Delta R.T. 0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	79	Resp:	11559
Ion	Ratio	Lower	Upper
79	100		
52	51.4	49.0	73.4
51	37.8	22.5	33.7#

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#4

n-Nitrosodimethylamine

Concen: 2.58 ng




RT: 3.49 min Scan# 137


Delta R.T. 0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	42	Resp:	3154
Ion	Ratio	Lower	Upper
42	100		
74	168.8	153.7	230.5
44	17.9	5.3	7.9#





#5

2-Fluorophenol

Concen: 5.51 ng

RT: 5.27 min Scan# 439

Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	112	Resp:	17813
Ion Ratio	Lower	Upper	
112	100		
64	57.0	44.5	66.7
63	27.4	21.4	32.0

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#6

Aniline

Concen: 2.89 ng




RT: 7.02 min Scan# 737


Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	93	Resp:	18805
Ion Ratio	Lower	Upper	
93	100		
66	34.6	27.5	41.3
65	15.5	15.1	22.7





#7

Phenol-d6

Concen: 5.74 ng

RT: 6.86 min Scan# 710

Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	99	Resp:	26149
Ion Ratio	Lower	Upper	
99	100		
42	9.4	10.0	15.0#
71	21.5	21.5	32.3

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#8

2-Chlorophenol

Concen: 2.56 ng




RT: 7.25 min Scan# 777


Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	128	Resp:	9851
Ion Ratio	Lower	Upper	
128	100		
130	37.9	11.6	51.6
64	49.2	24.2	64.2





#9

Benzaldehyde

Concen: 4.39 ng

RT: 6.84 min Scan# 706

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	77	Resp:	8764
Ion	Ratio	Lower	Upper
77	100		
105	88.7	76.8	116.8
106	103.6	80.1	120.1

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#10

Phenol

Concen: 2.69 ng




RT: 6.89 min Scan# 715


Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	94	Resp:	13404
Ion	Ratio	Lower	Upper
94	100		
65	28.3	3.8	43.8
66	25.7	7.4	47.4









#11
bis(2-Chloroethyl)ether
Concen: 2.95 ng
RT: 7.12 min Scan# 754
Delta R.T. -0.01 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:	93	Resp:	11880
Ion	Ratio	Lower	Upper
93	100		
63	66.2	44.1	84.1
95	33.9	12.5	52.5




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#12
1,3-Dichlorobenzene
Concen: 2.70 ng
RT: 7.58 min Scan# 833
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion:	146	Resp:	11082
Ion	Ratio	Lower	Upper
146	100		
148	67.9	50.9	76.3
75	28.5	23.3	34.9









#13
 1,4-Dichlorobenzene
 Concen: 2.84 ng
 RT: 7.73 min Scan# 858
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
146	100		
148	53.5	52.0	78.0
111	35.7	32.6	49.0




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#14
 1,2-Dichlorobenzene
 Concen: 2.76 ng
 RT: 8.04 min Scan# 911
 Delta R.T. -0.01 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
146	100		
148	59.0	50.9	76.3
111	39.8	34.7	52.1





#15

Benzyl Alcohol

Concen: 2.33 ng

RT: 7.93 min Scan# 892

Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	79	Resp:	7776
Ion	Ratio	Lower	Upper
79	100		
108	94.3	69.3	103.9
77	63.5	50.3	75.5

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 2.87 ng




RT: 8.23 min Scan# 943


Delta R.T. -0.01 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	45	Resp:	13016
Ion	Ratio	Lower	Upper
45	100		
77	14.8	0.0	35.3
79	12.8	0.0	32.3









#17
2-Methylphenol
Concen: 2.55 ng
RT: 8.14 min Scan# 927
Delta R.T. -0.01 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
107	100		
108	106.2	87.2	130.8
77	43.3	31.1	46.7
79	40.1	29.7	44.5




Manual Integrations
APPROVED




apatel
4/7/2015 5:08:10 PM



#18
Hexachloroethane
Concen: 3.06 ng
RT: 8.77 min Scan# 1035
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
117	100		
119	86.0	78.8	118.2
201	47.2	59.0	88.4#









#19
n-Nitroso-di-n-propylamine
Concen: 2.41 ng
RT: 8.49 min Scan# 988
Delta R.T. -0.01 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
70	100		
42	48.4	34.7	52.1
101	14.0	9.2	13.8
130	25.8	22.6	33.8


Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#20
3+4-Methylphenols
Concen: 2.39 ng
RT: 8.46 min Scan# 983
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
107	100		
108	88.2	69.5	109.5
77	24.7	8.2	48.2
79	24.5	3.8	43.8





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 10.48 min Scan# 1326

Delta R.T. -0.01 min

Lab File: BG016224.D


Acq: 6 Apr 2015 15:37

Instrument :

BNA_G

ClientSampleId :




SSTDICC2.5



Tgt Ion	Ratio	Resp	Lower	Upper
136	100	257957		
137	10.9	8.7	13.1	
54	8.3	6.3	9.5	
68	8.2	6.5	9.7	

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#22

Acetophenone


Concen: 2.61 ng

RT: 8.51 min Scan# 991



Delta R.T. -0.01 min


Lab File: BG016224.D

Acq: 6 Apr 2015 15:37



Tgt Ion	Ratio	Resp	Lower	Upper
105	100	15177		
71	0.0	2.0	3.0#	
51	24.4	17.9	26.9	
120	31.9	26.3	39.5	









#23
 Nitrobenzene-d5
 Concen: 5.17 ng
 RT: 8.85 min Scan# 1048
 Delta R.T. -0.01 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	21696		
128	43.9	37.8	56.8	
54	36.2	32.6	49.0	




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#24
 Nitrobenzene
 Concen: 2.61 ng
 RT: 8.89 min Scan# 1055
 Delta R.T. -0.01 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	11776		
123	40.5	38.8	58.2	
65	11.0	9.4	14.2	





#25

Isophorone

Concen: 2.53 ng

RT: 9.41 min Scan# 1144

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	82	Resp:	23984
Ion Ratio	Lower	Upper	
82	100		
95	5.5	5.0	7.6
138	18.6	16.2	24.2

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#27

2,4-Dimethylphenol

Concen: 2.45 ng




RT: 9.66 min Scan# 1187


Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	122	Resp:	10002
Ion Ratio	Lower	Upper	
122	100		
107	96.6	83.4	125.2
121	53.5	46.2	69.4









#28
bis(2-Chloroethoxy)methane
Concen: 2.70 ng
RT: 9.90 min Scan# 1227
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	14703		
95	32.5	26.1	39.1	
123	13.5	11.8	17.8	




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#29
2,4-Dichlorophenol
Concen: 2.00 ng
RT: 10.13 min Scan# 1267
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	7082		
164	70.9	45.0	85.0	
98	44.6	19.1	59.1	









#30
 1,2,4-Trichlorobenzene
 Concen: 2.36 ng
 RT: 10.35 min Scan# 1303
 Delta R.T. -0.01 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	8927		
182	89.3	76.3	114.5	
145	31.6	25.2	37.8	




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#31
 Naphthalene
 Concen: 2.68 ng
 RT: 10.53 min Scan# 1335
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	36771		
129	12.2	9.2	13.8	
127	12.8	10.6	16.0	






#33
4-Chloroaniline
Concen: 2.51 ng
RT: 10.64 min Scan# 1354
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5




Tgt Ion	Ratio	Resp	Lower	Upper
127	100	15256		
129	31.7	26.1	39.1	
65	26.1	21.2	31.8	
92	18.1	13.8	20.6	

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#36
4-Chloro-3-methylphenol
Concen: 2.37 ng
RT: 11.77 min Scan# 1546
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	10059		
144	27.2	22.0	33.0	
142	81.4	67.0	100.6	



Abundance Ion 107.00 (106.70 to 107.70): E
Ion 144.00 (143.70 to 144.70): E
Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 2.61 ng
RT: 12.15 min Scan# 1610
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
142	100		
141	77.3	69.3	103.9
115	28.7	25.0	37.6




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.34 min Scan# 1983
Delta R.T. 0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
164	100		
162	96.5	77.3	115.9
160	45.9	36.7	55.1






#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 2.30 ng
 RT: 12.52 min Scan# 1673
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion:	216	Resp:	8692
Ion Ratio	Lower	Upper	
216	100		
214	77.0	63.4	95.0
179	22.9	18.2	27.2
108	26.0	21.2	31.8

Manual Integrations
 APPROVED

apatel
 4/7/2015 5:08:10 PM






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E


Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#41
 2,4,6-Tribromophenol
 Concen: 2.56 ng
 RT: 15.83 min Scan# 2236
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37



Tgt Ion:	330	Resp:	4242
Ion Ratio	Lower	Upper	
330	100		
332	95.5	75.8	113.6
141	59.5	46.6	69.8




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 5.46 ng
 RT: 12.95 min Scan# 1747
 Delta R.T. -0.01 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	49726		
171	33.7		29.6	44.4
170	21.8		19.4	29.2




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#45
 1,1'-Biphenyl
 Concen: 2.81 ng
 RT: 13.17 min Scan# 1784
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	31645		
153	42.6		22.4	62.4
76	18.3		0.0	37.1









#46
 2-Chloronaphthalene
 Concen: 2.77 ng
 RT: 13.21 min Scan# 1790
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
162	100		
127	36.1	30.8	46.2
164	30.4	26.6	40.0




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#47
 2-Nitroaniline
 Concen: 2.16 ng
 RT: 13.42 min Scan# 1826
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
65	100		
92	72.3	61.0	91.6
138	106.0	98.7	148.1





#48

Acenaphthylene

Concen: 2.60 ng

RT: 14.06 min Scan# 1935

Delta R.T. -0.00 min

Lab File: BG016224.D


Acq: 6 Apr 2015 15:37

Instrument :

BNA_G

ClientSampleId :




SSTDICC2.5



Tgt Ion:	152	Resp:	41581
Ion Ratio	Lower	Upper	
152	100		
151	20.8	16.5	24.7
153	11.9	11.1	16.7

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#49

Dimethylphthalate


Concen: 2.69 ng

RT: 13.79 min Scan# 1890



Delta R.T. -0.01 min


Lab File: BG016224.D

Acq: 6 Apr 2015 15:37



Tgt Ion:	163	Resp:	29082
Ion Ratio	Lower	Upper	
163	100		
194	6.3	5.0	7.4
164	11.1	8.2	12.4









#50
2,6-Dinitrotoluene
Concen: 2.34 ng
RT: 13.91 min Scan# 1910
Delta R.T. -0.01 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	6006		
63	54.0	39.3	58.9	
89	49.0	41.3	61.9	




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#51
Acenaphthene
Concen: 2.70 ng
RT: 14.40 min Scan# 1993
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	25936		
153	112.6	87.5	131.3	
152	51.5	42.1	63.1	





#52

3-Nitroaniline

Concen: 2.36 ng

RT: 14.25 min Scan# 1967

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5




Tgt Ion:	138	Resp:	7505
Ion Ratio	Lower	Upper	
138	100		
108	10.4	7.4	11.0
92	126.1	91.4	137.0

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): B



#54

Dibenzofuran

Concen: 2.72 ng

RT: 14.73 min Scan# 2050




Delta R.T. -0.00 min


Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	168	Resp:	36527
Ion Ratio	Lower	Upper	
168	100		
139	39.6	31.0	46.6
169	13.9	11.2	16.8

Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#56
2,4-Dinitrotoluene
Concen: 2.07 ng
RT: 14.70 min Scan# 2045
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	7298		
63	44.2		31.4	47.2
89	76.8		58.8	88.2
182	3.3		4.1	6.1




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#57
Fluorene
Concen: 2.64 ng
RT: 15.39 min Scan# 2161
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
166	100	31523		
165	93.5		75.2	112.8
167	14.6		11.3	16.9









#59
Diethylphthalate
Concen: 2.73 ng
RT: 15.16 min Scan# 2123
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
149	100		
177	21.3	17.7	26.5
150	11.8	10.0	15.0




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#60
4-Chlorophenyl-phenylether
Concen: 2.44 ng
RT: 15.38 min Scan# 2160
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
204	100		
206	35.6	27.6	41.4
141	61.7	55.4	83.2









#61
4-Nitroaniline
Concen: 2.43 ng
RT: 15.40 min Scan# 2164
Delta R.T. -0.01 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
138	100		
92	44.5	26.3	66.3
108	53.5	33.6	73.6




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#62
Azobenzene
Concen: 2.99 ng
RT: 15.67 min Scan# 2210
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
77	100		
182	24.4	6.6	46.6
105	19.0	0.0	39.2
51	22.2	2.0	42.0









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.08 min Scan# 2449
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
188	100	332126		
94	17.0		12.2	18.4
80	16.6		12.8	19.2




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#65
n-Nitrosodiphenylamine
Concen: 2.66 ng
RT: 15.60 min Scan# 2197
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	28315		
168	62.9		53.4	80.0
167	38.0		28.3	42.5









#68
 Atrazine
 Concen: 2.32 ng
 RT: 16.54 min Scan# 2358
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
200	100		
173	34.9	10.2	50.2
215	51.5	34.4	74.4




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#70
 Phenanthrene
 Concen: 2.87 ng
 RT: 17.12 min Scan# 2456
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.3	16.2	24.4
179	13.8	12.7	19.1









#71
 Anthracene
 Concen: 2.67 ng
 RT: 17.21 min Scan# 2472
 Delta R.T. 0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
178	100		
176	17.4	15.9	23.9
179	14.9	13.0	19.6




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#72
 Carbazole
 Concen: 2.72 ng
 RT: 17.48 min Scan# 2518
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.0	17.9	26.9
139	13.2	11.5	17.3





#73

Di-n-butylphthalate

Concen: 2.78 ng

RT: 18.05 min Scan# 2614

Delta R.T. 0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	149	Resp:	57812
Ion Ratio	Lower	Upper	
149	100		
150	8.4	8.2	12.4
104	4.4	4.0	6.0

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#74

Fluoranthene

Concen: 2.57 ng




RT: 19.14 min Scan# 2800


Delta R.T. 0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	202	Resp:	52439
Ion Ratio	Lower	Upper	
202	100		
101	19.2	1.7	41.7
203	15.9	0.0	38.7





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.26 min Scan# 3160

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5

Tgt Ion:	240	Resp:	307477
Ion Ratio	Lower	Upper	
240	100		
120	24.1	18.6	28.0
236	26.7	21.7	32.5

Manual Integrations
APPROVED




apatel
4/7/2015 5:08:10 PM



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 3.26 ng


RT: 19.33 min Scan# 2832

Delta R.T. 0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37



Tgt Ion:	184	Resp:	28447
Ion Ratio	Lower	Upper	
184	100		
185	14.8	12.6	19.0
183	12.2	10.4	15.6




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 2.97 ng

RT: 19.50 min Scan# 2861

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	202	Resp:	57122
Ion Ratio	Lower	Upper	
202	100		
200	19.3	17.8	26.8
203	17.2	14.7	22.1

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#78

Terphenyl-d14

Concen: 6.18 ng




RT: 19.70 min Scan# 2896


Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	244	Resp:	75746
Ion Ratio	Lower	Upper	
244	100		
212	6.9	6.6	9.8
122	21.7	19.1	28.7









#79
Butylbenzylphthalate
Concen: 2.53 ng
RT: 20.40 min Scan# 3015
Delta R.T. 0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion:	149	Resp:	21020
Ion Ratio	Lower	Upper	
149	100		
91	71.4	59.4	89.2
206	20.2	15.6	23.4




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#80
Benzo(a)anthracene
Concen: 2.82 ng
RT: 21.24 min Scan# 3158
Delta R.T. 0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion:	228	Resp:	49141
Ion Ratio	Lower	Upper	
228	100		
226	24.6	23.5	35.3
229	18.2	16.8	25.2









#81
 3,3'-Dichlorobenzidine
 Concen: 2.26 ng
 RT: 21.18 min Scan# 3147
 Delta R.T. -0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	13637		
254	58.3	50.8	76.2	
126	19.2	16.0	24.0	




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#82
 Chrysene
 Concen: 3.09 ng
 RT: 21.30 min Scan# 3167
 Delta R.T. 0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	49726		
226	30.1	25.0	37.4	
229	20.6	17.0	25.4	





#83

Bis(2-ethylhexyl)phthalate

Concen: 2.70 ng

RT: 21.17 min Scan# 3146

Delta R.T. -0.00 min

Lab File: BG016224.D


Acq: 6 Apr 2015 15:37

Instrument :

BNA_G

ClientSampleId :




SSTDICC2.5



Tgt Ion:	149	Resp:	30620
Ion Ratio	Lower	Upper	
149	100		
167	30.9	25.8	38.6
279	5.8	4.7	7.1

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#84

Di-n-octyl phthalate

Concen: 2.33 ng




RT: 22.05 min Scan# 3295


Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	149	Resp:	44570
Ion Ratio	Lower	Upper	
149	100		
167	1.5	1.3	1.9
43	9.2	7.0	10.4









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 2.11 ng
 RT: 25.79 min Scan# 3931
 Delta R.T. -0.02 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion:	276	Resp:	42602
Ion Ratio	Lower	Upper	
276	100		
138	55.7	41.5	62.3
277	26.9	21.0	31.6




Manual Integrations
 APPROVED


apatel
 4/7/2015 5:08:10 PM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 23.51 min Scan# 3543
 Delta R.T. 0.00 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Tgt Ion:	264	Resp:	271608
Ion Ratio	Lower	Upper	
264	100		
260	24.3	20.4	30.6
265	21.2	17.2	25.8





#87

Benzo(b)fluoranthene

Concen: 2.68 ng

RT: 22.82 min Scan# 3427

Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	252	Resp:	42112
Ion Ratio	Lower	Upper	
252	100		
253	22.8	18.5	27.7
125	20.2	17.0	25.6

Manual Integrations
APPROVED

apatel
4/7/2015 5:08:10 PM



#88

Benzo(k)fluoranthene

Concen: 2.67 ng




RT: 22.87 min Scan# 3435


Delta R.T. -0.00 min

Lab File: BG016224.D

Acq: 6 Apr 2015 15:37

Tgt Ion:	252	Resp:	40575
Ion Ratio	Lower	Upper	
252	100		
253	21.5	17.5	26.3
125	21.1	16.3	24.5









#89
Benzo(a)pyrene
Concen: 2.56 ng
RT: 23.41 min Scan# 3526
Delta R.T. -0.00 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.5	17.6	26.4
125	21.1	17.6	26.4




Manual Integrations
APPROVED


apatel
4/7/2015 5:08:10 PM



#90
Dibenzo(a,h)anthracene
Concen: 2.43 ng
RT: 25.80 min Scan# 3933
Delta R.T. -0.02 min
Lab File: BG016224.D
Acq: 6 Apr 2015 15:37

Tgt Ion	Ratio	Lower	Upper
278	100		
139	31.6	27.9	41.9
279	23.8	19.5	29.3






#91
 Benzo(g,h,i)perylene
 Concen: 2.39 ng
 RT: 26.49 min Scan# 4050
 Delta R.T. -0.02 min
 Lab File: BG016224.D
 Acq: 6 Apr 2015 15:37

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.0	19.5	29.3
138	45.7	35.7	53.5

Manual Integrations
 APPROVED

apatel
 4/7/2015 5:08:10 PM



Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 277.00 (276.70 to 277.70): E
 Ion 138.00 (137.70 to 138.70): E

