


Data Path : Z:\HPCHEM1\BNA G\DATA\BG040618\  
 Data File : BG033646.D  
 Acq On : 7 Apr 2018 2:21  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG031918.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Apr 05 17:36:07 2018



AutoFind: Scans 2600, 2601, 2602; Background Corrected with Scan 2593

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.2      | 18369   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 259     | PASS             |
| 69          | 198          | 0.00         | 100          | 33.4      | 16067   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 58      | PASS             |
| 127         | 198          | 10           | 80           | 40.2      | 19333   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 48040   | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 3255    | PASS             |
| 275         | 198          | 10           | 60           | 27.1      | 13015   | PASS             |
| 365         | 198          | 1            | 100          | 3.2       | 1517    | PASS             |
| 441         | 443          | 0.01         | 100          | 91.6      | 5178    | PASS             |
| 442         | 198          | 50           | 100          | 67.9      | 32610   | PASS             |
| 443         | 442          | 15           | 24           | 17.3      | 5652    | PASS             |