

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040621\
 Data File : BG048598.D
 Acq On : 6 Apr 2021 10:46
 Operator : CG/JU
 Sample : PB135529BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB135529BL

Quant Time: Apr 06 11:28:06 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG033021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 02 10:20:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.090	152	22839	20.00	ng	# 0.00
21) Naphthalene-d8	10.916	136	89415	20.00	ng	# 0.00
39) Acenaphthene-d10	14.747	164	51308	20.00	ng	0.00
64) Phenanthrene-d10	17.503	188	143067	20.00	ng	0.00
76) Chrysene-d12	21.798	240	175190	20.00	ng	# 0.00
86) Perylene-d12	25.141	264	160540	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.634	112	197219	152.45	ng	0.00
7) Phenol-d6	7.244	99	239597	127.71	ng	0.00
23) Nitrobenzene-d5	9.260	82	173403	93.89	ng	0.00
42) 2,4,6-Tribromophenol	16.240	330	110822	164.31	ng	0.00
45) 2-Fluorobiphenyl	13.367	172	371933	107.74	ng	0.00
79) Terphenyl-d14	20.112	244	883911	92.27	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG040621\
Data File : BG048598.D
Acq On : 6 Apr 2021 10:46
Operator : CG/JU
Sample : PB135529BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB135529BL

Quant Time: Apr 06 11:28:06 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG033021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 02 10:20:56 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.090 min Scan# 809
Delta R.T. -0.004 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46




Instrument :
BNA_G
ClientSampleId :
PB135529BL


Tgt Ion:152 Resp: 22839
Ion Ratio Lower Upper
152 100
150 161.4 101.3 151.9#
115 62.5 42.3 63.5



#5
2-Fluorophenol
Concen: 152.45 ng
RT: 5.634 min Scan# 391
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46

Tgt Ion:112 Resp: 197219
Ion Ratio Lower Upper
112 100
64 53.8 50.4 75.6
63 33.2 31.4 47.2









#7
Phenol-d6
Concen: 127.71 ng
RT: 7.244 min Scan# 665
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46




Instrument :
BNA_G
ClientSampleId :
PB135529BL


Tgt Ion: 99 Resp: 239597
Ion Ratio Lower Upper
99 100
42 21.9 15.5 23.3
71 45.7 36.6 54.8



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.916 min Scan# 1290
Delta R.T. -0.009 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46

Tgt Ion:136 Resp: 89415
Ion Ratio Lower Upper
136 100
137 10.8 8.6 12.8
54 6.6 4.2 6.4#
68 4.4 3.0 4.4









#23
Nitrobenzene-d5
Concen: 93.89 ng
RT: 9.260 min Scan# 1008
Delta R.T. -0.009 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46




Instrument :
BNA_G
ClientSampleId :
PB135529BL


Tgt Ion: 82 Resp: 173403
Ion Ratio Lower Upper
82 100
128 39.3 26.2 39.4
54 49.3 38.9 58.3



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.747 min Scan# 1942
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46

Tgt Ion:164 Resp: 51308
Ion Ratio Lower Upper
164 100
162 100.5 87.0 130.6
160 50.3 38.4 57.6









#42
2,4,6-Tribromophenol
Concen: 164.31 ng
RT: 16.240 min Scan# 2196
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46




Instrument :
BNA_G
ClientSampleId :
PB135529BL


Tgt Ion:330 Resp: 110822
Ion Ratio Lower Upper
330 100
332 92.3 77.6 116.4
141 38.0 34.6 52.0



#45
2-Fluorobiphenyl
Concen: 107.74 ng
RT: 13.367 min Scan# 1707
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46

Tgt Ion:172 Resp: 371933
Ion Ratio Lower Upper
172 100
171 34.3 28.5 42.7
170 22.9 19.8 29.6









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.503 min Scan# 2411
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46




Instrument :
BNA_G
ClientSampleId :
PB135529BL


Tgt Ion:188 Resp: 143067
Ion Ratio Lower Upper
188 100
94 4.4 3.0 4.6
80 5.4 4.5 6.7



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.798 min Scan# 3142
Delta R.T. -0.003 min
Lab File: BG048598.D
Acq: 6 Apr 2021 10:46

Tgt Ion:240 Resp: 175190
Ion Ratio Lower Upper
240 100
120 4.8 2.6 3.8#
236 26.8 19.8 29.6









#79
 Terphenyl-d14
 Concen: 92.27 ng
 RT: 20.112 min Scan# 2855
 Delta R.T. -0.003 min
 Lab File: BG048598.D
 Acq: 6 Apr 2021 10:46

Instrument :
 BNA_G
 ClientSampleId :
 PB135529BL

Tgt Ion:244 Resp: 883911
 Ion Ratio Lower Upper
 244 100
 212 8.3 6.0 9.0
 122 6.8 5.8 8.6



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 25.141 min Scan# 3711
 Delta R.T. -0.009 min
 Lab File: BG048598.D
 Acq: 6 Apr 2021 10:46

Tgt Ion:264 Resp: 160540
 Ion Ratio Lower Upper
 264 100
 260 21.9 19.1 28.7
 265 22.5 16.1 24.1

