

Data Path : Z:\HPCHEM1\BNA_G\Data\BG040715\
 Data File : BG016274.D
 Acq On : 8 Apr 2015 1:20
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 08 03:49:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 08 03:25:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.576	0.493	14.4	106	0.00
3	Pyridine	1.722	1.622	5.8	113	0.00
4	n-Nitrosodimethylamine	0.512	0.472	7.8	110	0.00
5 S	2-Fluorophenol	1.267	1.219	3.8	116	0.00
6	Aniline	2.717	2.609	4.0	115	0.00
7 S	Phenol-d6	1.902	1.878	1.3	118	0.00
8	2-Chlorophenol	1.522	1.525	-0.2	121	0.00
9	Benzaldehyde	1.114	1.001	10.1	113	0.00
10 C	Phenol	2.035	2.020	0.7	118	0.00
11	bis(2-Chloroethyl)ether	1.623	1.522	6.2	114	0.00
12	1,3-Dichlorobenzene	1.537	1.488	3.2	118	0.00
13 C	1,4-Dichlorobenzene	1.594	1.525	4.3	120	0.00
14	1,2-Dichlorobenzene	1.525	1.474	3.3	119	0.00
15	Benzyl Alcohol	1.326	1.332	-0.5	117	0.00
16	2,2'-oxybis(1-Chloropropane	1.828	1.687	7.7	111	0.00
17	2-Methylphenol	1.365	1.370	-0.4	119	0.00
18	Hexachloroethane	0.610	0.570	6.6	116	0.00
19 P	n-Nitroso-di-n-propylamine	1.363	1.322	3.0	112	0.00
20	3+4-Methylphenols	1.891	1.923	-1.7	119	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
22	Acetophenone	0.464	0.432	6.9	114	0.00
23 S	Nitrobenzene-d5	0.345	0.327	5.2	114	0.00
24	Nitrobenzene	0.369	0.345	6.5	114	0.00
25	Isophorone	0.767	0.717	6.5	113	0.00
26 C	2-Nitrophenol	0.172	0.191	-11.0	127	0.00
27	2,4-Dimethylphenol	0.321	0.317	1.2	119	0.00
28	bis(2-Chloroethoxy)methane	0.438	0.413	5.7	116	0.00
29 C	2,4-Dichlorophenol	0.254	0.269	-5.9	125	0.00
30	1,2,4-Trichlorobenzene	0.265	0.257	3.0	120	0.00
31	Naphthalene	1.042	0.987	5.3	118	0.00
32	Benzoic acid	0.182	0.230	-26.4#	148	0.00
33	4-Chloroaniline	0.489	0.483	1.2	119	0.00
34 C	Hexachlorobutadiene	0.116	0.114	1.7	123	0.00
35	Caprolactam	0.160	0.162	-1.3	115	0.01
36 C	4-Chloro-3-methylphenol	0.335	0.340	-1.5	120	0.00
37	2-Methylnaphthalene	0.750	0.719	4.1	118	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
39	1,2,4,5-Tetrachlorobenzene	0.440	0.429	2.5	123	0.00
40 P	Hexachlorocyclopentadiene	0.202	0.193	4.5	118	0.00
41 S	2,4,6-Tribromophenol	0.135	0.140	-3.7	123	0.00
42 C	2,4,6-Trichlorophenol	0.327	0.346	-5.8	127	0.00
43	2,4,5-Trichlorophenol	0.350	0.370	-5.7	129	0.00
44 S	2-Fluorobiphenyl	1.175	1.104	6.0	118	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG040715\
 Data File : BG016274.D
 Acq On : 8 Apr 2015 1:20
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 08 03:49:29 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 08 03:25:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.534	1.442	6.0	119	0.00
46	2-Chloronaphthalene	1.169	1.101	5.8	120	0.00
47	2-Nitroaniline	0.377	0.366	2.9	115	0.00
48	Acenaphthylene	2.078	1.949	6.2	117	0.00
49	Dimethylphthalate	1.466	1.361	7.2	115	0.00
50	2,6-Dinitrotoluene	0.355	0.346	2.5	118	0.00
51 C	Acenaphthene	1.242	1.160	6.6	119	0.00
52	3-Nitroaniline	0.453	0.441	2.6	117	0.00
53 P	2,4-Dinitrophenol	0.167	0.174	-4.2	124	0.00
54	Dibenzofuran	1.723	1.605	6.8	118	0.00
55 P	4-Nitrophenol	0.374	0.357	4.5	116	0.00
56	2,4-Dinitrotoluene	0.483	0.475	1.7	118	0.00
57	Fluorene	1.520	1.414	7.0	117	0.00
58	2,3,4,6-Tetrachlorophenol	0.270	0.280	-3.7	127	0.00
59	Diethylphthalate	1.560	1.427	8.5	113	0.00
60	4-Chlorophenyl-phenylether	0.622	0.586	5.8	119	0.00
61	4-Nitroaniline	0.514	0.495	3.7	116	0.00
62	Azobenzene	1.661	1.463	11.9	111	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.134	-12.6	120	0.00
65 c	n-Nitrosodiphenylamine	0.676	0.664	1.8	117	0.00
66	4-Bromophenyl-phenylether	0.160	0.160	0.0	118	0.00
67	Hexachlorobenzene	0.166	0.166	0.0	119	0.00
68	Atrazine	0.188	0.185	1.6	116	0.00
69 C	Pentachlorophenol	0.101	0.109	-7.9	124	0.00
70	Phenanthrene	1.125	1.060	5.8	115	0.00
71	Anthracene	1.115	1.072	3.9	116	0.00
72	Carbazole	1.119	1.057	5.5	113	0.00
73	Di-n-butylphthalate	1.369	1.288	5.9	110	0.00
74 C	Fluoranthene	1.159	1.081	6.7	112	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
76	Benzidine	0.848	0.759	10.5	102	0.00
77	Pyrene	1.393	1.329	4.6	112	0.00
78 S	Terphenyl-d14	0.855	0.811	5.1	112	0.00
79	Butylbenzylphthalate	0.684	0.697	-1.9	113	0.00
80	Benzo(a)anthracene	1.182	1.124	4.9	113	0.00
81	3,3'-Dichlorobenzidine	0.389	0.391	-0.5	115	0.00
82	Chrysene	1.141	1.070	6.2	112	0.00
83	Bis(2-ethylhexyl)phthalate	0.926	0.940	-1.5	114	0.00
84 c	Di-n-octyl phthalate	1.509	1.542	-2.2	112	-0.01
85	Indeno(1,2,3-cd)pyrene	1.186	1.185	0.1	117	0.00
86 I	Perylene-d12	1.000	1.000	0.0	115	0.00
87	Benzo(b)fluoranthene	1.171	1.079	7.9	110	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG040715\
Data File : BG016274.D
Acq On : 8 Apr 2015 1:20
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 08 03:49:29 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 08 03:25:31 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.146	1.107	3.4	119	-0.01
89 C	Benzo(a)pyrene	1.098	1.046	4.7	114	-0.01
90	Dibenzo(a,h)anthracene	1.068	1.028	3.7	116	-0.01
91	Benzo(g,h,i)perylene	1.067	1.028	3.7	117	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0