

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040915\
 Data File : BG016291.D
 Acq On : 9 Apr 2015 16:15
 Operator : TP/IZ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 09 17:36:54 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 17:18:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	79373	20.00	ng	0.00
4) Naphthalene-d8	10.48	136	397693	20.00	ng	0.00
7) Acenaphthene-d10	14.34	164	240549	20.00	ng	0.00
10) Phenanthrene-d10	17.08	188	486960	20.00	ng	0.00
11) Chrysene-d12	21.25	240	403893	20.00	ng	0.00
13) Perylene-d12	23.50	264	379036	20.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	5.28	112	105126	10.27	ng	0.00
3) Phenol-d6	6.88	99	159937	10.27	ng	0.00
5) Nitrobenzene-d5	8.85	82	111496	9.88	ng	0.00
8) 2,4,6-Tribromophenol	15.83	330	27293	10.15	ng	0.00
9) 2-Fluorobiphenyl	12.96	172	261511	10.58	ng	0.00
12) Terphenyl-d14	19.70	244	321347	11.03	ng	0.00
Target Compounds						
6) Terpineol	10.54	59	4416	9.41	ng	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040915\
Data File : BG016291.D
Acq On : 9 Apr 2015 16:15
Operator : TP/IZ
Sample : SSTDICC010
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Quant Time: Apr 09 17:36:54 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 09 17:18:12 2015
Response via : Initial Calibration

