

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040915\
 Data File : BG016295.D
 Acq On : 9 Apr 2015 18:33
 Operator : TP/IZ
 Sample : SSTDICC060
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Quant Time: Apr 09 19:15:13 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 18:23:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	79878	20.00	ng	0.00
4) Naphthalene-d8	10.49	136	405813	20.00	ng	0.00
7) Acenaphthene-d10	14.34	164	241512	20.00	ng	0.00
10) Phenanthrene-d10	17.08	188	488422	20.00	ng	0.00
11) Chrysene-d12	21.25	240	401072	20.00	ng	0.00
13) Perylene-d12	23.50	264	374878	20.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	5.29	112	595304	57.32	ng	0.00
3) Phenol-d6	6.88	99	913352	56.92	ng	0.00
5) Nitrobenzene-d5	8.86	82	687877	59.39	ng	0.00
8) 2,4,6-Tribromophenol	15.83	330	178453	62.57	ng	0.00
9) 2-Fluorobiphenyl	12.96	172	1389909	54.22	ng	0.00
12) Terphenyl-d14	19.71	244	1582667	51.62	ng	0.01
Target Compounds						
6) Terpineol	10.54	59	27981	60.04	ng	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG040915\
Data File : BG016295.D
Acq On : 9 Apr 2015 18:33
Operator : TP/IZ
Sample : SSTDICC060
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC060

Quant Time: Apr 09 19:15:13 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 09 18:23:47 2015
Response via : Initial Calibration

