

Method Path : Z:\HPCHEM1\BNA_G\METHODS\
 Method File : 625-TERPINEOL-BG040915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 09 19:35:19 2015
 Response Via : Initial Calibration

Calibration Files

2.5 =BG016290.D 10 =BG016291.D 25 =BG016292.D 40 =BG016293.D 50 =BG016294.D 60 =BG016295.D 80 =BG016296.D

	Compound	2.5	10	25	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2) S	2-Fluorophenol	1.368	1.324	1.285	1.261	1.262	1.242	1.232	1.282	3.79
3) S	Phenol-d6	2.088	2.015	2.021	1.964	1.956	1.906	1.890	1.977	3.52
4) I	Naphthalene-d8	-----ISTD-----								
5) S	Nitrobenzene-d5	0.291	0.280	0.284	0.283	0.289	0.283	0.287	0.285	1.36
6)	Terpineol	0.023	0.022	0.023	0.024	0.024	0.023	0.024	0.023	3.21
7) I	Acenaphthene-d10	-----ISTD-----								
8) S	2,4,6-Tribromo...	0.101	0.113	0.123	0.125	0.127	0.123	0.125	0.120	7.72
9) S	2-Fluorobiphenyl	1.168	1.087	1.046	0.999	1.008	0.959	0.936	1.029	7.70
10) I	Phenanthrene-d10	-----ISTD-----								
11) I	Chrysene-d12	-----ISTD-----								
12) S	Terphenyl-d14	0.863	0.796	0.769	0.709	0.685	0.658	0.605	0.726	12.15
13) I	Perylene-d12	-----ISTD-----								

(#) = Out of Range