

Data Path : Z:\HPCHEM1\BNA G\Data\BG040915\
 Data File : BG016302.D
 Acq On : 9 Apr 2015 22:35
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 10 03:33:26 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 19:35:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 S	2-Fluorophenol	1.282	1.245	2.9	102	0.00
3 S	Phenol-d6	1.977	1.931	2.3	102	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
5 S	Nitrobenzene-d5	0.285	0.288	-1.1	102	0.00
6	Terpineol	0.023	0.023	0.0	98	0.00
7 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
8 S	2,4,6-Tribromophenol	0.120	0.125	-4.2	101	0.00
9 S	2-Fluorobiphenyl	1.029	0.993	3.5	101	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
11 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
12 S	Terphenyl-d14	0.726	0.706	2.8	103	0.00
13 I	Perylene-d12	1.000	1.000	0.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0