

Data Path : Z:\HPCHEM1\BNA G\Data\BG040915\
 Data File : BG016302.D
 Acq On : 9 Apr 2015 22:35
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 10 03:33:26 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\625-TERPINEOL-BG040915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 09 19:35:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2 S	2-Fluorophenol	80.000	77.695	2.9	102	0.00
3 S	Phenol-d6	80.000	78.120	2.3	102	0.00
4 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
5 S	Nitrobenzene-d5	80.000	80.910	-1.1	102	0.00
6	Terpineol	40.000	39.696	0.8	98	0.00
7 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
8 S	2,4,6-Tribromophenol	80.000	83.407	-4.3	101	0.00
9 S	2-Fluorobiphenyl	80.000	77.187	3.5	101	0.00
10 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
11 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
12 S	Terphenyl-d14	80.000	77.710	2.9	103	0.00
13 I	Perylene-d12	20.000	20.000	0.0	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0