

Data Path : Z:\HPCHEM1\BNA G\DATA\BG041017\
 Data File : BG026610.D
 Acq On : 10 Apr 2017 16:48
 Operator : SJ/MA
 Sample : 40 PPB(EPA-STD+SURR)
 Misc : For Confirmation
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Apr 11 03:25:35 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 03:24:57 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.03	152	179497	20.00	ng/ul	0.00
18) Naphthalene-d8	10.83	136	763112	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.64	164	462025	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.37	188	1095552	20.00	ng/ul	0.00
77) Chrysene-d12	21.62	240	1197043	20.00	ng/ul	0.00
85) Perylene-d12	24.79	264	1203142	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	7.20	99	558376	42.25	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.36	67	317463	42.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.57	132	486743	41.07	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	456579	42.88	ng/ul	0.00
19) Nitrobenzene-d5	9.19	128	246967	45.28	ng/ul	0.00
22) 2-Nitrophenol-d4	9.92	143	293251	45.88	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.46	165	494827	42.18	ng/ul	0.00
29) 4-Chloroaniline-d4	10.96	131	653016	55.93	ng/ul	0.00
43) Dimethylphthalate-d6	14.04	166	1505976	42.09	ng/ul	0.00
46) Acenaphthylene-d8	14.34	160	1843032	41.35	ng/ul	0.00
51) 4-Nitrophenol-d4	14.83	143	303113	43.55	ng/ul	0.00
57) Fluorene-d10	15.63	176	1331864	40.98	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.75	200	310120	44.02	ng/ul	0.00
70) Anthracene-d10	17.47	188	2018045	40.56	ng/ul	0.00
78) Pyrene-d10	19.75	212	2245887	42.06	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.58	264	2199721	42.55	ng/ul	0.01

Target Compounds

						Ovalue
6) Phenol	7.22	94	550926	39.91	ng/ul	93
8) Bis(2-Chloroethyl)ether	7.45	93	439817	41.75	ng/ul	94
10) 2-Chlorophenol	7.60	128	456932	39.85	ng/ul	98
11) 2-Methylphenol	8.47	108	437849	40.57	ng/ul	98
12) 2,2'-oxybis(1-Chloropropan	8.57	45	416738	41.95	ng/ul	99
14) Acetophenone	8.85	105	675701	42.25	ng/ul	93
15) N-Nitroso-di-n-propylamine	8.84	70	316548	42.63	ng/ul	95
16) 4-Methylphenol	8.80	108	468732	40.12	ng/ul	99
17) Hexachloroethane	9.12	117	169716	39.97	ng/ul	86
20) Nitrobenzene	9.24	77	451004	41.62	ng/ul	93
21) Isophorone	9.76	82	886762	42.98	ng/ul	93
23) 2-Nitrophenol	9.95	139	289917	43.12	ng/ul#	86
24) 2,4-Dimethylphenol	10.01	107	489693	40.55	ng/ul	93
25) Bis(2-Chloroethoxy)methane	10.23	93	587655	41.22	ng/ul	99
27) 2,4-Dichlorophenol	10.48	162	458668	39.59	ng/ul	97
28) Naphthalene	10.89	128	1452342	39.66	ng/ul	98
30) 4-Chloroaniline	10.99	127	596484	53.55	ng/ul	97
31) Hexachlorobutadiene	11.17	225	294030	39.88	ng/ul	98
33) 4-Chloro-3-methylphenol	12.10	107	455032	40.56	ng/ul	87
34) 2-Methylnaphthalene	12.48	142	1136870	40.14	ng/ul	99
36) 1,2,4,5-Tetrachlorobenzene	12.85	216	561573	38.61	ng/ul#	97
37) Hexachlorocyclopentadiene	12.83	237	298209	41.04	ng/ul	97
38) 2,4,6-Trichlorophenol	13.08	196	387483	40.15	ng/ul	99

Data Path : Z:\HPCHEM1\BNA G\DATA\BG041017\
 Data File : BG026610.D
 Acq On : 10 Apr 2017 16:48
 Operator : SJ/MA
 Sample : 40 PPB(EPA-STD+SURR)
 Misc : For Confirmation
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Apr 11 03:25:35 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 03:24:57 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 2,4,5-Trichlorophenol	13.15	196	396099	39.87	ng/ul	98
40) 1,1'-Biphenyl	13.48	154	1454533	39.24	ng/ul	98
41) 2-Chloronaphthalene	13.52	162	1094525	39.12	ng/ul	92
42) 2-Nitroaniline	13.72	65	282721	43.46	ng/ul	92
44) Dimethylphthalate	14.09	163	1369443	39.50	ng/ul#	97
45) 2,6-Dinitrotoluene	14.21	165	323272	43.20	ng/ul	87
47) Acenaphthylene	14.36	152	1723876	38.84	ng/ul	94
48) 3-Nitroaniline	14.54	138	356662	51.74	ng/ul#	91
49) Acenaphthene	14.71	153	1200610	38.98	ng/ul	99
50) 2,4-Dinitrophenol	14.75	184	180377	42.80	ng/ul#	80
52) 4-Nitrophenol	14.85	109	156224	42.30	ng/ul#	60
53) Dibenzofuran	15.04	168	1709936	39.04	ng/ul	91
54) 2,4-Dinitrotoluene	15.00	165	471159	43.55	ng/ul	90
55) 2,3,4,6-Tetrachlorophenol	15.26	232	382886	40.07	ng/ul#	90
56) Diethylphthalate	15.45	149	1403463	40.72	ng/ul	94
58) Fluorene	15.69	166	1407314	38.65	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.67	204	698769	39.07	ng/ul#	86
60) 4-Nitroaniline	15.70	138	379977	44.22	ng/ul#	77
63) 4,6-Dinitro-2-methylphenol	15.76	198	316299	42.78	ng/ul#	86
64) N-Nitrosodiphenylamine	15.89	169	1229193	39.22	ng/ul	95
65) 4-Bromophenyl-phenylether	16.56	248	478316	39.70	ng/ul#	85
66) Hexachlorobenzene	16.69	284	521081	40.34	ng/ul#	86
68) Pentachlorophenol	17.03	266	303935	39.54	ng/ul	99
69) Phenanthrene	17.42	178	2180355	37.98	ng/ul	94
71) Anthracene	17.51	178	2241885	38.17	ng/ul	93
74) Carbazole	17.77	167	2080288	41.20	ng/ul	95
75) Di-n-butylphthalate	18.32	149	2364432	39.72	ng/ul	96
76) Fluoranthene	19.42	202	2549225	41.56	ng/ul	96
79) Pyrene	19.78	202	2573907	38.62	ng/ul	95
80) Butylbenzylphthalate	20.65	149	1094300	41.51	ng/ul	94
81) 3,3'-Dichlorobenzidine	21.52	252	1051044	50.59	ng/ul	99
82) Benzo(a)anthracene	21.60	228	2514938	38.85	ng/ul	92
83) Bis(2-ethylhexyl)phthalate	21.50	149	1539345	40.86	ng/ul#	91
84) Chrysene	21.67	228	2314574	38.67	ng/ul	94
86) Di-n-octyl phthalate	22.68	149	2644798	42.73	ng/ul#	97
87) Benzo(b)fluoranthene	23.79	252	2661412	39.00	ng/ul	97
88) Benzo(k)fluoranthene	23.85	252	2635192	40.43	ng/ul	97
90) Benzo(a)pyrene	24.65	252	2642889	40.03	ng/ul	97
91) Indeno(1,2,3-cd)pyrene	28.42	276	3251425	40.84	ng/ul#	94
92) Dibenzo(a,h)anthracene	28.47	278	2743883	41.29	ng/ul	97
93) Benzo(g,h,i)perylene	29.55	276	2701374	40.80	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG041017\
Data File : BG026610.D
Acq On : 10 Apr 2017 16:48
Operator : SJ/MA
Sample : 40 PPB(EPA-STD+SURR)
Misc : For Confirmation
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :

Quant Time: Apr 11 03:25:35 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 11 03:24:57 2017
Response via : Initial Calibration





#6

Phenol

Concen: 39.91 ng/ul

RT: 7.22 min Scan# 703

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Instrument :

BNA_G

ClientSampleId :

Tot Ion: 94 Resp: 550926

Ion	Ratio	Lower	Upper
94	100		
65	25.6	21.0	31.6
66	30.0	29.3	43.9



Abundance



Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

7.22

Time-->



#8

Bis(2-Chloroethyl)ether

Concen: 41.75 ng/ul

RT: 7.45 min Scan# 742


Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Tgt Ion: 93 Resp: 439817

Ion	Ratio	Lower	Upper
93	100		
63	73.7	53.5	80.3
95	32.4	26.2	39.2



Abundance


Ion 93.00 (92.70 to 93.70): BGC


Ion 63.00 (62.70 to 63.70): BGC

Ion 95.00 (94.70 to 95.70): BGC

7.45

Time-->









#10
2-Chlorophenol
Concen: 39.85 ng/ul
RT: 7.60 min Scan# 768
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
128	100		
64	41.7	31.9	47.9
130	32.9	25.6	38.4



#11
2-Methylphenol
Concen: 40.57 ng/ul
RT: 8.47 min Scan# 916
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
108	100		
107	92.0	72.4	108.6









#12
2,2'-oxvbis(1-Chloropropane)
Concen: 41.95 ng/ul
RT: 8.57 min Scan# 932
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampled :


Tot Ion	Ratio	Lower	Upper
45	100		
77	19.3	15.9	23.9
79	14.9	12.5	18.7



#14
Acetophenone
Concen: 42.25 ng/ul
RT: 8.85 min Scan# 981
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
105	100		
77	67.4	59.0	88.4
51	18.1	16.4	24.6





#15

N-Nitroso-di-n-propylamine

Concen: 42.63 ng/ul

RT: 8.84 min Scan# 978

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :

Tot Ion: 70 Resp: 316548


Ion Ratio Lower Upper

70 100

42 40.5 35.4 53.0

101 11.4 8.4 12.6

130 25.3 18.2 27.4






Abundance Ion 70.00 (69.70 to 70.70): BG0

Ion 42.00 (41.70 to 42.70): BG0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#16

4-Methylphenol

Concen: 40.12 ng/ul

RT: 8.80 min Scan# 972

Delta R.T. -0.00 min

Lab File: BG026610.D


Acq: 10 Apr 2017 16:48

Tgt Ion: 108 Resp: 468732

Ion Ratio Lower Upper



108 100


107 113.9 90.3 135.5



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E





#17

Hexachloroethane

Concen: 39.97 ng/ul

RT: 9.12 min Scan# 1026

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :


Tot Ion:117 Resp: 169716

Ion Ratio Lower Upper




117 100

201 116.7 81.5 122.3

199 73.1 50.6 76.0



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 201.00 (200.70 to 201.70): E
Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 41.62 ng/ul

RT: 9.24 min Scan# 1046

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion: 77 Resp: 451004

Ion Ratio Lower Upper



77 100


123 55.0 39.0 58.4

65 12.8 10.4 15.6



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 123.00 (122.70 to 123.70): E
Ion 65.00 (64.70 to 65.70): BGC





#21

Isophorone

Concen: 42.98 ng/ul

RT: 9.76 min Scan# 1135

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 82 Resp: 886762

Ion Ratio Lower Upper

82 100

95 6.9 6.4 9.6

138 22.2 14.9 22.3



#23

2-Nitrophenol

Concen: 43.12 ng/ul

RT: 9.95 min Scan# 1167

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48




Tgt Ion: 139 Resp: 289917


Ion Ratio Lower Upper

139 100

65 39.0 36.3 54.5

109 21.7 26.4 39.6#









#24
 2,4-Dimethylphenol
 Concen: 40.55 ng/ul
 RT: 10.01 min Scan# 1177
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion:107 Resp: 489693
 Ion Ratio Lower Upper
 107 100
 121 55.1 40.6 61.0
 122 91.7 67.8 101.6




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E
 Ion 122.00 (121.70 to 122.70): E






#25
 Bis(2-Chloroethoxy)methane
 Concen: 41.22 ng/ul
 RT: 10.23 min Scan# 1216
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion: 93 Resp: 587655
 Ion Ratio Lower Upper
 93 100
 95 32.6 25.8 38.8
 123 12.6 9.8 14.8



Abundance Ion 93.00 (92.70 to 93.70): BGC
 Ion 95.00 (94.70 to 95.70): BGC
 Ion 123.00 (122.70 to 123.70): E









#27
2,4-Dichlorophenol
Concen: 39.59 ng/ul
RT: 10.48 min Scan# 1258
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Instrument :
BNA_G
ClientSampleId :

Tot Ion:162 Resp: 458668
Ion Ratio Lower Upper
162 100
164 65.9 52.1 78.1
98 31.5 28.4 42.6




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): B






#28
Naphthalene
Concen: 39.66 ng/ul
RT: 10.89 min Scan# 1327
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion:128 Resp: 1452342
Ion Ratio Lower Upper
128 100
129 11.5 8.8 13.2
127 14.5 10.5 15.7



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#30
 4-Chloroaniline
 Concen: 53.55 ng/ul
 RT: 10.99 min Scan# 1344
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:127 Resp: 596484
 Ion Ratio Lower Upper
 127 100
 129 33.5 25.4 38.2



#31
 Hexachlorobutadiene
 Concen: 39.88 ng/ul
 RT: 11.17 min Scan# 1375
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:225 Resp: 294030
 Ion Ratio Lower Upper
 225 100
 223 62.0 51.4 77.2
 227 62.7 50.7 76.1









#33
 4-Chloro-3-methylphenol
 Concen: 40.56 ng/ul
 RT: 12.10 min Scan# 1534
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
107	100		
144	29.0	20.4	30.6
142	91.3	62.5	93.7




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#34
 2-Methylnaphthalene
 Concen: 40.14 ng/ul
 RT: 12.48 min Scan# 1599
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.4	71.6	107.4



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 38.61 ng/ul
 RT: 12.85 min Scan# 1661
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
216	100		
214	79.9	62.5	93.7
179	21.2	18.2	27.2
108	13.7	14.2	21.2



#37
 Hexachlorocyclopentadiene
 Concen: 41.04 ng/ul
 RT: 12.83 min Scan# 1658
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
237	100		
235	65.0	50.4	75.6
272	13.7	9.6	14.4









#38
 2,4,6-Trichlorophenol
 Concen: 40.15 ng/ul
 RT: 13.08 min Scan# 1701
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:196 Resp: 387483
 Ion Ratio Lower Upper
 196 100
 198 96.7 76.6 114.8
 200 31.2 24.7 37.1



#39
 2,4,5-Trichlorophenol
 Concen: 39.87 ng/ul
 RT: 13.15 min Scan# 1713
 Delta R.T. -0.01 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:196 Resp: 396099
 Ion Ratio Lower Upper
 196 100
 198 98.0 76.9 115.3
 200 32.3 25.0 37.4









#40
 1,1'-Biphenyl
 Concen: 39.24 ng/ul
 RT: 13.48 min Scan# 1768
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion:154 Resp: 1454533




Ion	Ratio	Lower	Upper
154	100		
153	41.4	32.1	48.1
76	11.8	9.5	14.3




#41
 2-Chloronaphthalene
 Concen: 39.12 ng/ul
 RT: 13.52 min Scan# 1776
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:162 Resp: 1094525

Ion	Ratio	Lower	Upper
162	100		
164	34.4	25.9	38.9
127	35.7	34.1	51.1





#42

2-Nitroaniline

Concen: 43.46 ng/ul

RT: 13.72 min Scan# 1810

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :





Tot Ion: 65 Resp: 282721

Ion Ratio Lower Upper

65 100

92 76.2 60.4 90.6

138 138.1 99.8 149.6



#44

Dimethylphthalate

Concen: 39.50 ng/ul

RT: 14.09 min Scan# 1872

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48




Tgt Ion:163 Resp: 1369443


Ion Ratio Lower Upper

163 100

194 4.9 3.2 4.8#

164 11.2 7.8 11.8









#45
 2,6-Dinitrotoluene
 Concen: 43.20 ng/ul
 RT: 14.21 min Scan# 1893
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
165	100		
89	42.5	42.4	63.6
121	16.9	16.6	24.8



#47
 Acenaphthylene
 Concen: 38.84 ng/ul
 RT: 14.36 min Scan# 1919
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
152	100		
151	22.8	15.7	23.5
153	14.6	10.5	15.7









#48
 3-Nitroaniline
 Concen: 51.74 ng/ul
 RT: 14.54 min Scan# 1949
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:138 Resp: 356662
 Ion Ratio Lower Upper
 138 100
 108 8.5 8.7 13.1#
 92 100.3 88.2 132.4



#49
 Acenaphthene
 Concen: 38.98 ng/ul
 RT: 14.71 min Scan# 1977
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:153 Resp: 1200610
 Ion Ratio Lower Upper
 153 100
 152 49.0 38.3 57.5
 154 90.7 71.8 107.8









#50
2,4-Dinitrophenol
Concen: 42.80 ng/ul
RT: 14.75 min Scan# 1985
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
184	100		
63	44.8	46.5	69.7#
154	50.5	54.6	82.0#



#52
4-Nitrophenol
Concen: 42.30 ng/ul
RT: 14.85 min Scan# 2002
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
109	100		
139	191.3	103.7	155.5#
65	138.4	89.4	134.2#





#53

Dibenzofuran

Concen: 39.04 ng/ul

RT: 15.04 min Scan# 2034

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G


ClientSampleId :

Tot Ion:168 Resp: 1709936

Ion Ratio Lower Upper

168 100

139 34.7 32.2 48.2



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

1200000

1000000

800000

600000

400000

200000



0

15.00

15.04

15.10

Time-->



#54

2,4-Dinitrotoluene

Concen: 43.55 ng/ul

RT: 15.00 min Scan# 2027

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:165 Resp: 471159

Ion Ratio Lower Upper

165 100

63 30.5 29.8 44.6

182 2.9 2.2 3.4



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 182.00 (181.70 to 182.70): E

400000

300000

200000

100000


0


14.95

15.00

15.05

Time-->






#55
 2,3,4,6-Tetrachlorophenol
 Concen: 40.07 ng/ul
 RT: 15.26 min Scan# 2072
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Instrument :
 BNA_G
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
232	100		
131	40.1	39.4	59.2
130	2.0	2.2	3.2
166	29.9	26.6	39.8






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E


Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#56
 Diethylphthalate
 Concen: 40.72 ng/ul
 RT: 15.45 min Scan# 2103
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48



Tgt Ion	Ratio	Lower	Upper
149	100		
177	25.8	17.4	26.2
150	13.3	10.0	15.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E









#58
 Fluorene
 Concen: 38.65 ng/ul
 RT: 15.69 min Scan# 2144
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:166 Resp: 1407314
 Ion Ratio Lower Upper
 166 100
 165 98.9 78.5 117.7
 167 14.4 10.7 16.1



#59
 4-Chlorophenyl-phenylether
 Concen: 39.07 ng/ul
 RT: 15.67 min Scan# 2141
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:204 Resp: 698769
 Ion Ratio Lower Upper
 204 100
 206 34.0 26.2 39.4
 141 52.7 54.6 81.8#









#60
4-Nitroaniline
Concen: 44.22 ng/ul
RT: 15.70 min Scan# 2147
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampled :


Tot Ion	Ratio	Lower	Upper
138	100		
92	43.1	40.6	60.8
108	56.2	67.0	100.6



#63
4,6-Dinitro-2-methylphenol
Concen: 42.78 ng/ul
RT: 15.76 min Scan# 2157
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
198	100		
182	3.6	3.1	4.7
77	17.2	20.1	30.1





#64

N-Nitrosodiphenylamine

Concen: 39.22 ng/ul

RT: 15.89 min Scan# 2178

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Instrument :

BNA_G

ClientSampled :

Tot Ion:169 Resp: 1229193

Ion	Ratio	Lower	Upper
169	100		
168	71.3	53.8	80.8
167	38.9	29.3	43.9






Abundance

Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#65

4-Bromophenyl-phenylether

Concen: 39.70 ng/ul

RT: 16.56 min Scan# 2293


Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Tgt Ion:248 Resp: 478316

Ion	Ratio	Lower	Upper
248	100		
250	98.4	78.7	118.1
141	61.2	72.2	108.2






Abundance

Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#66
Hexachlorobenzene
Concen: 40.34 ng/ul
RT: 16.69 min Scan# 2315
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampled :


Tot Ion	Ratio	Lower	Upper
284	100		
142	28.4	33.1	49.7#
249	31.0	27.1	40.7



#68
Pentachlorophenol
Concen: 39.54 ng/ul
RT: 17.03 min Scan# 2372
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.4	50.5	75.7
268	64.9	50.9	76.3





#69

Phenanthrene

Concen: 37.98 ng/ul

RT: 17.42 min Scan# 2439

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Instrument :

BNA_G

ClientSampleId :

Tot Ion:178 Resp: 2180355

Ion	Ratio	Lower	Upper
178	100		
179	17.9	12.2	18.4
176	22.0	15.6	23.4






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 38.17 ng/ul

RT: 17.51 min Scan# 2454


Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Tgt Ion:178 Resp: 2241885

Ion	Ratio	Lower	Upper
178	100		
179	18.1	12.1	18.1
176	21.5	14.8	22.2






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#74

Carbazole

Concen: 41.20 ng/ul

RT: 17.77 min Scan# 2499

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Instrument :

BNA_G

ClientSampleId :

Tot Ion:167 Resp: 2080288




Ion	Ratio	Lower	Upper
167	100		
166	25.1	17.0	25.6
139	13.4	11.0	16.4



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#75

Di-n-butylphthalate

Concen: 39.72 ng/ul

RT: 18.32 min Scan# 2593


Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Tgt Ion:149 Resp: 2364432



Ion	Ratio	Lower	Upper
149	100		
150	11.0	7.4	11.0
104	4.7	4.0	6.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E









#76
 Fluoranthene
 Concen: 41.56 ng/ul
 RT: 19.42 min Scan# 2779
 Delta R.T. -0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48




Instrument :
 BNA_G
 ClientSampleId :


Tot Ion:202 Resp: 2549225
 Ion Ratio Lower Upper
 202 100
 101 12.5 8.8 13.2
 100 9.7 6.6 9.8



#79
 Pyrene
 Concen: 38.62 ng/ul
 RT: 19.78 min Scan# 2841
 Delta R.T. 0.00 min
 Lab File: BG026610.D
 Acq: 10 Apr 2017 16:48

Tgt Ion:202 Resp: 2573907
 Ion Ratio Lower Upper
 202 100
 101 14.7 10.2 15.2
 100 12.6 8.5 12.7





#80

Butylbenzylphthalate

Concen: 41.51 ng/ul

RT: 20.65 min Scan# 2989

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleID :


Tot Ion:149 Resp: 1094300

Ion	Ratio	Lower	Upper
149	100		
91	61.9	53.4	80.2
206	23.5	16.6	24.8

149 100

91 61.9 53.4 80.2




206 23.5 16.6 24.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E



#81

3,3'-Dichlorobenzidine

Concen: 50.59 ng/ul

RT: 21.52 min Scan# 3136

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:252 Resp: 1051044

Ion	Ratio	Lower	Upper
252	100		
254	65.9	52.2	78.2
126	10.1	9.4	14.2

252 100

254 65.9 52.2 78.2



126 10.1 9.4 14.2




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#82
Benzo(a)anthracene
Concen: 38.85 ng/ul
RT: 21.60 min Scan# 3151
Delta R.T. 0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48




Instrument :
BNA_G
ClientSampleId :


Tot Ion:228 Resp: 2514938
Ion Ratio Lower Upper
228 100
229 22.8 15.7 23.5
226 31.4 21.8 32.6



#83
Bis(2-ethylhexyl)phthalate
Concen: 40.86 ng/ul
RT: 21.50 min Scan# 3133
Delta R.T. -0.00 min
Lab File: BG026610.D
Acq: 10 Apr 2017 16:48

Tgt Ion:149 Resp: 1539345
Ion Ratio Lower Upper
149 100
167 31.9 21.6 32.4
279 5.2 3.0 4.4#





#84

Chrysene

Concen: 38.67 ng/ul

RT: 21.67 min Scan# 3162

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :


Tot Ion:228 Resp: 2314574

Ion Ratio Lower Upper

228 100

226 33.7 24.0 36.0




229 22.1 15.6 23.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#86

Di-n-octyl phthalate

Concen: 42.73 ng/ul

RT: 22.68 min Scan# 3335

Delta R.T. -0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:149 Resp: 2644798

Ion Ratio Lower Upper

149 100

167 1.9 1.2 1.8#



43 5.9 5.6 8.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BG0





#87

Benzo(b)fluoranthene

Concen: 39.00 ng/ul

RT: 23.79 min Scan# 3523

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :


Tot Ion:252 Resp: 2661412

Ion	Ratio	Lower	Upper
252	100		
253	24.0	17.5	26.3
125	10.8	8.5	12.7

252 100

253 24.0 17.5 26.3

125 10.8 8.5 12.7






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 40.43 ng/ul

RT: 23.85 min Scan# 3534

Delta R.T. 0.00 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:252 Resp: 2635192

Ion	Ratio	Lower	Upper
252	100		
253	23.4	17.3	25.9
125	11.0	8.2	12.2

252 100

253 23.4 17.3 25.9

125 11.0 8.2 12.2






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#90

Benzo(a)pyrene

Concen: 40.03 ng/ul

RT: 24.65 min Scan# 3669

Delta R.T. 0.01 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :


Tot Ion:252 Resp: 2642889

Ion Ratio Lower Upper

252 100

253 23.3 17.2 25.8




125 11.9 9.0 13.6



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Indeno(1,2,3-cd)pyrene

Concen: 40.84 ng/ul

RT: 28.42 min Scan# 4311

Delta R.T. 0.02 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:276 Resp: 3251425

Ion Ratio Lower Upper

276 100

138 21.8 22.3 33.5#



277 25.4 20.2 30.2




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#92

Dibenzo(a,h)anthracene

Concen: 41.29 ng/ul

RT: 28.47 min Scan# 4319

Delta R.T. 0.01 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48

Instrument :

BNA_G

ClientSampleId :


Tot Ion:278 Resp: 2743883

Ion	Ratio	Lower	Upper
278	100		
139	16.4	13.9	20.9
279	25.3	19.0	28.4

278 100

139 16.4 13.9 20.9




279 25.3 19.0 28.4



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#93

Benzo(g,h,i)perylene

Concen: 40.80 ng/ul

RT: 29.55 min Scan# 4503

Delta R.T. 0.02 min

Lab File: BG026610.D

Acq: 10 Apr 2017 16:48


Tgt Ion:276 Resp: 2701374

Ion	Ratio	Lower	Upper
276	100		
138	22.6	18.9	28.3
277	24.5	19.2	28.8

276 100

138 22.6 18.9 28.3

277 24.5 19.2 28.8



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

