

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041023\
 Data File : BG057040.D
 Acq On : 10 Apr 2023 11:19
 Operator : CG/JU
 Sample : 02224-13
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 OU3-FC-BACKFILL-07

Quant Time: Apr 11 04:11:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG033023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 08 04:45:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.404	152	15095	20.000	ng	0.00
21) Naphthalene-d8	11.247	136	56591	20.000	ng	# 0.00
39) Acenaphthene-d10	15.019	164	40770	20.000	ng	0.00
64) Phenanthrene-d10	17.762	188	96637	20.000	ng	0.00
76) Chrysene-d12	22.074	240	85401	20.000	ng	0.00
86) Perylene-d12	25.622	264	95349	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.919	112	53571	56.799	ng	0.00
7) Phenol-d6	7.535	99	102476	72.645	ng	0.00
23) Nitrobenzene-d5	9.579	82	71148	51.443	ng	0.00
42) 2,4,6-Tribromophenol	16.499	330	22554	34.628	ng	0.00
45) 2-Fluorobiphenyl	13.644	172	166001	55.212	ng	0.00
79) Terphenyl-d14	20.329	244	302082	59.855	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041023\
Data File : BG057040.D
Acq On : 10 Apr 2023 11:19
Operator : CG/JU
Sample : 02224-13
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
OU3-FC-BACKFILL-07

Quant Time: Apr 11 04:11:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 08 04:45:47 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.404 min Scan# 81
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19




Instrument :
BNA_G
ClientSampleId :
OU3-FC-BACKFILL-07


Tgt Ion:152 Resp: 15095
Ion Ratio Lower Upper
152 100
150 140.8 124.1 186.1
115 58.2 49.1 73.7



#5
2-Fluorophenol
Concen: 56.799 ng
RT: 5.919 min Scan# 405
Delta R.T. 0.002 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19

Tgt Ion:112 Resp: 53571
Ion Ratio Lower Upper
112 100
64 64.2 53.2 79.8
63 38.0 32.6 48.8









#7
Phenol-d6
Concen: 72.645 ng
RT: 7.535 min Scan# 61
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19




Instrument :
BNA_G
ClientSampleId :
OU3-FC-BACKFILL-07


Tgt Ion: 99 Resp: 102476
Ion Ratio Lower Upper
99 100
42 19.1 15.7 23.5
71 41.8 32.5 48.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.247 min Scan# 1312
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19

Tgt Ion: 136 Resp: 56591
Ion Ratio Lower Upper
136 100
137 11.9 8.6 13.0
54 7.2 6.9 10.3
68 5.8 3.8 5.6#









#23
Nitrobenzene-d5
Concen: 51.443 ng
RT: 9.579 min Scan# 1028
Delta R.T. -0.009 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19




Instrument :
BNA_G
ClientSampleId :
OU3-FC-BACKFILL-07


Tgt Ion: 82 Resp: 71148
Ion Ratio Lower Upper
82 100
128 36.8 27.9 41.9
54 49.2 39.4 59.2



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 15.019 min Scan# 1954
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19


Tgt Ion: 164 Resp: 40770
Ion Ratio Lower Upper
164 100
162 104.3 77.8 116.8
160 43.4 33.5 50.3








#42
2,4,6-Tribromophenol
Concen: 34.628 ng
RT: 16.499 min Scan# 21
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19

Instrument :
BNA_G
ClientSampleId :
OU3-FC-BACKFILL-07







Tgt Ion:330 Resp: 22554
Ion Ratio Lower Upper
330 100
332 95.9 76.1 114.1
141 29.4 25.8 38.6



#45
2-Fluorobiphenyl
Concen: 55.212 ng
RT: 13.644 min Scan# 1720
Delta R.T. -0.004 min
Lab File: BG057040.D
Acq: 10 Apr 2023 11:19

Tgt Ion:172 Resp: 166001
Ion Ratio Lower Upper
172 100
171 39.9 30.6 46.0
170 25.3 20.6 30.8





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.762 min Scan# 2421

Delta R.T. 0.002 min

Lab File: BG057040.D

Acq: 10 Apr 2023 11:19

Instrument :

BNA_G

ClientSampleId :

OU3-FC-BACKFILL-07





Tgt Ion:188 Resp: 96637

Ion Ratio Lower Upper

188 100

94 6.2 4.6 6.8

80 7.9 5.5 8.3



#76

Chrysene-d12

Concen: 20.000 ng

RT: 22.074 min Scan# 3155

Delta R.T. -0.009 min

Lab File: BG057040.D

Acq: 10 Apr 2023 11:19




Tgt Ion:240 Resp: 85401


Ion Ratio Lower Upper

240 100

120 5.3 3.8 5.6

236 28.2 21.0 31.6





#79

Terphenyl-d14

Concen: 59.855 ng

RT: 20.329 min Scan# 21

Delta R.T. 0.002 min

Lab File: BG057040.D

Acq: 10 Apr 2023 11:19

Instrument :

BNA_G

ClientSampleId :

OU3-FC-BACKFILL-07





Tgt Ion:244 Resp: 302082

Ion Ratio Lower Upper

244 100

212 7.1 6.1 9.1

122 5.6 5.0 7.4



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.622 min Scan# 3759

Delta R.T. -0.009 min

Lab File: BG057040.D

Acq: 10 Apr 2023 11:19

Tgt Ion:264 Resp: 95349

Ion Ratio Lower Upper

264 100

260 24.3 19.0 28.4

265 20.0 17.8 26.6

