

Data Path : Z:\HPCHEM1\BNA G\Data\BG041118\
 Data File : BG033765.D
 Acq On : 12 Apr 2018 4:11
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 07:23:38 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 10 16:51:36 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.542	0.563	-3.9	116	0.00
3	Pyridine	1.497	1.551	-3.6	104	0.00
4	n-Nitrosodimethylamine	0.614	0.682	-11.1	122	0.00
5 S	2-Fluorophenol	1.067	1.180	-10.6	113	0.00
6	Aniline	2.155	2.335	-8.4	111	0.00
7 S	Phenol-d6	1.833	1.910	-4.2	112	0.00
8	2-Chlorophenol	1.219	1.294	-6.2	108	0.00
9	Benzaldehyde	1.165	0.956	17.9	93	0.00
10 C	Phenol	1.809	1.919	-6.1	111	0.00
11	bis(2-Chloroethyl)ether	1.477	1.521	-3.0	105	0.00
12	1,3-Dichlorobenzene	1.594	1.551	2.7	106	0.00
13 C	1,4-Dichlorobenzene	1.597	1.463	8.4	101	0.00
14	1,2-Dichlorobenzene	1.558	1.575	-1.1	107	0.00
15	Benzyl Alcohol	1.443	1.539	-6.7	110	0.00
16	2,2'-oxybis(1-Chloropropane	2.408	2.604	-8.1	117	0.00
17	2-Methylphenol	1.233	1.294	-4.9	113	0.00
18	Hexachloroethane	0.616	0.640	-3.9	114	0.00
19 P	n-Nitroso-di-n-propylamine	1.343	1.439	-7.1	109	0.00
20	3+4-Methylphenols	1.755	1.879	-7.1	110	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
22	Acetophenone	0.548	0.540	1.5	101	0.00
23 S	Nitrobenzene-d5	0.435	0.426	2.1	104	0.00
24	Nitrobenzene	0.466	0.457	1.9	104	0.00
25	Isophorone	0.845	0.843	0.2	106	0.00
26 C	2-Nitrophenol	0.176	0.178	-1.1	102	0.00
27	2,4-Dimethylphenol	0.256	0.259	-1.2	113	0.00
28	bis(2-Chloroethoxy)methane	0.422	0.408	3.3	102	0.00
29 C	2,4-Dichlorophenol	0.315	0.321	-1.9	107	0.00
30	1,2,4-Trichlorobenzene	0.409	0.390	4.6	103	0.00
31	Naphthalene	1.036	1.006	2.9	105	0.00
32	Benzoic acid	0.238	0.200	16.0	103	0.01
33	4-Chloroaniline	0.464	0.440	5.2	100	0.00
34 C	Hexachlorobutadiene	0.310	0.278	10.3	100	0.00
35	Caprolactam	0.123	0.127	-3.3	108	0.00
36 C	4-Chloro-3-methylphenol	0.411	0.409	0.5	102	0.00
37	2-Methylnaphthalene	0.804	0.769	4.4	106	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
39	1,2,4,5-Tetrachlorobenzene	0.724	0.668	7.7	98	0.00
40 P	Hexachlorocyclopentadiene	0.425	0.423	0.5	104	0.00
41 S	2,4,6-Tribromophenol	0.299	0.279	6.7	97	0.00
42 C	2,4,6-Trichlorophenol	0.421	0.399	5.2	96	0.00
43	2,4,5-Trichlorophenol	0.498	0.488	2.0	96	0.00
44 S	2-Fluorobiphenyl	1.385	1.328	4.1	101	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG041118\
 Data File : BG033765.D
 Acq On : 12 Apr 2018 4:11
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 07:23:38 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 10 16:51:36 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.537	1.474	4.1	101	0.00
46	2-Chloronaphthalene	1.185	1.145	3.4	102	0.00
47	2-Nitroaniline	0.444	0.476	-7.2	110	0.00
48	Acenaphthylene	1.978	1.929	2.5	103	0.00
49	Dimethylphthalate	1.741	1.711	1.7	104	0.00
50	2,6-Dinitrotoluene	0.356	0.361	-1.4	103	0.00
51 C	Acenaphthene	1.275	1.299	-1.9	106	0.00
52	3-Nitroaniline	0.373	0.351	5.9	98	0.00
53 P	2,4-Dinitrophenol	0.149	0.185	-24.2	122	0.00
54	Dibenzofuran	1.883	1.826	3.0	103	0.00
55 P	4-Nitrophenol	0.262	0.248	5.3	97	0.00
56	2,4-Dinitrotoluene	0.524	0.525	-0.2	105	0.00
57	Fluorene	1.604	1.567	2.3	105	0.00
58	2,3,4,6-Tetrachlorophenol	0.468	0.421	10.0	92	0.00
59	Diethylphthalate	1.690	1.678	0.7	106	0.00
60	4-Chlorophenyl-phenylether	0.922	0.868	5.9	102	0.00
61	4-Nitroaniline	0.378	0.387	-2.4	107	0.00
62	Azobenzene	1.637	1.639	-0.1	106	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
64	4,6-Dinitro-2-methylphenol	0.120	0.119	0.8	104	0.00
65 c	n-Nitrosodiphenylamine	0.571	0.544	4.7	105	0.00
66	4-Bromophenyl-phenylether	0.235	0.212	9.8	98	0.00
67	Hexachlorobenzene	0.248	0.232	6.5	104	0.00
68	Atrazine	0.231	0.212	8.2	100	0.00
69 C	Pentachlorophenol	0.170	0.137	19.4	88	0.00
70	Phenanthrene	1.042	0.981	5.9	104	0.00
71	Anthracene	1.055	1.024	2.9	107	0.00
72	Carbazole	0.935	0.911	2.6	106	0.00
73	Di-n-butylphthalate	1.088	1.097	-0.8	110	0.00
74 C	Fluoranthene	1.289	1.248	3.2	107	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
76	Benzidine	0.542	0.490	9.6	94	0.00
77	Pyrene	1.334	1.262	5.4	106	0.00
78 S	Terphenyl-d14	0.935	0.873	6.6	106	0.00
79	Butylbenzylphthalate	0.492	0.490	0.4	111	0.00
80	Benzo(a)anthracene	1.274	1.213	4.8	108	0.00
81	3,3'-Dichlorobenzidine	0.453	0.453	0.0	108	0.00
82	Chrysene	1.233	1.183	4.1	109	0.00
83	Bis(2-ethylhexyl)phthalate	0.679	0.704	-3.7	115	0.00
84 c	Di-n-octyl phthalate	1.039	1.134	-9.1	120	0.00
85	Indeno(1,2,3-cd)pyrene	1.333	1.360	-2.0	113	0.00
86 I	Perylene-d12	1.000	1.000	0.0	120	0.00
87	Benzo(b)fluoranthene	1.155	1.077	6.8	110	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG041118\
Data File : BG033765.D
Acq On : 12 Apr 2018 4:11
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 12 07:23:38 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 10 16:51:36 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.169	1.094	6.4	110	0.00
89 C	Benzo(a)pyrene	1.110	1.051	5.3	111	0.00
90	Dibenzo(a,h)anthracene	1.045	1.042	0.3	114	0.00
91	Benzo(a,h,i)perylene	1.035	1.001	3.3	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0