

Data Path : Z:\HPCHEM1\BNA G\DATA\BG041118\
 Data File : BG033767.D
 Acq On : 12 Apr 2018 5:42
 Operator : SJ/JU
 Sample : J2321-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 041018-JETB-E-U-M

Quant Time: Apr 12 07:31:25 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 10 16:51:36 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.86	152	26692	20.00	ng	0.00
21) Naphthalene-d8	11.74	136	118011	20.00	ng	0.00
38) Acenaphthene-d10	15.47	164	92752	20.00	ng	0.00
63) Phenanthrene-d10	18.20	188	264100	20.00	ng	0.00
75) Chrysene-d12	22.55	240	286539	20.00	ng	0.00
86) Perylene-d12	26.31	264	303229	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.28	112	58858	41.35	ng	0.00
7) Phenol-d6	7.97	99	53819	22.00	ng	0.00
23) Nitrobenzene-d5	10.06	82	237787	92.58	ng	0.00
41) 2,4,6-Tribromophenol	16.95	330	123208	88.82	ng	0.00
44) 2-Fluorobiphenyl	14.10	172	569123	88.58	ng	0.00
78) Terphenyl-d14	20.73	244	1269263	94.73	ng	0.00
Target Compounds						
10) Phenol	8.00	94	7373	3.053	ng	80
49) Dimethylphthalate	14.89	163	33371	4.134	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG041118\
Data File : BG033767.D
Acq On : 12 Apr 2018 5:42
Operator : SJ/JU
Sample : J2321-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
041018-JETB-E-U-M

Quant Time: Apr 12 07:31:25 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 10 16:51:36 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.86 min Scan# 948
 Delta R.T. 0.00 min
 Lab File: BG033767.D
 Acq: 12 Apr 2018 5:42

Instrument :
 BNA_G
 ClientSampleId :
 041018-JETB-E-U-M

Tot Ion	Ratio	Lower	Upper
152	100		
150	142.8	126.6	189.8
115	60.5	53.0	79.4




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5




2-Fluorophenol
 Concen: 41.348 ng
 RT: 6.28 min Scan# 510
 Delta R.T. 0.00 min
 Lab File: BG033767.D
 Acq: 12 Apr 2018 5:42

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.8	56.3	84.5
63	36.5	30.1	45.1



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BG
 Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 22.004 ng

RT: 7.97 min Scan# 797

Delta R.T. 0.00 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42

Instrument :

BNA_G

ClientSampleId :

041018-JETB-E-U-M





Tot Ion: 99 Resp: 53819

Ion Ratio Lower Upper

99 100

42 20.2 14.1 21.1

71 48.2 26.1 39.1#



#10

Phenol

Concen: 3.053 ng

RT: 8.00 min Scan# 802

Delta R.T. 0.00 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42


Tgt Ion: 94 Resp: 7373


Ion Ratio Lower Upper

94 100

65 41.8 11.9 51.9

66 55.9 22.2 62.2





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 11.74 min Scan# 1439

Delta R.T. 0.01 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42

Instrument :

BNA_G

ClientSampleId :

041018-JETB-E-U-M

Tot Ion:136 Resp: 118011


Ion Ratio Lower Upper

136 100

137 10.0 9.2 13.8

54 9.9 10.3 15.5#

68 5.8 4.5 6.7






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC



#23

Nitrobenzene-d5

Concen: 92.575 ng

RT: 10.06 min Scan# 1152

Delta R.T. 0.00 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42


Tgt Ion: 82 Resp: 237787

Ion Ratio Lower Upper

82 100

128 36.3 29.7 44.5



54 57.5 43.1 64.7




Abundance Ion 82.00 (81.70 to 82.70): BGC

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BGC









#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 15.47 min Scan# 2074
Delta R.T. 0.00 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42

Instrument :
BNA_G
ClientSampleId :
041018-JETB-E-U-M

Tot Ion:164 Resp: 92752
Ion Ratio Lower Upper
164 100
162 99.9 78.8 118.2
160 45.7 35.4 53.0




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#41
2,4,6-Tribromophenol
Concen: 88.817 ng
RT: 16.95 min Scan# 2325
Delta R.T. 0.00 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42

Tgt Ion:330 Resp: 123208
Ion Ratio Lower Upper
330 100
332 97.6 77.0 115.6
141 38.9 25.2 37.8#



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#44
2-Fluorobiphenyl
Concen: 88.581 ng
RT: 14.10 min Scan# 1840
Delta R.T. -0.01 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42




Instrument :
BNA_G
ClientSampleId :
041018-JETB-E-U-M


Tot Ion:172 Resp: 569123
Ion Ratio Lower Upper
172 100
171 37.7 30.0 45.0
170 25.3 19.5 29.3



#49
Dimethylphthalate
Concen: 4.134 ng
RT: 14.89 min Scan# 1975
Delta R.T. -0.01 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42

Tgt Ion:163 Resp: 33371
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 7.8 8.2 12.4#









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 18.20 min Scan# 2538
Delta R.T. -0.01 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42

Instrument :
BNA_G
ClientSampleId :
041018-JETB-E-U-M

Tot Ion	Ratio	Lower	Upper
188	100		
94	7.1	6.9	10.3
80	8.1	7.7	11.5




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#75
Chrysene-d12
Concen: 20.000 ng
RT: 22.55 min Scan# 3278
Delta R.T. -0.01 min
Lab File: BG033767.D
Acq: 12 Apr 2018 5:42

Tgt Ion	Ratio	Lower	Upper
240	100		
120	6.6	6.8	10.2#
236	26.6	20.9	31.3



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 94.732 ng

RT: 20.73 min Scan# 2968

Delta R.T. 0.00 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42

Instrument :

BNA_G

ClientSampleId :

041018-JETB-E-U-M


Tot Ion:244 Resp: 1269263

Ion Ratio Lower Upper

244 100

212 8.4 5.8 8.8




122 7.3 7.0 10.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 26.31 min Scan# 3918

Delta R.T. 0.00 min

Lab File: BG033767.D

Acq: 12 Apr 2018 5:42


Tgt Ion:264 Resp: 303229

Ion Ratio Lower Upper

264 100

260 21.4 17.4 26.0

265 21.5 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

