

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041122\
 Data File : BG053102.D
 Acq On : 11 Apr 2022 15:38
 Operator : CG/JU
 Sample : N2340-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-04-COMP

Quant Time: Apr 12 06:07:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.120	152	30281	20.000	ng	# 0.00
21) Naphthalene-d8	10.934	136	121887	20.000	ng	0.00
39) Acenaphthene-d10	14.741	164	95108	20.000	ng	0.00
64) Phenanthrene-d10	17.484	188	231177	20.000	ng	0.00
76) Chrysene-d12	21.766	240	246214	20.000	ng	0.00
86) Perylene-d12	25.068	264	249866	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.682	112	193619	109.606	ng	0.00
7) Phenol-d6	7.280	99	285394	104.765	ng	0.00
23) Nitrobenzene-d5	9.283	82	200669	66.848	ng	0.00
42) 2,4,6-Tribromophenol	16.227	330	158834	104.970	ng	0.00
45) 2-Fluorobiphenyl	13.366	172	460069	66.702	ng	0.00
79) Terphenyl-d14	20.086	244	930421	63.822	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041122\
Data File : BG053102.D
Acq On : 11 Apr 2022 15:38
Operator : CG/JU
Sample : N2340-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-04-COMP

Quant Time: Apr 12 06:07:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration

