

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041122\
 Data File : BG053094.D
 Acq On : 11 Apr 2022 10:18
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 06:06:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	42.765	-6.9	112	0.00
3	Pyridine	40.000	44.086	-10.2	109	0.00
4	n-Nitrosodimethylamine	40.000	42.581	-6.5	107	0.00
5 S	2-Fluorophenol	80.000	81.797	-2.2	106	0.00
6	Aniline	40.000	41.712	-4.3	106	0.00
7 S	Phenol-d6	80.000	82.607	-3.3	105	0.00
8	2-Chlorophenol	40.000	40.076	-0.2	102	0.00
9	Benzaldehyde	40.000	37.977	5.1	110	0.00
10 C	Phenol	40.000	40.999	-2.5	107	0.00
11	bis(2-Chloroethyl)ether	40.000	42.139	-5.3	110	0.00
12	1,3-Dichlorobenzene	40.000	40.535	-1.3	105	0.00
13 C	1,4-Dichlorobenzene	40.000	40.407	-1.0	104	0.00
14	1,2-Dichlorobenzene	40.000	39.349	1.6	104	0.00
15	Benzyl Alcohol	40.000	40.723	-1.8	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.857	-2.1	103	0.00
17	2-Methylphenol	40.000	40.909	-2.3	106	0.00
18	Hexachloroethane	40.000	40.855	-2.1	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.546	-1.4	103	0.00
20	3+4-Methylphenols	40.000	41.664	-4.2	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	41.306	-3.3	102	0.00
23 S	Nitrobenzene-d5	80.000	82.894	-3.6	103	0.00
24	Nitrobenzene	40.000	41.434	-3.6	105	0.00
25	Isophorone	40.000	41.542	-3.9	101	0.00
26 C	2-Nitrophenol	40.000	39.839	0.4	97	0.00
27	2,4-Dimethylphenol	40.000	43.091	-7.7	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.447	-6.1	104	0.00
29 C	2,4-Dichlorophenol	40.000	42.105	-5.3	102	0.00
30	1,2,4-Trichlorobenzene	40.000	41.233	-3.1	102	0.00
31	Naphthalene	40.000	41.737	-4.3	103	0.00
32	Benzoic acid	40.000	40.155	-0.4	95	0.00
33	4-Chloroaniline	40.000	42.136	-5.3	102	0.00
34 C	Hexachlorobutadiene	40.000	41.698	-4.2	104	0.00
35	Caprolactam	40.000	41.000	-2.5	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.789	-7.0	104	0.00
37	2-Methylnaphthalene	40.000	41.484	-3.7	103	0.00
38	1-Methylnaphthalene	40.000	41.529	-3.8	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.221	-0.6	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.794	-2.0	99	0.00
42 S	2,4,6-Tribromophenol	80.000	79.374	0.8	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.873	0.3	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.061	-0.2	99	0.00
45 S	2-Fluorobiphenyl	80.000	80.428	-0.5	102	0.00
46	1,1'-Biphenyl	40.000	40.407	-1.0	103	0.00
47	2-Chloronaphthalene	40.000	40.055	-0.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041122\
 Data File : BG053094.D
 Acq On : 11 Apr 2022 10:18
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 12 06:06:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.641	-1.6	98	0.00
49	Acenaphthylene	40.000	40.506	-1.3	101	0.00
50	Dimethylphthalate	40.000	40.338	-0.8	101	0.00
51	2,6-Dinitrotoluene	40.000	39.983	0.0	100	0.00
52 C	Acenaphthene	40.000	40.684	-1.7	101	0.00
53	3-Nitroaniline	40.000	40.592	-1.5	100	0.00
54 P	2,4-Dinitrophenol	40.000	39.856	0.4	96	0.00
55	Dibenzofuran	40.000	40.787	-2.0	104	0.00
56 P	4-Nitrophenol	40.000	43.173	-7.9	102	0.00
57	2,4-Dinitrotoluene	40.000	41.302	-3.3	101	0.00
58	Fluorene	40.000	40.497	-1.2	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.703	-1.8	100	0.00
60	Diethylphthalate	40.000	39.471	1.3	99	0.00
61	4-Chlorophenyl-phenylether	40.000	40.298	-0.7	101	0.00
62	4-Nitroaniline	40.000	40.371	-0.9	102	0.00
63	Azobenzene	40.000	40.812	-2.0	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.583	1.0	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.653	0.9	100	0.00
67	4-Bromophenyl-phenylether	40.000	40.883	-2.2	101	0.00
68	Hexachlorobenzene	40.000	40.109	-0.3	100	0.00
69	Atrazine	40.000	38.346	4.1	98	0.00
70 C	Pentachlorophenol	40.000	42.655	-6.6	99	0.00
71	Phenanthrene	40.000	39.969	0.1	101	0.00
72	Anthracene	40.000	39.883	0.3	101	0.00
73	Carbazole	40.000	40.473	-1.2	102	0.00
74	Di-n-butylphthalate	40.000	40.027	-0.1	101	0.00
75 C	Fluoranthene	40.000	40.675	-1.7	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	43.102	-7.8	119	0.00
78	Pyrene	40.000	40.352	-0.9	103	0.00
79 S	Terphenyl-d14	80.000	78.965	1.3	102	0.00
80	Butylbenzylphthalate	40.000	40.532	-1.3	103	0.00
81	Benzo(a)anthracene	40.000	40.909	-2.3	105	0.00
82	3,3'-Dichlorobenzidine	40.000	40.966	-2.4	106	0.00
83	Chrysene	40.000	40.952	-2.4	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.533	-3.8	108	0.00
85 c	Di-n-octyl phthalate	40.000	41.209	-3.0	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	110	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.739	-1.8	109	0.00
88	Benzo(b)fluoranthene	40.000	41.467	-3.7	111	0.00
89	Benzo(k)fluoranthene	40.000	38.962	2.6	105	-0.01
90 C	Benzo(a)pyrene	40.000	40.840	-2.1	109	0.00
91	Dibenzo(a,h)anthracene	40.000	40.298	-0.7	108	-0.01
92	Benzo(g,h,i)perylene	40.000	40.596	-1.5	109	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041122\
Data File : BG053094.D
Acq On : 11 Apr 2022 10:18
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 12 06:06:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0