



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041224\
Data File : BG060871.D
Acq On : 12 Apr 2024 11:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Mar 30 03:12:26 2024






AutoFind: Scans 2518, 2519, 2520; Background Corrected with Scan 2510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.1	56939	PASS
68	69	0.00	2	1.5	606	PASS
69	198	0.00	100	30.9	40854	PASS
70	69	0.00	2	0.7	267	PASS
127	198	10	80	38.9	51336	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	132053	PASS
199	198	5	9	7.4	9717	PASS
275	198	10	60	31.4	41459	PASS
365	198	1	100	5.5	7265	PASS
441	198	0.01	100	17.5	23085	PASS
442	442	50	100	100.0	144101	PASS
443	442	15	24	19.5	28099	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041224\
Data File : BG060871.D
Acq On : 12 Apr 2024 11:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 12 12:11:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration



TIC: BG060871.D\data.ms

(70) Pentachlorophenol (C)

16.983min (+ 0.003) 22983.25 ng




response 174902

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.30	65.97
264.00	64.70	63.64
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041224\
Data File : BG060871.D
Acq On : 12 Apr 2024 11:02
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Apr 12 12:11:01 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration



TIC: BG060871.D\data.ms

(77) Benzidine**19.568min (-0.003) 99031.22 ng****response 904874**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	13.98
183.00	11.90	12.73
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/12/2024	BNA_G	<u>BG060871.D</u>
Compound Name	Response	Retention Time
DDT	475223	20.814
DDD	14707	20.367
DDE	2525	19.868
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
17232	492455	3.50

Instrument :
BNA_G
ClientSampleId :
DFTPP