


Data Path : Z:\HPCHEM1\BNA\_G\Data\BG041315\  
 Data File : BG016376.D  
 Acq On : 14 Apr 2015 1:24  
 Operator : TP/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG040615.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Apr 14 14:00:56 2015



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2514

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.8      | 56965   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 44.5      | 65234   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 317     | PASS             |
| 127         | 198          | 10           | 80           | 57.4      | 84197   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 146717  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 9963    | PASS             |
| 275         | 198          | 10           | 60           | 20.6      | 30285   | PASS             |
| 365         | 198          | 1            | 100          | 2.1       | 3012    | PASS             |
| 441         | 442          | 0.01         | 24           | 19.2      | 17531   | PASS             |
| 442         | 198          | 50           | 100          | 62.3      | 91333   | PASS             |
| 443         | 442          | 15           | 24           | 19.5      | 17848   | PASS             |