




Data Path : Z:\HPCHEM1\BNA_G\DATA\BG041316\
Data File : BG021627.D
Acq On : 13 Apr 2016 18:53
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 14 04:40:33 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG041316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 15:48:13 2016
Response via : Initial Calibration



TIC: BG021627.D

(69) Pentachlorophenol (C)




17.198min (-0.001) 0.00ng

response 101548

Ion	Exp%	Act%
265.70	100	100
268.00	68.10	62.26
264.00	67.20	62.19
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG041316\
Data File : BG021627.D
Acq On : 13 Apr 2016 18:53
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 14 04:40:33 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG041316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 15:48:13 2016
Response via : Initial Calibration



TIC: BG021627.D

(76) Benzidine

19.772min (+0.005) 15957.26ng

response 664270

Ion	Exp%	Act%
184.00	100	100
185.00	14.40	16.21
183.00	12.80	14.25
0.00	0.00	0.00