

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041322\
 Data File : BG053131.D
 Acq On : 13 Apr 2022 10:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 02:55:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.557	0.571	-2.5	107	0.00
3	Pyridine	1.470	1.670	-13.6	112	0.00
4	n-Nitrosodimethylamine	0.728	0.782	-7.4	107	0.00
5 S	2-Fluorophenol	1.167	1.293	-10.8	114	0.00
6	Aniline	2.085	2.246	-7.7	109	0.00
7 S	Phenol-d6	1.799	1.960	-8.9	110	0.00
8	2-Chlorophenol	1.260	1.351	-7.2	109	0.00
9	Benzaldehyde	1.075	1.069	0.6	114	0.00
10 C	Phenol	1.791	1.921	-7.3	112	0.00
11	bis(2-Chloroethyl)ether	1.291	1.388	-7.5	112	0.00
12	1,3-Dichlorobenzene	1.464	1.574	-7.5	111	0.00
13 C	1,4-Dichlorobenzene	1.492	1.607	-7.7	111	0.00
14	1,2-Dichlorobenzene	1.439	1.512	-5.1	110	0.00
15	Benzyl Alcohol	1.599	1.706	-6.7	107	0.00
16	2,2'-oxybis(1-Chloropropane	2.155	2.244	-4.1	105	0.00
17	2-Methylphenol	1.212	1.298	-7.1	111	0.00
18	Hexachloroethane	0.560	0.575	-2.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.317	1.389	-5.5	107	0.00
20	3+4-Methylphenols	1.711	1.827	-6.8	108	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.555	0.583	-5.0	108	0.00
23 S	Nitrobenzene-d5	0.493	0.504	-2.2	105	0.00
24	Nitrobenzene	0.479	0.487	-1.7	107	0.00
25	Isophorone	0.837	0.870	-3.9	105	0.00
26 C	2-Nitrophenol	0.200	0.205	-2.5	104	0.00
27	2,4-Dimethylphenol	0.287	0.304	-5.9	107	0.00
28	bis(2-Chloroethoxy)methane	0.470	0.487	-3.6	105	0.00
29 C	2,4-Dichlorophenol	0.361	0.380	-5.3	106	0.00
30	1,2,4-Trichlorobenzene	0.408	0.421	-3.2	106	0.00
31	Naphthalene	1.067	1.099	-3.0	105	0.00
32	Benzoic acid	0.155	0.212	-36.8#	113	0.00
33	4-Chloroaniline	0.457	0.468	-2.4	103	0.00
34 C	Hexachlorobutadiene	0.303	0.316	-4.3	108	0.00
35	Caprolactam	0.127	0.128	-0.8	101	0.00
36 C	4-Chloro-3-methylphenol	0.420	0.439	-4.5	105	0.00
37	2-Methylnaphthalene	0.790	0.808	-2.3	105	0.00
38	1-Methylnaphthalene	0.771	0.777	-0.8	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.679	0.701	-3.2	106	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.366	-4.3	103	0.00
42 S	2,4,6-Tribromophenol	0.318	0.320	-0.6	101	0.00
43 C	2,4,6-Trichlorophenol	0.468	0.487	-4.1	104	0.00
44	2,4,5-Trichlorophenol	0.499	0.509	-2.0	102	0.00
45 S	2-Fluorobiphenyl	1.450	1.476	-1.8	105	0.00
46	1,1'-Biphenyl	1.440	1.463	-1.6	105	0.00
47	2-Chloronaphthalene	1.149	1.162	-1.1	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041322\
 Data File : BG053131.D
 Acq On : 13 Apr 2022 10:08
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 02:55:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.436	0.438	-0.5	98	0.00
49	Acenaphthylene	1.799	1.827	-1.6	103	0.00
50	Dimethylphthalate	1.600	1.597	0.2	102	0.00
51	2,6-Dinitrotoluene	0.334	0.329	1.5	100	0.00
52 C	Acenaphthene	1.220	1.249	-2.4	103	0.00
53	3-Nitroaniline	0.353	0.350	0.8	99	0.00
54 P	2,4-Dinitrophenol	0.212	0.231	-9.0	106	0.00
55	Dibenzofuran	1.910	1.918	-0.4	104	0.00
56 P	4-Nitrophenol	0.269	0.284	-5.6	101	0.00
57	2,4-Dinitrotoluene	0.475	0.477	-0.4	100	0.00
58	Fluorene	1.543	1.550	-0.5	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.485	0.490	-1.0	101	0.00
60	Diethylphthalate	1.676	1.618	3.5	98	0.00
61	4-Chlorophenyl-phenylether	0.875	0.873	0.2	102	0.00
62	4-Nitroaniline	0.372	0.376	-1.1	103	0.00
63	Azobenzene	1.690	1.691	-0.1	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.140	0.152	-8.6	104	0.00
66 c	n-Nitrosodiphenylamine	0.575	0.602	-4.7	102	0.00
67	4-Bromophenyl-phenylether	0.256	0.264	-3.1	99	0.00
68	Hexachlorobenzene	0.278	0.289	-4.0	101	0.00
69	Atrazine	0.225	0.220	2.2	97	0.00
70 C	Pentachlorophenol	0.152	0.169	-11.2	101	0.00
71	Phenanthrene	1.092	1.147	-5.0	103	0.00
72	Anthracene	1.082	1.128	-4.3	102	0.00
73	Carbazole	1.052	1.104	-4.9	103	0.00
74	Di-n-butylphthalate	1.193	1.218	-2.1	100	0.00
75 C	Fluoranthene	1.400	1.459	-4.2	103	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
77	Benzidine	0.571	0.546	4.4	106	0.00
78	Pyrene	1.424	1.425	-0.1	103	0.00
79 S	Terphenyl-d14	1.184	1.183	0.1	104	0.00
80	Butylbenzylphthalate	0.531	0.541	-1.9	104	0.00
81	Benzo(a)anthracene	1.367	1.399	-2.3	106	0.00
82	3,3'-Dichlorobenzidine	0.503	0.505	-0.4	104	0.00
83	Chrysene	1.269	1.300	-2.4	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.718	0.739	-2.9	107	0.00
85 c	Di-n-octyl phthalate	1.201	1.254	-4.4	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	109	0.00
87	Indeno(1,2,3-cd)pyrene	1.486	1.520	-2.3	108	-0.02
88	Benzo(b)fluoranthene	1.262	1.282	-1.6	108	-0.01
89	Benzo(k)fluoranthene	1.240	1.276	-2.9	109	-0.01
90 C	Benzo(a)pyrene	1.075	1.104	-2.7	108	0.00
91	Dibenzo(a,h)anthracene	1.223	1.239	-1.3	108	-0.02
92	Benzo(g,h,i)perylene	1.205	1.230	-2.1	108	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041322\
Data File : BG053131.D
Acq On : 13 Apr 2022 10:08
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 14 02:55:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0