




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041323\
Data File : BG057077.D
Acq On : 13 Apr 2023 10:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 13 13:46:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 13 05:15:23 2023
Response via : Initial Calibration



TIC: BG057077.D\data.ms

(70) Pentachlorophenol (C)




17.405min (-0.000) 101090.04 ng

response 47907

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.80	59.53
264.00	58.50	58.10
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041323\
Data File : BG057077.D
Acq On : 13 Apr 2023 10:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 13 13:46:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG033023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 13 05:15:23 2023
Response via : Initial Calibration



TIC: BG057077.D\data.ms

(77) Benzidine**19.954min (-0.000) 2219.26 ng****response 254831**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	14.54
183.00	14.20	15.48
0.00	0.00	0.00