

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041324\
 Data File : BG060895.D
 Acq On : 13 Apr 2024 15:53
 Operator : MA/JU
 Sample : P2070-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-1

Quant Time: Apr 15 01:18:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.953	152	56609	20.000	ng	0.00
21) Naphthalene-d8	10.744	136	211473	20.000	ng	-0.01
39) Acenaphthene-d10	14.575	164	171841	20.000	ng	-0.01
64) Phenanthrene-d10	17.324	188	422995	20.000	ng	-0.01
76) Chrysene-d12	21.590	240	383447	20.000	ng	#-0.01
86) Perylene-d12	24.727	264	442973	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.538	112	385345	113.399	ng	0.00
7) Phenol-d6	7.119	99	480187	95.197	ng	0.00
23) Nitrobenzene-d5	9.105	82	401862	108.434	ng	0.00
42) 2,4,6-Tribromophenol	16.067	330	369354	189.347	ng	0.00
45) 2-Fluorobiphenyl	13.206	172	991665	84.694	ng	-0.01
79) Terphenyl-d14	19.951	244	1985145	91.160	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041324\
Data File : BG060895.D
Acq On : 13 Apr 2024 15:53
Operator : MA/JU
Sample : P2070-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-1

Quant Time: Apr 15 01:18:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration

