

Data Path : Z:\HPCHEM1\BNA_G\Data\BG041415\
 Data File : BG016413.D
 Acq On : 15 Apr 2015 00:51
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 03:40:47 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 14 14:00:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.576	0.508	11.8	113	-0.01
3	Pyridine	1.722	1.507	12.5	109	-0.01
4	n-Nitrosodimethylamine	0.512	0.456	10.9	110	-0.01
5 S	2-Fluorophenol	1.267	1.178	7.0	116	0.00
6	Aniline	2.717	2.421	10.9	110	0.00
7 S	Phenol-d6	1.902	1.753	7.8	114	0.00
8	2-Chlorophenol	1.522	1.449	4.8	118	0.00
9	Benzaldehyde	1.114	0.885	20.6	103	0.00
10 C	Phenol	2.035	1.864	8.4	112	0.00
11	bis(2-Chloroethyl)ether	1.623	1.408	13.2	109	0.00
12	1,3-Dichlorobenzene	1.537	1.466	4.6	121	0.00
13 C	1,4-Dichlorobenzene	1.594	1.503	5.7	122	0.00
14	1,2-Dichlorobenzene	1.525	1.409	7.6	117	0.00
15	Benzyl Alcohol	1.326	1.209	8.8	110	0.00
16	2,2'-oxybis(1-Chloropropane	1.828	1.485	18.8	101	0.00
17	2-Methylphenol	1.365	1.255	8.1	112	0.00
18	Hexachloroethane	0.610	0.547	10.3	115	0.00
19 P	n-Nitroso-di-n-propylamine	1.363	1.176	13.7	103	0.00
20	3+4-Methylphenols	1.891	1.761	6.9	112	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
22	Acetophenone	0.464	0.428	7.8	108	0.00
23 S	Nitrobenzene-d5	0.345	0.319	7.5	107	0.00
24	Nitrobenzene	0.369	0.335	9.2	106	0.00
25	Isophorone	0.767	0.676	11.9	102	0.00
26 C	2-Nitrophenol	0.172	0.190	-10.5	121	0.00
27	2,4-Dimethylphenol	0.321	0.313	2.5	113	0.00
28	bis(2-Chloroethoxy)methane	0.438	0.393	10.3	105	0.00
29 C	2,4-Dichlorophenol	0.254	0.265	-4.3	118	0.00
30	1,2,4-Trichlorobenzene	0.265	0.264	0.4	118	0.00
31	Naphthalene	1.042	0.987	5.3	113	0.00
32	Benzoic acid	0.182	0.226	-24.2	139	0.00
33	4-Chloroaniline	0.489	0.476	2.7	112	0.00
34 C	Hexachlorobutadiene	0.116	0.118	-1.7	121	0.00
35	Caprolactam	0.160	0.160	0.0	109	0.00
36 C	4-Chloro-3-methylphenol	0.335	0.328	2.1	111	0.00
37	2-Methylnaphthalene	0.750	0.716	4.5	113	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
39	1,2,4,5-Tetrachlorobenzene	0.440	0.425	3.4	117	0.00
40 P	Hexachlorocyclopentadiene	0.202	0.179	11.4	105	0.00
41 S	2,4,6-Tribromophenol	0.135	0.140	-3.7	119	0.00
42 C	2,4,6-Trichlorophenol	0.327	0.342	-4.6	121	0.00
43	2,4,5-Trichlorophenol	0.350	0.365	-4.3	122	0.00
44 S	2-Fluorobiphenyl	1.175	1.095	6.8	113	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG041415\
 Data File : BG016413.D
 Acq On : 15 Apr 2015 00:51
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 03:40:47 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 14 14:00:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.534	1.419	7.5	113	0.00
46	2-Chloronaphthalene	1.169	1.092	6.6	114	0.00
47	2-Nitroaniline	0.377	0.360	4.5	108	0.00
48	Acenaphthylene	2.078	1.949	6.2	112	0.00
49	Dimethylphthalate	1.466	1.378	6.0	112	0.00
50	2,6-Dinitrotoluene	0.355	0.348	2.0	115	0.00
51 C	Acenaphthene	1.242	1.157	6.8	114	0.00
52	3-Nitroaniline	0.453	0.450	0.7	114	0.00
53 P	2,4-Dinitrophenol	0.167	0.158	5.4	109	0.00
54	Dibenzofuran	1.723	1.614	6.3	114	0.00
55 P	4-Nitrophenol	0.374	0.373	0.3	116	0.00
56	2,4-Dinitrotoluene	0.483	0.487	-0.8	116	0.00
57	Fluorene	1.520	1.439	5.3	114	0.00
58	2,3,4,6-Tetrachlorophenol	0.270	0.284	-5.2	124	0.00
59	Diethylphthalate	1.560	1.452	6.9	111	0.00
60	4-Chlorophenyl-phenylether	0.622	0.591	5.0	115	0.00
61	4-Nitroaniline	0.514	0.508	1.2	114	0.00
62	Azobenzene	1.661	1.422	14.4	103	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
64	4,6-Dinitro-2-methylphenol	0.119	0.132	-10.9	121	0.00
65 c	n-Nitrosodiphenylamine	0.676	0.640	5.3	114	0.00
66	4-Bromophenyl-phenylether	0.160	0.152	5.0	115	0.00
67	Hexachlorobenzene	0.166	0.159	4.2	116	0.00
68	Atrazine	0.188	0.186	1.1	118	0.00
69 C	Pentachlorophenol	0.101	0.102	-1.0	119	0.00
70	Phenanthrene	1.125	1.042	7.4	115	0.00
71	Anthracene	1.115	1.055	5.4	116	0.00
72	Carbazole	1.119	1.064	4.9	116	0.00
73	Di-n-butylphthalate	1.369	1.288	5.9	112	0.00
74 C	Fluoranthene	1.159	1.115	3.8	118	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
76	Benzidine	0.848	0.784	7.5	109	0.00
77	Pyrene	1.393	1.334	4.2	117	0.00
78 S	Terphenyl-d14	0.855	0.797	6.8	114	0.00
79	Butylbenzylphthalate	0.684	0.705	-3.1	119	0.00
80	Benzo(a)anthracene	1.182	1.127	4.7	118	0.00
81	3,3'-Dichlorobenzidine	0.389	0.389	0.0	119	0.00
82	Chrysene	1.141	1.080	5.3	118	0.00
83	Bis(2-ethylhexyl)phthalate	0.926	0.949	-2.5	120	0.00
84 c	Di-n-octyl phthalate	1.509	1.570	-4.0	118	0.00
85	Indeno(1,2,3-cd)pyrene	1.186	1.170	1.3	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	118	0.00
87	Benzo(b)fluoranthene	1.171	1.128	3.7	118	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG041415\
Data File : BG016413.D
Acq On : 15 Apr 2015 00:51
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 15 03:40:47 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 14 14:00:56 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.146	1.083	5.5	119	0.00
89 C	Benzo(a)pyrene	1.098	1.056	3.8	118	0.00
90	Dibenzo(a,h)anthracene	1.068	1.034	3.2	120	0.00
91	Benzo(g,h,i)perylene	1.067	1.034	3.1	120	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0