

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041421\
 Data File : BG048691.D
 Acq On : 14 Apr 2021 13:15
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 14 15:33:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 14 15:25:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	18645	20.00	ng	0.00
21) Naphthalene-d8	10.907	136	78180	20.00	ng	0.00
39) Acenaphthene-d10	14.738	164	58001	20.00	ng	0.00
64) Phenanthrene-d10	17.493	188	132421	20.00	ng	0.00
76) Chrysene-d12	21.794	240	135873	20.00	ng	0.00
86) Perylene-d12	25.137	264	140309	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.625	112	24745	23.43	ng	0.00
7) Phenol-d6	7.235	99	34350	22.43	ng	0.00
23) Nitrobenzene-d5	9.256	82	35715	22.12	ng	0.00
42) 2,4,6-Tribromophenol	16.230	330	15022	19.70	ng	0.00
45) 2-Fluorobiphenyl	13.357	172	81827	20.97	ng	0.00
79) Terphenyl-d14	20.102	244	165575	22.29	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.492	88	4796	11.29	ng	97
3) Pyridine	3.909	79	11286	10.59	ng	98
4) n-Nitrosodimethylamine	3.815	42	8620	12.39	ng	# 96
6) Aniline	7.399	93	19090	10.85	ng	96
8) 2-Chlorophenol	7.640	128	13087	10.97	ng	97
9) Benzaldehyde	7.205	77	12705	12.98	ng	87
10) Phenol	7.264	94	16829	10.98	ng	94
11) bis(2-Chloroethyl)ether	7.493	93	11868	11.15	ng	83
12) 1,3-Dichlorobenzene	7.969	146	14351	10.67	ng	92
13) 1,4-Dichlorobenzene	8.116	146	15653	11.55	ng	# 89
14) 1,2-Dichlorobenzene	8.439	146	14618	11.26	ng	91
15) Benzyl Alcohol	8.316	79	14347	11.44	ng	92
16) 2,2'-oxybis(1-Chloropr...	8.598	45	11751	11.45	ng	88
17) 2-Methylphenol	8.527	107	10560	10.64	ng	# 83
18) Hexachloroethane	9.179	117	6359	12.62	ng	# 80
19) n-Nitroso-di-n-propyla...	8.886	70	11248	10.65	ng	88
20) 3+4-Methylphenols	8.856	107	14122	10.25	ng	87
22) Acetophenone	8.909	105	21190	10.51	ng	# 97
24) Nitrobenzene	9.291	77	17470	11.02	ng	91
25) Isophorone	9.814	82	32834	11.01	ng	# 92
26) 2-Nitrophenol	10.008	139	7116	9.77	ng	97
27) 2,4-Dimethylphenol	10.067	122	12039	10.51	ng	91
28) bis(2-Chloroethoxy)met...	10.302	93	13971	10.22	ng	92
29) 2,4-Dichlorophenol	10.554	162	13097	10.10	ng	94
30) 1,2,4-Trichlorobenzene	10.772	180	15919	10.68	ng	93
31) Naphthalene	10.960	128	40115	9.98	ng	# 94
32) Benzoic acid	10.178	122	5671	6.43	ng	# 86
33) 4-Chloroaniline	11.065	127	16856	9.94	ng	# 93
34) Hexachlorobutadiene	11.242	225	10705	9.83	ng	88
35) Caprolactam	11.817	113	4594	9.71	ng	# 89
36) 4-Chloro-3-methylphenol	12.188	107	15339	10.53	ng	97
37) 2-Methylnaphthalene	12.564	142	33660	10.52	ng	90
38) 1-Methylnaphthalene	12.781	142	31735	10.40	ng	93
40) 1,2,4,5-Tetrachloroben...	12.928	216	18057	10.04	ng	95
41) Hexachlorocyclopentadiene	12.904	237	5969	5.73	ng	78

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041421\
 Data File : BG048691.D
 Acq On : 14 Apr 2021 13:15
 Operator : CG/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 14 15:33:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 14 15:25:16 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	13.251	196	13012	10.53	ng	88
44) 2,4,5-Trichlorophenol	13.251	196	13012	9.84	ng	# 86
46) 1,1'-Biphenyl	13.562	154	42905	10.12	ng	96
47) 2-Chloronaphthalene	13.609	162	32733	10.14	ng	93
48) 2-Nitroaniline	13.809	65	11689	11.39	ng	92
49) Acenaphthylene	14.461	152	51554	10.19	ng	94
50) Dimethylphthalate	14.179	163	44462	10.33	ng	# 95
51) 2,6-Dinitrotoluene	14.303	165	9444	9.59	ng	90
52) Acenaphthene	14.802	154	31911	8.72	ng	96
53) 3-Nitroaniline	14.644	138	9681	10.02	ng	# 77
54) 2,4-Dinitrophenol	14.849	184	3637	6.57	ng	# 86
55) Dibenzofuran	15.131	168	53819	10.07	ng	96
56) 4-Nitrophenol	14.961	139	7482	9.60	ng	100
57) 2,4-Dinitrotoluene	15.090	165	14243	10.22	ng	97
58) Fluorene	15.789	166	46240	10.33	ng	92
59) 2,3,4,6-Tetrachlorophenol	15.366	232	11726	9.09	ng	94
60) Diethylphthalate	15.542	149	44544	10.07	ng	99
61) 4-Chlorophenyl-phenyle...	15.777	204	24346	10.11	ng	97
62) 4-Nitroaniline	15.801	138	12076	11.95	ng	99
63) Azobenzene	16.071	77	46367	10.65	ng	91
65) 4,6-Dinitro-2-methylph...	15.860	198	8906	11.19	ng	81
66) n-Nitrosodiphenylamine	15.989	169	38134	10.12	ng	98
67) 4-Bromophenyl-phenylether	16.676	248	16009	10.91	ng	93
68) Hexachlorobenzene	16.794	284	15383	10.04	ng	92
69) Atrazine	16.935	200	17197	11.19	ng	96
70) Pentachlorophenol	17.146	266	7289	7.65	ng	# 86
71) Phenanthrene	17.534	178	73544	10.70	ng	95
72) Anthracene	17.628	178	72341	10.56	ng	96
73) Carbazole	17.899	167	68997	10.62	ng	97
74) Di-n-butylphthalate	18.445	149	80174	11.05	ng	98
75) Fluoranthene	19.550	202	92718	10.88	ng	97
77) Benzidine	19.726	184	25477	5.54	ng	# 93
78) Pyrene	19.908	202	93128	9.97	ng	96
80) Butylbenzylphthalate	20.789	149	38081	10.97	ng	88
81) Benzo(a)anthracene	21.770	228	97378	10.73	ng	95
82) 3,3'-Dichlorobenzidine	21.682	252	36691	11.76	ng	91
83) Chrysene	21.841	228	90976	10.50	ng	97
84) Bis(2-ethylhexyl)phtha...	21.665	149	52066	10.77	ng	97
85) Di-n-octyl phthalate	22.916	149	86554	10.27	ng	99
87) Indeno(1,2,3-cd)pyrene	28.962	276	105727	10.75	ng	# 91
88) Benzo(b)fluoranthene	24.068	252	94627	10.38	ng	# 96
89) Benzo(k)fluoranthene	24.132	252	94256	10.76	ng	# 100
90) Benzo(a)pyrene	24.973	252	88340	10.52	ng	# 99
91) Dibenzo(a,h)anthracene	29.021	278	88656	10.56	ng	# 92
92) Benzo(g,h,i)perylene	30.149	276	87197	10.61	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 14 15:33:07 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 14 15:25:16 2021
Response via : Initial Calibration






Abundance Scan 807 (8.079 min): BG048693.D\data.ms (-800) (#1)




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.081 min Scan# 807
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDICC010

Tgt Ion	Ratio	Lower	Upper
152	100		
150	171.0	119.4	179.0
115	67.9	60.2	90.4







Abundance Scan 27 (3.496 min): BG048693.D\data.ms (-20) (#2)



1,4-Dioxane
Concen: 11.29 ng
RT: 3.492 min Scan# 26
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion	Ratio	Lower	Upper
88	100		
58	72.6	55.7	83.5
43	37.0	29.3	43.9









Pyridine
Concen: 10.59 ng
RT: 3.909 min Scan# 97
Delta R.T. 0.008 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010


Tgt Ion:	79	Resp:	11286
Ion	Ratio	Lower	Upper
79	100		
52	74.5	58.4	87.6
51	44.4	37.0	55.6



n-Nitrosodimethylamine
Concen: 12.39 ng
RT: 3.815 min Scan# 81
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	42	Resp:	8620
Ion	Ratio	Lower	Upper
42	100		
74	82.9	64.0	96.0
44	0.0	2.7	4.1#









2-Fluorophenol
 Concen: 23.43 ng
 RT: 5.625 min Scan# 389
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleID :
 SSTDICC010


Tgt Ion:	112	Resp:	24745
Ion	Ratio	Lower	Upper
112	100		
64	56.2	46.7	70.1
63	38.2	35.3	52.9



Aniline
 Concen: 10.85 ng
 RT: 7.399 min Scan# 691
 Delta R.T. 0.002 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	93	Resp:	19090
Ion	Ratio	Lower	Upper
93	100		
66	41.8	30.6	45.8
65	19.1	15.0	22.4









Phenol-d6
 Concen: 22.43 ng
 RT: 7.235 min Scan# 663
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010


Tgt Ion:	99	Resp:	34350
Ion	Ratio	Lower	Upper
99	100		
42	27.0	20.5	30.7
71	46.6	41.5	62.3



2-Chlorophenol
 Concen: 10.97 ng
 RT: 7.640 min Scan# 732
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	128	Resp:	13087
Ion	Ratio	Lower	Upper
128	100		
130	28.5	10.0	50.0
64	50.6	28.1	68.1









Benzaldehyde
 Concen: 12.98 ng
 RT: 7.205 min Scan# 658
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleID :
 SSTDICC010


Tgt Ion:	77	Resp:	12705
Ion	Ratio	Lower	Upper
77	100		
105	76.8	68.8	108.8
106	58.0	47.9	87.9



Phenol
 Concen: 10.98 ng
 RT: 7.264 min Scan# 668
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	94	Resp:	16829
Ion	Ratio	Lower	Upper
94	100		
65	27.9	12.9	52.9
66	62.6	39.4	79.4









bis(2-Chloroethyl)ether
 Concen: 11.15 ng
 RT: 7.493 min Scan# 707
 Delta R.T. 0.002 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleID :
 SSTDICC010

Tgt Ion:	93	Resp:	11868
Ion	Ratio	Lower	Upper
93	100		
63	87.4	50.4	90.4
95	29.5	15.3	55.3



1,3-Dichlorobenzene
 Concen: 10.67 ng
 RT: 7.969 min Scan# 788
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15


Tgt Ion:	146	Resp:	14351
Ion	Ratio	Lower	Upper
146	100		
148	74.1	52.6	78.8
75	41.5	35.7	53.5






Abundance Scan 814 (8.120 min): BG048693.D\data.ms (-806) #13

1,4-Dichlorobenzene
Concen: 11.55 ng
RT: 8.116 min Scan# 813
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampled :
SSTDIC010




Tgt Ion:	146	Resp:	15653
Ion	Ratio	Lower	Upper
146	100		
148	53.9	54.0	81.0#
111	41.7	34.3	51.5







Abundance Scan 868 (8.437 min): BG048693.D\data.ms (-860) #14

1,2-Dichlorobenzene
Concen: 11.26 ng
RT: 8.439 min Scan# 868
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	146	Resp:	14618
Ion	Ratio	Lower	Upper
146	100		
148	53.3	47.5	71.3
111	41.6	39.1	58.7









Benzyl Alcohol
 Concen: 11.44 ng
 RT: 8.316 min Scan# 847
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleID :
 SSTDICC010

Tgt Ion:	79	Resp:	14347
Ion	Ratio	Lower	Upper
79	100		
108	67.0	48.2	72.4
77	66.2	57.3	85.9



2,2'-oxybis(1-Chloropropane)
 Concen: 11.45 ng
 RT: 8.598 min Scan# 895
 Delta R.T. -0.010 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	45	Resp:	11751
Ion	Ratio	Lower	Upper
45	100		
77	15.8	2.2	42.2
79	17.8	0.0	34.3



Abundance Scan 883 (8.525 min): BG048693.D\data.ms (-875) #17

2-Methylphenol

Concen: 10.64 ng

RT: 8.527 min Scan# 883

Delta R.T. 0.002 min

Lab File: BG048691.D


Acq: 14 Apr 2021 13:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC010



Tgt Ion:107 Resp: 10560



Ion Ratio Lower Upper

107 100


108 121.4 84.6 127.0

77 79.2 50.7 76.1#

79 75.9 49.4 74.0#



Abundance Scan 883 (8.527 min): BG048691.D\data.ms (-832) (-)



Abundance Scan 994 (9.177 min): BG048693.D\data.ms (-986) #18

Hexachloroethane

Concen: 12.62 ng

RT: 9.179 min Scan# 994

Delta R.T. 0.002 min

Lab File: BG048691.D

Acq: 14 Apr 2021 13:15


Tgt Ion:117 Resp: 6359

Ion Ratio Lower Upper



117 100


119 61.4 74.4 111.6#

201 101.2 88.6 133.0



Abundance Scan 994 (9.179 min): BG048691.D\data.ms (-903) (-)









n-Nitroso-di-n-propylamine
Concen: 10.65 ng
RT: 8.886 min Scan# 944
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion:	70	Resp:	11248
Ion	Ratio	Lower	Upper
70	100		
42	56.0	54.5	81.7
101	13.2	9.0	13.6
130	19.0	17.0	25.6



3+4-Methylphenols
Concen: 10.25 ng
RT: 8.856 min Scan# 939
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Tgt Ion:	107	Resp:	14122
Ion	Ratio	Lower	Upper
107	100		
108	75.5	69.3	109.3
77	54.2	25.2	65.2
79	31.5	15.1	55.1






Abundance Scan 1289 (10.911 min): BG048693.D\data.ms (-1289) (-)

Naphthalene-d8
Concen: 20.00 ng
RT: 10.907 min Scan# 1288
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion	Ratio	Lower	Upper
136	100		
137	12.1	8.2	12.2
54	6.5	5.5	8.3
68	3.0	2.3	3.5







Abundance Scan 949 (8.913 min): BG048693.D\data.ms (-941) (-)

Acetophenone
Concen: 10.51 ng
RT: 8.909 min Scan# 948
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion	Ratio	Lower	Upper
105	100		
71	0.0	0.6	0.8#
51	35.7	26.6	40.0
120	21.2	16.2	24.4









Nitrobenzene-d5
 Concen: 22.12 ng
 RT: 9.256 min Scan# 1007
 Delta R.T. 0.002 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Tgt Ion:	82	Resp:	35715
Ion	Ratio	Lower	Upper
82	100		
128	34.2	29.8	44.6
54	49.4	45.8	68.8



Nitrobenzene
 Concen: 11.02 ng
 RT: 9.291 min Scan# 1013
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15


Tgt Ion:	77	Resp:	17470
Ion	Ratio	Lower	Upper
77	100		
123	28.9	27.5	41.3
65	12.6	12.6	18.8





Abundance Scan 1103 (9.818 min): BG048693.D\data.ms (-1025) (-)

Isophorone
Concen: 11.01 ng
RT: 9.814 min Scan# 1102
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Instrument :
BNA_G
ClientSampleId :
SSTDICC010



Tgt Ion:	82	Resp:	32834
Ion	Ratio	Lower	Upper
82	100		
95	6.5	8.5	12.7#
138	17.4	11.8	17.6




Abundance Scan 1102 (9.814 min): BG048691.D\data.ms (-1012) (-)





Abundance Scan 1136 (10.012 min): BG048693.D\data.ms (-1129) (-)


2-Nitrophenol
Concen: 9.77 ng
RT: 10.008 min Scan# 1135
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	139	Resp:	7116
Ion	Ratio	Lower	Upper
139	100		
109	49.1	37.8	56.6
65	62.6	51.8	77.8




Abundance Scan 1135 (10.008 min): BG048691.D\data.ms (-1045) (-)





Abundance Scan 1146 (10.070 min): BG048693.D\data.ms (-1129) (-)

2,4-Dimethylphenol
Concen: 10.51 ng
RT: 10.067 min Scan# 1145
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Instrument :
BNA_G
ClientSampleId :
SSTDIC010



Tgt Ion:	122	Resp:	12039
Ion	Ratio	Lower	Upper
122	100		
107	123.3	110.1	165.1
121	58.8	49.0	73.6




Abundance Scan 1145 (10.067 min): BG048691.D\data.ms (-1055) (-)





Abundance Scan 1186 (10.306 min): BG048693.D\data.ms (-1129) (-)


bis(2-Chloroethoxy)methane
Concen: 10.22 ng
RT: 10.302 min Scan# 1185
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	93	Resp:	13971
Ion	Ratio	Lower	Upper
93	100		
95	32.9	22.2	33.4
123	9.8	8.8	13.2




Abundance Scan 1185 (10.302 min): BG048691.D\data.ms (-1095) (-)






Abundance Scan 1229 (10.558 min): BG048693.D\data.ms (-1229) (-)

2,4-Dichlorophenol
Concen: 10.10 ng
RT: 10.554 min Scan# 1228
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampled :
SSTDIC010




Tgt Ion:	162	Resp:	13097
Ion	Ratio	Lower	Upper
162	100		
164	55.2	39.4	79.4
98	35.3	18.7	58.7







Abundance Scan 1265 (10.770 min): BG048693.D\data.ms (-1265) (-)

1,2,4-Trichlorobenzene
Concen: 10.68 ng
RT: 10.772 min Scan# 1265
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	180	Resp:	15919
Ion	Ratio	Lower	Upper
180	100		
182	90.9	78.9	118.3
145	30.4	26.2	39.4









#30 (-)
Naphthalene
Concen: 9.98 ng
RT: 10.960 min Scan# 1297
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010


Tgt Ion:	128	Resp:	40115
Ion	Ratio	Lower	Upper
128	100		
129	13.6	8.5	12.7#
127	14.6	10.2	15.4



#30 (-)
Benzoic acid
Concen: 6.43 ng
RT: 10.178 min Scan# 1164
Delta R.T. -0.033 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	122	Resp:	5671
Ion	Ratio	Lower	Upper
122	100		
105	105.6	92.3	132.3
77	108.4	66.5	106.5#









4-Chloroaniline
Concen: 9.94 ng
RT: 11.065 min Scan# 1315
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010


Tgt Ion:	127	Resp:	16856
Ion	Ratio	Lower	Upper
127	100		
129	28.7	26.7	40.1
65	41.5	32.8	49.2
92	27.4	16.2	24.4#



Hexachlorobutadiene
Concen: 9.83 ng
RT: 11.242 min Scan# 1345
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	225	Resp:	10705
Ion	Ratio	Lower	Upper
225	100		
223	74.6	49.8	74.8
227	63.3	46.4	69.6









Caprolactam
Concen: 9.71 ng
RT: 11.817 min Scan# 1443
Delta R.T. -0.027 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010


Tgt Ion:	113	Resp:	4594
Ion	Ratio	Lower	Upper
113	100		
55	134.6	113.8	153.8
56	150.9	105.6	145.6#



4-Chloro-3-methylphenol
Concen: 10.53 ng
RT: 12.188 min Scan# 1506
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	107	Resp:	15339
Ion	Ratio	Lower	Upper
107	100		
144	26.0	24.8	37.2
142	78.5	62.2	93.4









2-Methylnaphthalene
Concen: 10.52 ng
RT: 12.564 min Scan# 1570
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion:	142	Resp:	33660
Ion	Ratio	Lower	Upper
142	100		
141	75.3	68.8	103.2
115	40.9	36.4	54.6



1-Methylnaphthalene
Concen: 10.40 ng
RT: 12.781 min Scan# 1607
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Tgt Ion:	142	Resp:	31735
Ion	Ratio	Lower	Upper
142	100		
141	81.9	70.6	106.0
116	5.1	4.0	6.0






Abundance Scan 1940 (14.736 min): BG048693.D\data.ms (-1839) (-)

Acenaphthene-d10
Concen: 20.00 ng
RT: 14.738 min Scan# 1940
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion:	164	Resp:	58001
Ion Ratio	Lower	Upper	
164	100		
162	93.2	81.4	122.0
160	43.0	34.9	52.3






Abundance Scan 1633 (12.932 min): BG048693.D\data.ms (-1626) (-)

1,2,4,5-Tetrachlorobenzene
Concen: 10.04 ng
RT: 12.928 min Scan# 1632
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Tgt Ion:	216	Resp:	18057
Ion Ratio	Lower	Upper	
216	100		
214	82.8	61.4	92.2
179	21.0	17.3	25.9
108	19.3	16.3	24.5





Abundance Scan 1629 (12.908 min): BG048693.D\data.ms (-1621) (-)

Hexachlorocyclopentadiene
Concen: 5.73 ng
RT: 12.904 min Scan# 1628
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Instrument :
BNA_G
ClientSampleID :
SSTDIC010



Tgt Ion	Ratio	Lower	Upper
237	100		
235	75.1	36.6	76.6
272	9.8	0.0	32.5




Abundance Scan 1628 (12.904 min): BG048691.D\data.ms (-1538) (-)





Abundance Scan 2194 (16.228 min): BG048693.D\data.ms (-2192) (-)


2,4,6-Tribromophenol
Concen: 19.70 ng
RT: 16.230 min Scan# 2194
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion	Ratio	Lower	Upper
330	100		
332	103.0	77.3	115.9
141	44.7	35.6	53.4




Abundance Scan 2194 (16.230 min): BG048691.D\data.ms (-2103) (-)





Abundance Scan 1673 (13.167 min): BG048693.D\data.ms (-1643) (-)

2,4,6-Trichlorophenol
Concen: 10.53 ng
RT: 13.251 min Scan# 1687
Delta R.T. 0.084 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Instrument :
BNA_G
ClientSampleID :
SSTDIC010



Tgt Ion	196	198	200
Ratio	100	88.9	25.4
Lower		82.1	23.8
Upper		123.1	35.8






Abundance Scan 1687 (13.251 min): BG048691.D\data.ms (-1582) (-)




Abundance Scan 1687 (13.249 min): BG048693.D\data.ms (-1641) (-)

2,4,5-Trichlorophenol
Concen: 9.84 ng
RT: 13.251 min Scan# 1687
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion	196	198	97	132
Ratio	100	88.9	50.3	20.2
Lower		87.4	43.3	21.4
Upper		131.0	64.9	32.2#





Abundance Scan 1687 (13.251 min): BG048691.D\data.ms (-1596) (-)





Abundance Scan 1705 (13.355 min): BG048693.D\data.ms (-1615) (-)

2-Fluorobiphenyl
Concen: 20.97 ng
RT: 13.357 min Scan# 1705
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010





Tgt Ion:	172	Resp:	81827
Ion	Ratio	Lower	Upper
172	100		
171	31.1	28.6	43.0
170	20.5	18.6	27.8





Abundance Scan 1741 (13.566 min): BG048693.D\data.ms (-1741) (-)

1,1'-Biphenyl
Concen: 10.12 ng
RT: 13.562 min Scan# 1740
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	154	Resp:	42905
Ion	Ratio	Lower	Upper
154	100		
153	38.2	20.5	60.5
76	8.2	0.0	29.3



Abundance Scan 1749 (13.613 min): BG048693.D\data.ms (-1741) (-)

2-Chloronaphthalene

Concen: 10.14 ng

RT: 13.609 min Scan# 1748

Delta R.T. -0.004 min

Lab File: BG048691.D


Acq: 14 Apr 2021 13:15

Instrument :

BNA_G

ClientSampleId :

SSTDIC010






Tgt Ion:162 Resp: 32733

Ion Ratio Lower Upper

162 100

127 33.7 32.7 49.1

164 33.0 27.0 40.4



Abundance Scan 1784 (13.819 min): BG048693.D\data.ms (-1746) (-)

2-Nitroaniline

Concen: 11.39 ng

RT: 13.809 min Scan# 1782

Delta R.T. -0.010 min

Lab File: BG048691.D

Acq: 14 Apr 2021 13:15




Tgt Ion: 65 Resp: 11689

Ion Ratio Lower Upper

65 100

92 54.7 47.2 70.8


138 94.0 68.0 102.0






Abundance Scan 1893 (14.459 min): BG048693.D\data.ms (-1802) (-)

Acenaphthylene
Concen: 10.19 ng
RT: 14.461 min Scan# 1893
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion:	152	Resp:	51554
Ion	Ratio	Lower	Upper
152	100		
151	19.8	18.2	27.4
153	12.3	11.7	17.5







Abundance Scan 1846 (14.183 min): BG048693.D\data.ms (-1750) (-)

Dimethylphthalate
Concen: 10.33 ng
RT: 14.179 min Scan# 1845
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion:	163	Resp:	44462
Ion	Ratio	Lower	Upper
163	100		
194	5.0	3.0	4.6#
164	8.6	8.6	12.8#









2,6-Dinitrotoluene
Concen: 9.59 ng
RT: 14.303 min Scan# 1866
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion:	165	Resp:	9444
Ion Ratio	Lower	Upper	
165	100		
63	85.0	63.3	94.9
89	90.3	63.4	95.0



Acenaphthene
Concen: 8.72 ng
RT: 14.802 min Scan# 1951
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Tgt Ion:	154	Resp:	31911
Ion Ratio	Lower	Upper	
154	100		
153	113.9	87.0	130.4
152	49.7	40.1	60.1






Abundance Scan 1924 (14.642 min): BG048693.D\data.ms (-1954) (-)

3-Nitroaniline
Concen: 10.02 ng
RT: 14.644 min Scan# 1924
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion:	138	Resp:	9681
Ion	Ratio	Lower	Upper
138	100		
108	11.6	8.3	12.5
92	92.7	96.4	144.6#





Abundance Scan 1959 (14.847 min): BG048693.D\data.ms (-1954) (-)


2,4-Dinitrophenol
Concen: 6.57 ng
RT: 14.849 min Scan# 1959
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Tgt Ion:	184	Resp:	3637
Ion	Ratio	Lower	Upper
184	100		
63	88.4	68.3	102.5
154	101.3	62.6	93.8#



Abundance Scan 1959 (14.849 min): BG048691.D\data.ms (-1868) (-)






Abundance Scan 2008 (15.135 min): BG048693.D\data.ms (-1917) (-)




Dibenzenofuran
Concen: 10.07 ng
RT: 15.131 min Scan# 2007
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampled :
SSTDIC010

Tgt Ion:	168	Resp:	53819
Ion	Ratio	Lower	Upper
168	100		
139	43.7	33.4	50.2
169	12.7	12.3	18.5






Abundance Scan 1978 (14.959 min): BG048693.D\data.ms (-1917) (-)



4-Nitrophenol
Concen: 9.60 ng
RT: 14.961 min Scan# 1978
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Tgt Ion:	139	Resp:	7482
Ion	Ratio	Lower	Upper
139	100		
109	136.6	116.3	156.3
65	124.2	105.0	145.0






Abundance Scan 2002 (15.100 min): BG048693.D\data.ms (-1957) (-)

2,4-Dinitrotoluene
Concen: 10.22 ng
RT: 15.090 min Scan# 2000
Delta R.T. -0.010 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampled :
SSTDIC010




Tgt Ion:	165	Resp:	14243
Ion	Ratio	Lower	Upper
165	100		
63	59.8	48.0	72.0
89	86.3	65.4	98.0
182	2.3	2.0	3.0






Abundance Scan 2119 (15.787 min): BG048693.D\data.ms (-2150) (-)

Fluorene
Concen: 10.33 ng
RT: 15.789 min Scan# 2119
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Tgt Ion:	166	Resp:	46240
Ion	Ratio	Lower	Upper
166	100		
165	88.2	77.7	116.5
167	13.6	10.8	16.2






Abundance Scan 2047 (15.364 min): BG048693.D\data.ms (-2659) (-)

2,3,4,6-Tetrachlorophenol
Concen: 9.09 ng
RT: 15.366 min Scan# 2047
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampled :
SSTDIC010




Tgt Ion:	232	Resp:	11726
Ion	Ratio	Lower	Upper
232	100		
131	51.7	37.2	55.8
130	4.1	2.8	4.2
166	28.3	24.0	36.0






Abundance Scan 2078 (15.546 min): BG048693.D\data.ms (-2660) (-)

Diethylphthalate
Concen: 10.07 ng
RT: 15.542 min Scan# 2077
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Tgt Ion:	149	Resp:	44544
Ion	Ratio	Lower	Upper
149	100		
177	24.1	19.0	28.6
150	13.8	10.2	15.4






Abundance Scan 2117 (15.776 min): BG048693.D\data.ms (-2160) (-)

4-Chlorophenyl-phenylether
Concen: 10.11 ng
RT: 15.777 min Scan# 2117
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion:	204	Resp:	24346
Ion Ratio	Lower	Upper	
204	100		
206	32.2	24.6	37.0
141	62.4	47.8	71.8






Abundance Scan 2122 (15.805 min): BG048693.D\data.ms (-2160) (-)

4-Nitroaniline
Concen: 11.95 ng
RT: 15.801 min Scan# 2121
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15




Tgt Ion:	138	Resp:	12076
Ion Ratio	Lower	Upper	
138	100		
92	54.3	32.3	72.3
108	152.1	132.4	172.4






Abundance Scan 2167 (16.069 min): BG048693.D\data.ms (-2160) (-)

Azobenzene
Concen: 10.65 ng
RT: 16.071 min Scan# 2167
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010







Tgt Ion:	77	Resp:	46367
Ion	Ratio	Lower	Upper
77	100		
182	22.0	7.9	47.9
105	20.1	0.0	37.9
51	41.5	27.3	67.3




Abundance Scan 2410 (17.497 min): BG048693.D\data.ms (-2404) (-)

Phenanthrene-d10
Concen: 20.00 ng
RT: 17.493 min Scan# 2409
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	188	Resp:	132421
Ion	Ratio	Lower	Upper
188	100		
94	5.5	5.4	8.0
80	5.3	3.7	5.5





Abundance Scan 2132 (15.864 min): BG048693.D\data.ms (-2132) (-)




4,6-Dinitro-2-methylphenol
Concen: 11.19 ng
RT: 15.860 min Scan# 2131
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleID :
SSTDIC010


Tgt Ion:	198	Resp:	8906
Ion	Ratio	Lower	Upper
198	100		
51	38.8	29.4	69.4
105	48.7	15.1	55.1



Abundance Scan 2131 (15.860 min): BG048691.D\data.ms (-2041) (-)





Abundance Scan 2154 (15.993 min): BG048693.D\data.ms (-2154) (-)




n-Nitrosodiphenylamine
Concen: 10.12 ng
RT: 15.989 min Scan# 2153
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	169	Resp:	38134
Ion	Ratio	Lower	Upper
169	100		
168	71.2	56.3	84.5
167	32.4	28.4	42.6




Abundance Scan 2153 (15.989 min): BG048691.D\data.ms (-2063) (-)





Abundance Scan 2270 (16.674 min): BG048693.D\data.ms (-2267) (-)

4-Bromophenyl-phenylether
Concen: 10.91 ng
RT: 16.676 min Scan# 2270
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Instrument :
BNA_G
ClientSampleId :
SSTDIC010



Tgt Ion:	248	Resp:	16009
Ion	Ratio	Lower	Upper
248	100		
250	103.5	76.1	114.1
141	86.7	65.4	98.2






Abundance Scan 2270 (16.676 min): BG048691.D\data.ms (-2179) (-)




Abundance Scan 2291 (16.798 min): BG048693.D\data.ms (-2268) (-)


Hexachlorobenzene
Concen: 10.04 ng
RT: 16.794 min Scan# 2290
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	284	Resp:	15383
Ion	Ratio	Lower	Upper
284	100		
142	38.5	27.0	40.6
249	31.0	28.2	42.2



Abundance Scan 2290 (16.794 min): BG048691.D\data.ms (-2200) (-)









Scan 2315 (16.939 min): BG048693.D\data.ms (-2309) (-)
 Atrazine
 Concen: 11.19 ng
 RT: 16.935 min Scan# 2314
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampled :
 SSTDICC010

Tgt Ion:	200	Resp:	17197
Ion Ratio	Lower	Upper	
200	100		
173	23.9	3.7	43.7
215	43.1	26.9	66.9




Scan 2350 (17.145 min): BG048693.D\data.ms (-2372) (-)
 Pentachlorophenol
 Concen: 7.65 ng
 RT: 17.146 min Scan# 2350
 Delta R.T. 0.002 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	266	Resp:	7289
Ion Ratio	Lower	Upper	
266	100		
268	81.5	49.0	73.4#
264	59.4	49.0	73.6






Abundance Scan 2417 (17.538 min): BG048693.D\data.ms (-2479) (-)




Phenanthrene
Concen: 10.70 ng
RT: 17.534 min Scan# 2416
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleID :
SSTDIC010

Tgt Ion:	178	Resp:	73544
Ion Ratio	Lower	Upper	
178	100		
176	16.8	16.0	24.0
179	16.2	14.2	21.4






Abundance Scan 2432 (17.626 min): BG048693.D\data.ms (-2479) (-)



Anthracene
Concen: 10.56 ng
RT: 17.628 min Scan# 2432
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15


Tgt Ion:	178	Resp:	72341
Ion Ratio	Lower	Upper	
178	100		
176	16.5	14.6	21.8
179	14.9	13.4	20.0






Abundance Scan 2478 (17.897 min): BG048693.D\data.ms (-2478) (-)

Carbazole
Concen: 10.62 ng
RT: 17.899 min Scan# 2478
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010







Tgt Ion:	167	Resp:	68997
Ion	Ratio	Lower	Upper
167	100		
166	20.5	17.4	26.2
139	14.8	10.7	16.1




Abundance Scan 2572 (18.449 min): BG048693.D\data.ms (-2572) (-)

Di-n-butylphthalate
Concen: 11.05 ng
RT: 18.445 min Scan# 2571
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	149	Resp:	80174
Ion	Ratio	Lower	Upper
149	100		
150	10.6	7.5	11.3
104	8.9	6.6	10.0









Fluoranthene
 Concen: 10.88 ng
 RT: 19.550 min Scan# 2759
 Delta R.T. 0.002 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Tgt Ion:	202	Resp:	92718
Ion Ratio	Lower	Upper	
202	100		
101	4.7	0.0	24.7
203	16.6	0.0	38.5




Chrysene-d12
 Concen: 20.00 ng
 RT: 21.794 min Scan# 3141
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15

Tgt Ion:	240	Resp:	135873
Ion Ratio	Lower	Upper	
240	100		
120	4.4	4.2	6.2
236	26.7	21.8	32.6






Abundance Scan 2789 (19.724 min): BG048693.D\data.ms (-2789) (-)




Benzidine
Concen: 5.54 ng
RT: 19.726 min Scan# 2789
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion:	184	Resp:	25477
Ion Ratio	Lower	Upper	
184	100		
185	16.3	10.7	16.1#
183	14.6	9.4	14.0#






Abundance Scan 2821 (19.912 min): BG048693.D\data.ms (-2821) (-)




Pyrene
Concen: 9.97 ng
RT: 19.908 min Scan# 2820
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	202	Resp:	93128
Ion Ratio	Lower	Upper	
202	100		
200	21.8	15.8	23.8
203	19.9	14.6	21.8





Abundance Scan 2854 (20.106 min): BG048693.D\data.ms (-2679) (-)




Terphenyl-d14
Concen: 22.29 ng
RT: 20.102 min Scan# 2853
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010


Tgt Ion:	244	Resp:	165575
Ion Ratio	Lower	Upper	
244	100		
212	7.4	6.9	10.3
122	6.0	5.0	7.6



Abundance Scan 2853 (20.102 min): BG048691.D\data.ms (-2763) (-)





Abundance Scan 2970 (20.787 min): BG048693.D\data.ms (-2680) (-)





Butylbenzylphthalate
Concen: 10.97 ng
RT: 20.789 min Scan# 2970
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	149	Resp:	38081
Ion Ratio	Lower	Upper	
149	100		
91	67.9	63.4	95.2
206	19.8	18.4	27.6



Abundance Scan 2970 (20.789 min): BG048691.D\data.ms (-2879) (-)









Benzo(a)anthracene
 Concen: 10.73 ng
 RT: 21.770 min Scan# 3137
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Tgt Ion:	228	Resp:	97378
Ion	Ratio	Lower	Upper
228	100		
226	26.0	22.7	34.1
229	23.0	16.0	24.0



3,3'-Dichlorobenzidine
 Concen: 11.76 ng
 RT: 21.682 min Scan# 3122
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15


Tgt Ion:	252	Resp:	36691
Ion	Ratio	Lower	Upper
252	100		
254	72.9	52.4	78.6
126	8.6	5.9	8.9






Abundance Scan 3150 (21.845 min): BG048693.D\data.ms (-3182) (-)

Chrysene
Concen: 10.50 ng
RT: 21.841 min Scan# 3149
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.5	23.1	34.7
229	19.5	15.3	22.9






Abundance Scan 3119 (21.663 min): BG048693.D\data.ms (-3182) (-)


Bis(2-ethylhexyl)phthalate
Concen: 10.77 ng
RT: 21.665 min Scan# 3119
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion	Ratio	Lower	Upper
149	100		
167	29.3	22.3	33.5
279	5.5	3.8	5.8






Abundance Scan 3333 (22.920 min): BG048693.D\data.ms (-3325) (-)




Di-n-octyl phthalate
Concen: 10.27 ng
RT: 22.916 min Scan# 3332
Delta R.T. -0.004 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion:	149	Resp:	86554
Ion Ratio	Lower	Upper	
149	100		
167	1.8	1.4	2.2
43	11.1	8.6	12.8






Abundance Scan 3710 (25.135 min): BG048693.D\data.ms (-3696) (-)




Perylene-d12
Concen: 20.00 ng
RT: 25.137 min Scan# 3710
Delta R.T. 0.002 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion:	264	Resp:	140309
Ion Ratio	Lower	Upper	
264	100		
260	20.7	19.4	29.2
265	22.4	16.9	25.3






Abundance Scan 4364 (28.978 min): BG048693.D\data.ms (-4364) (-)




Indeno(1,2,3-cd)pyrene
Concen: 10.75 ng
RT: 28.962 min Scan# 4361
Delta R.T. -0.016 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010

Tgt Ion	Ratio	Lower	Upper
276	100		
138	2.1	5.1	7.7#
277	20.9	19.8	29.6







Abundance Scan 3530 (24.078 min): BG048693.D\data.ms (-3530) (-)



Benzo(b)fluoranthene
Concen: 10.38 ng
RT: 24.068 min Scan# 3528
Delta R.T. -0.010 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	16.2	24.4
125	6.4	3.0	4.4#









Scan 3541 (24.142 min): BG048693.D\data.ms (-3539) (-)
 Benzo(k)fluoranthene
 Concen: 10.76 ng
 RT: 24.132 min Scan# 3539
 Delta R.T. -0.010 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC010

Tgt Ion:	252	Resp:	94256
Ion Ratio	Lower	Upper	
252	100		
253	22.5	18.0	27.0
125	4.8	3.1	4.7#



Scan 3683 (24.977 min): BG048693.D\data.ms (-3680) (-)
 Benzo(a)pyrene
 Concen: 10.52 ng
 RT: 24.973 min Scan# 3682
 Delta R.T. -0.004 min
 Lab File: BG048691.D
 Acq: 14 Apr 2021 13:15


Tgt Ion:	252	Resp:	88340
Ion Ratio	Lower	Upper	
252	100		
253	20.7	16.6	25.0
125	5.3	3.5	5.3#






Abundance Scan 4374 (29.036 min): BG048693.D\data.ms (-4390) (-)

Dibenzo(a,h)anthracene
Concen: 10.56 ng
RT: 29.021 min Scan# 4371
Delta R.T. -0.016 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15

Instrument :
BNA_G
ClientSampleId :
SSTDIC010




Tgt Ion	Ratio	Lower	Upper
278	100		
139	6.1	4.6	6.8
279	17.5	17.9	26.9#



Abundance Scan 4563 (30.149 min): BG048693.D\data.ms (-4542) (-)

Benzo(g,h,i)perylene
Concen: 10.61 ng
RT: 30.149 min Scan# 4563
Delta R.T. -0.033 min
Lab File: BG048691.D
Acq: 14 Apr 2021 13:15



Tgt Ion	Ratio	Lower	Upper
276	100		
277	20.5	18.4	27.6
138	12.7	7.0	10.6#

