


Hexachlorocyclopentadiene

Response Ratio




$$\text{Response} = 3.831\text{e-}001 * \text{Amt} - 1.122\text{e-}001$$

Coef of Det (r^2) = 0.997930 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041421.M

Calibration Table Last Updated: Thu Apr 15 10:59:37 2021

2,4-Dinitrophenol



Response = 2.224e-001 * Amt - 5.874e-002
 Coef of Det (r^2) = 0.997339 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041421.M
 Calibration Table Last Updated: Thu Apr 15 10:59:37 2021