

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041421\
 Data File : BG048698.D
 Acq On : 14 Apr 2021 19:02
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG041421

Quant Time: Apr 15 02:04:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 01:59:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	40.423	-1.1	112	0.00
3	Pyridine	40.000	40.978	-2.4	109	0.00
4	n-Nitrosodimethylamine	40.000	41.422	-3.6	113	0.00
5 S	2-Fluorophenol	80.000	76.800	4.0	110	0.00
6	Aniline	40.000	39.846	0.4	108	0.00
7 S	Phenol-d6	80.000	77.448	3.2	108	0.00
8	2-Chlorophenol	40.000	38.669	3.3	111	0.00
9	Benzaldehyde	40.000	37.690	5.8	110	0.00
10 C	Phenol	40.000	38.279	4.3	109	0.00
11	bis(2-Chloroethyl)ether	40.000	38.221	4.4	110	0.00
12	1,3-Dichlorobenzene	40.000	39.095	2.3	114	0.00
13 C	1,4-Dichlorobenzene	40.000	38.406	4.0	109	0.00
14	1,2-Dichlorobenzene	40.000	40.085	-0.2	112	0.00
15	Benzyl Alcohol	40.000	38.887	2.8	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.498	1.3	108	0.00
17	2-Methylphenol	40.000	38.553	3.6	106	0.00
18	Hexachloroethane	40.000	36.855	7.9	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.490	1.3	111	0.00
20	3+4-Methylphenols	40.000	38.829	2.9	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	39.375	1.6	108	0.00
23 S	Nitrobenzene-d5	80.000	81.061	-1.3	108	0.00
24	Nitrobenzene	40.000	41.896	-4.7	112	0.00
25	Isophorone	40.000	40.739	-1.8	109	0.00
26 C	2-Nitrophenol	40.000	39.535	1.2	104	0.00
27	2,4-Dimethylphenol	40.000	40.481	-1.2	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.883	-2.2	109	0.00
29 C	2,4-Dichlorophenol	40.000	39.208	2.0	105	0.00
30	1,2,4-Trichlorobenzene	40.000	40.333	-0.8	112	0.00
31	Naphthalene	40.000	39.446	1.4	104	0.00
32	Benzoic acid	40.000	35.874	10.3	94	0.01
33	4-Chloroaniline	40.000	38.943	2.6	102	0.00
34 C	Hexachlorobutadiene	40.000	39.584	1.0	103	0.00
35	Caprolactam	40.000	40.575	-1.4	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.746	-1.9	112	0.00
37	2-Methylnaphthalene	40.000	39.472	1.3	106	0.00
38	1-Methylnaphthalene	40.000	39.927	0.2	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.859	-4.6	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.424	-1.1	109	0.00
42 S	2,4,6-Tribromophenol	80.000	83.814	-4.8	109	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.725	-4.3	106	0.00
44	2,4,5-Trichlorophenol	40.000	42.015	-5.0	108	0.00
45 S	2-Fluorobiphenyl	80.000	82.295	-2.9	108	0.00
46	1,1'-Biphenyl	40.000	41.321	-3.3	108	0.00
47	2-Chloronaphthalene	40.000	40.234	-0.6	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041421\
 Data File : BG048698.D
 Acq On : 14 Apr 2021 19:02
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG041421

Quant Time: Apr 15 02:04:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 15 01:59:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.069	-5.2	103	0.00
49	Acenaphthylene	40.000	40.209	-0.5	105	0.00
50	Dimethylphthalate	40.000	40.327	-0.8	105	0.00
51	2,6-Dinitrotoluene	40.000	39.764	0.6	101	0.00
52 C	Acenaphthene	40.000	41.414	-3.5	106	0.00
53	3-Nitroaniline	40.000	40.691	-1.7	109	0.00
54 P	2,4-Dinitrophenol	40.000	39.624	0.9	102	0.00
55	Dibenzofuran	40.000	41.058	-2.6	106	0.00
56 P	4-Nitrophenol	40.000	40.988	-2.5	103	0.00
57	2,4-Dinitrotoluene	40.000	40.601	-1.5	104	0.00
58	Fluorene	40.000	40.488	-1.2	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.996	-5.0	107	0.00
60	Diethylphthalate	40.000	41.118	-2.8	108	0.00
61	4-Chlorophenyl-phenylether	40.000	40.744	-1.9	107	0.00
62	4-Nitroaniline	40.000	39.249	1.9	106	0.00
63	Azobenzene	40.000	40.788	-2.0	107	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.528	-1.3	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.067	-0.2	106	0.00
67	4-Bromophenyl-phenylether	40.000	40.331	-0.8	113	0.00
68	Hexachlorobenzene	40.000	41.020	-2.6	108	0.00
69	Atrazine	40.000	39.690	0.8	107	0.00
70 C	Pentachlorophenol	40.000	41.273	-3.2	111	0.00
71	Phenanthrene	40.000	40.322	-0.8	108	0.00
72	Anthracene	40.000	39.858	0.4	107	0.00
73	Carbazole	40.000	39.095	2.3	106	0.00
74	Di-n-butylphthalate	40.000	39.970	0.1	109	0.00
75 C	Fluoranthene	40.000	39.609	1.0	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	112	0.00
77	Benzydine	40.000	43.820	-9.6	101	0.00
78	Pyrene	40.000	39.491	1.3	108	0.00
79 S	Terphenyl-d14	80.000	76.594	4.3	107	0.00
80	Butylbenzylphthalate	40.000	39.028	2.4	108	0.00
81	Benzo(a)anthracene	40.000	39.172	2.1	109	0.00
82	3,3'-Dichlorobenzidine	40.000	39.155	2.1	109	0.00
83	Chrysene	40.000	38.625	3.4	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.825	0.4	110	0.00
85 c	Di-n-octyl phthalate	40.000	38.942	2.6	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.269	-0.7	109	0.00
88	Benzo(b)fluoranthene	40.000	40.648	-1.6	109	0.00
89	Benzo(k)fluoranthene	40.000	40.364	-0.9	110	0.00
90 C	Benzo(a)pyrene	40.000	40.299	-0.7	109	0.00
91	Dibenzo(a,h)anthracene	40.000	39.270	1.8	108	0.00
92	Benzo(g,h,i)perylene	40.000	39.563	1.1	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041421\
Data File : BG048698.D
Acq On : 14 Apr 2021 19:02
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG041421

Quant Time: Apr 15 02:04:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG041421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 15 01:59:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0