

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041422\
 Data File : BG053150.D
 Acq On : 14 Apr 2022 10:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 02:20:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	102	0.00
2	1,4-Dioxane	40.000	41.853	-4.6	103	0.00
3	Pyridine	40.000	42.993	-7.5	99	0.00
4	n-Nitrosodimethylamine	40.000	41.268	-3.2	96	0.00
5 S	2-Fluorophenol	80.000	85.542	-6.9	103	0.00
6	Aniline	40.000	41.581	-4.0	98	0.00
7 S	Phenol-d6	80.000	84.534	-5.7	100	0.00
8	2-Chlorophenol	40.000	41.846	-4.6	99	0.00
9	Benzaldehyde	40.000	38.719	3.2	104	0.00
10 C	Phenol	40.000	42.006	-5.0	102	0.00
11	bis(2-Chloroethyl)ether	40.000	41.856	-4.6	102	0.00
12	1,3-Dichlorobenzene	40.000	42.081	-5.2	102	0.00
13 C	1,4-Dichlorobenzene	40.000	42.158	-5.4	102	0.00
14	1,2-Dichlorobenzene	40.000	41.147	-2.9	101	0.00
15	Benzyl Alcohol	40.000	41.677	-4.2	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.737	-1.8	96	0.00
17	2-Methylphenol	40.000	41.462	-3.7	101	0.00
18	Hexachloroethane	40.000	41.746	-4.4	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.207	-3.0	98	0.00
20	3+4-Methylphenols	40.000	42.319	-5.8	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	41.274	-3.2	98	0.00
23 S	Nitrobenzene-d5	80.000	81.917	-2.4	98	0.00
24	Nitrobenzene	40.000	40.328	-0.8	99	0.00
25	Isophorone	40.000	40.831	-2.1	96	0.00
26 C	2-Nitrophenol	40.000	40.689	-1.7	96	0.00
27	2,4-Dimethylphenol	40.000	41.266	-3.2	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.970	-2.4	97	0.00
29 C	2,4-Dichlorophenol	40.000	41.117	-2.8	96	0.00
30	1,2,4-Trichlorobenzene	40.000	40.478	-1.2	96	0.00
31	Naphthalene	40.000	41.250	-3.1	98	0.00
32	Benzoic acid	40.000	44.219	-10.5	102	0.00
33	4-Chloroaniline	40.000	41.736	-4.3	98	0.00
34 C	Hexachlorobutadiene	40.000	39.928	0.2	96	0.00
35	Caprolactam	40.000	42.606	-6.5	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.100	-5.3	98	0.00
37	2-Methylnaphthalene	40.000	40.854	-2.1	97	0.00
38	1-Methylnaphthalene	40.000	41.394	-3.5	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.075	-0.2	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.101	7.2	88	0.00
42 S	2,4,6-Tribromophenol	80.000	83.229	-4.0	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.032	-2.6	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.091	-0.2	96	0.00
45 S	2-Fluorobiphenyl	80.000	80.110	-0.1	99	0.00
46	1,1'-Biphenyl	40.000	40.066	-0.2	99	0.00
47	2-Chloronaphthalene	40.000	40.025	-0.1	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041422\
 Data File : BG053150.D
 Acq On : 14 Apr 2022 10:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 15 02:20:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 15:31:24 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.081	-5.2	99	0.00
49	Acenaphthylene	40.000	40.745	-1.9	98	0.00
50	Dimethylphthalate	40.000	40.454	-1.1	98	0.00
51	2,6-Dinitrotoluene	40.000	40.239	-0.6	98	0.00
52 C	Acenaphthene	40.000	40.548	-1.4	98	0.00
53	3-Nitroaniline	40.000	42.366	-5.9	101	0.00
54 P	2,4-Dinitrophenol	40.000	43.747	-9.4	102	0.00
55	Dibenzofuran	40.000	41.370	-3.4	102	0.00
56 P	4-Nitrophenol	40.000	45.542	-13.9	104	0.00
57	2,4-Dinitrotoluene	40.000	42.086	-5.2	100	0.00
58	Fluorene	40.000	41.202	-3.0	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.593	-4.0	99	0.00
60	Diethylphthalate	40.000	40.929	-2.3	100	0.00
61	4-Chlorophenyl-phenylether	40.000	40.308	-0.8	98	0.00
62	4-Nitroaniline	40.000	44.128	-10.3	108	0.00
63	Azobenzene	40.000	40.623	-1.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.429	-8.6	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.550	-1.4	101	0.00
67	4-Bromophenyl-phenylether	40.000	41.932	-4.8	103	0.00
68	Hexachlorobenzene	40.000	40.478	-1.2	100	0.00
69	Atrazine	40.000	38.385	4.0	97	0.00
70 C	Pentachlorophenol	40.000	43.581	-9.0	100	0.00
71	Phenanthrene	40.000	41.202	-3.0	102	0.00
72	Anthracene	40.000	41.239	-3.1	103	0.00
73	Carbazole	40.000	41.791	-4.5	104	0.00
74	Di-n-butylphthalate	40.000	41.909	-4.8	105	0.00
75 C	Fluoranthene	40.000	40.930	-2.3	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	38.555	3.6	106	0.00
78	Pyrene	40.000	40.534	-1.3	104	0.00
79 S	Terphenyl-d14	80.000	80.845	-1.1	104	0.00
80	Butylbenzylphthalate	40.000	41.195	-3.0	104	0.00
81	Benzo(a)anthracene	40.000	40.886	-2.2	105	0.00
82	3,3'-Dichlorobenzidine	40.000	38.839	2.9	100	0.00
83	Chrysene	40.000	41.146	-2.9	106	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.938	-2.3	106	0.00
85 c	Di-n-octyl phthalate	40.000	40.551	-1.4	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.400	-1.0	104	0.00
88	Benzo(b)fluoranthene	40.000	39.972	0.1	103	0.00
89	Benzo(k)fluoranthene	40.000	40.671	-1.7	105	0.00
90 C	Benzo(a)pyrene	40.000	40.442	-1.1	104	0.00
91	Dibenzo(a,h)anthracene	40.000	40.909	-2.3	106	0.00
92	Benzo(g,h,i)perylene	40.000	40.622	-1.6	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041422\
Data File : BG053150.D
Acq On : 14 Apr 2022 10:13
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 15 02:20:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 15:31:24 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0