


Data Path : Z:\HPCHEM1\BNA_G\Data\BG041615\
Data File : BG016456.D
Acq On : 16 Apr 2015 13:35
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Apr 14 14:00:56 2015



AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.0	18238	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	44.0	22285	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	56.5	28650	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	50666	PASS
199	198	5	9	6.6	3359	PASS
275	198	10	60	21.7	10986	PASS
365	198	1	100	1.9	940	PASS
441	442	0.01	24	19.0	6256	PASS
442	198	50	100	64.8	32850	PASS
443	442	15	24	19.2	6291	PASS