

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG No.: BG041624

Instrument ID: BNA_G Calibration Date/Time: 4/16/2024 11:36

Lab File ID: BG060909.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
n-Nitrosodimethylamine	0.887	0.832		-6.201	
Pyridine	1.485	1.247		-16.027	
2-Fluorophenol	1.201	1.069		-10.991	
Benzaldehyde	0.983	0.833		-15.259	
Aniline	2.250	1.839		-18.267	
Phenol-d6	1.782	1.510		-15.264	
Phenol	1.819	1.509		-17.042	20.0
bis(2-Chloroethyl) ether	1.468	1.186		-19.21	
2-Chlorophenol	1.361	1.211		-11.021	
1,2-Dichlorobenzene	1.415	1.295		-8.481	
1,3-Dichlorobenzene	1.415	1.286		-9.117	
1,4-Dichlorobenzene	1.444	1.320		-8.587	20.0
Benzyl Alcohol	1.444	1.311		-9.211	
2-Methylphenol	1.373	1.213		-11.653	
2,2-oxybis(1-Chloropropane)	3.314	2.832		-14.544	
Acetophenone	0.551	0.525		-4.719	
3+4-Methylphenols	1.880	1.649		-12.287	
n-Nitroso-di-n-propylamine	1.359	1.105	0.050	-18.69	
Nitrobenzene-d5	0.350	0.412		17.714	
Hexachloroethane	0.513	0.488		-4.873	
Nitrobenzene	0.352	0.409		16.193	
Isophorone	0.807	0.722		-10.533	
2-Nitrophenol	0.120	0.179		49.167	20.0
2,4-Dimethylphenol	0.324	0.296		-8.642	
bis(2-Chloroethoxy) methane	0.477	0.423		-11.321	
2,4-Dichlorophenol	0.298	0.324		8.725	20.0
1,2,4-Trichlorobenzene	0.337	0.353		4.748	
Benzoic acid	0.191	0.245		28.272	
Naphthalene	1.082	1.002		-7.394	
4-Chloroaniline	0.503	0.450		-10.537	
Hexachlorobutadiene	0.225	0.263		16.889	20.0
Caprolactam	0.118	0.108		-8.475	
4-Chloro-3-methylphenol	0.380	0.374		-1.579	20.0
2-Methylnaphthalene	0.795	0.745		-6.289	
Hexachlorocyclopentadiene	0.292	0.348	0.050	19.178	
2,4,6-Trichlorophenol	0.370	0.409		10.541	20.0
2-Fluorobiphenyl	1.363	1.297		-4.842	
2,4,5-Trichlorophenol	0.442	0.481		8.824	
1,1-Biphenyl	1.413	1.305		-7.643	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG No.: BG041624

Instrument ID: BNA_G Calibration Date/Time: 4/16/2024 11:36

Lab File ID: BG060909.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Chloronaphthalene	1.073	0.998		-6.99	
2-Nitroaniline	0.307	0.400		30.293	
Dimethylphthalate	1.581	1.445		-8.602	
Acenaphthylene	1.803	1.654		-8.264	
2,6-Dinitrotoluene	0.247	0.305		23.482	
3-Nitroaniline	0.277	0.323		16.606	
Acenaphthene	1.132	1.093		-3.445	20.0
2,4-Dinitrophenol	0.110	0.177	0.050	60.909	
4-Nitrophenol	0.231	0.250	0.050	8.225	
Dibenzofuran	1.786	1.642		-8.063	
2,4-Dinitrotoluene	0.320	0.442		38.125	
Diethylphthalate	1.590	1.427		-10.252	
4-Chlorophenyl-phenylether	0.758	0.743		-1.979	
Fluorene	1.433	1.335		-6.839	
4-Nitroaniline	0.289	0.336		16.263	
4,6-Dinitro-2-methylphenol	0.073	0.115		57.534	
n-Nitrosodiphenylamine	0.572	0.535		-6.469	20.0
Azobenzene	1.550	1.319		-14.903	
2,4,6-Tribromophenol	0.227	0.304		33.921	
4-Bromophenyl-phenylether	0.203	0.224		10.345	
Hexachlorobenzene	0.229	0.241		5.24	
Atrazine	0.200	0.186		-7.0	
Pentachlorophenol	0.150	0.172		14.667	20.0
Phenanthrene	1.040	0.950		-8.654	
Anthracene	1.056	0.979		-7.292	
Carbazole	0.931	0.842		-9.56	
Di-n-butylphthalate	1.161	1.015		-12.575	
Fluoranthene	1.262	1.165		-7.686	20.0
Benzidine	0.527	0.451		-14.421	
Pyrene	1.399	1.283		-8.292	
Terphenyl-d14	1.136	1.032		-9.155	
Butylbenzylphthalate	0.510	0.482		-5.49	
3,3-Dichlorobenzidine	0.476	0.467		-1.891	
Benzo(a)anthracene	1.410	1.296		-8.085	
Chrysene	1.359	1.227		-9.713	
Bis(2-ethylhexyl)phthalate	0.813	0.676		-16.851	
Di-n-octyl phthalate	1.325	1.187		-10.415	20.0
Benzo(b)fluoranthene	1.149	1.042		-9.312	
Benzo(k)fluoranthene	1.176	1.075		-8.588	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: _____

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG No.: BG041624

Instrument ID: BNA_G Calibration Date/Time: 4/16/2024 11:36

Lab File ID: BG060909.D Init. Calib. Date(s): _____

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): _____

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Benzo(a)pyrene	1.117	1.030		-7.789	20.0
Indeno(1,2,3-cd)pyrene	1.408	1.300		-7.67	
Dibenzo(a,h)anthracene	1.187	1.077		-9.267	
Benzo(g,h,i)perylene	1.155	1.054		-8.745	
1,2,4,5-Tetrachlorobenzene	0.575	0.615		6.957	
1,4-Dioxane	0.492	0.388		-21.138	20.0
2,3,4,6-Tetrachlorophenol	0.385	0.445		15.584	
1-Methylnaphthalene	0.751	0.702		-6.525	

All other compounds must meet a minimum RRF of 0.010.

Report of Analysis

Client:		Date Collected:	4/12/2024 12:00:00 AM
Project:		Date Received:	4/12/2024 12:00:00 AM
Client Sample ID:	PB160180TB	SDG No.:	BG041624
Lab Sample ID:	PB160180TB	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060910.D	1	4/12/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160203

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	10.3	U	10.3	80	100	ug/L
110-86-1	Pyridine	15.5	U	15.5	40	50	ug/L
100-52-7	Benzaldehyde	40	U	40	80	100	ug/L
62-53-3	Aniline	16	U	16	40	50	ug/L
108-95-2	Phenol	9.3	U	9.3	40	50	ug/L
111-44-4	bis(2-Chloroethyl)ether	11.9	U	11.9	40	50	ug/L
95-57-8	2-Chlorophenol	7.1	U	7.1	40	50	ug/L
95-50-1	1,2-Dichlorobenzene	8.8	U	8.8	40	50	ug/L
541-73-1	1,3-Dichlorobenzene	8	U	8	40	50	ug/L
106-46-7	1,4-Dichlorobenzene	8.4	U	8.4	40	50	ug/L
100-51-6	Benzyl Alcohol	16	U	16	80	100	ug/L
95-48-7	2-Methylphenol	11.3	U	11.3	40	50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	13.5	U	13.5	40	50	ug/L
98-86-2	Acetophenone	11	U	11	40	50	ug/L
65794-96-9	3+4-Methylphenols	11.5	U	11.5	80	100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	14.8	U	14.8	25	25	ug/L
67-72-1	Hexachloroethane	10.1	U	10.1	40	50	ug/L
98-95-3	Nitrobenzene	12.7	U	12.7	40	50	ug/L
78-59-1	Isophorone	11.4	U	11.4	40	50	ug/L
88-75-5	2-Nitrophenol	19.6	U	19.6	40	50	ug/L
105-67-9	2,4-Dimethylphenol	15.1	U	15.1	40	50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	10.2	U	10.2	40	50	ug/L
120-83-2	2,4-Dichlorophenol	8.8	U	8.8	40	50	ug/L
120-82-1	1,2,4-Trichlorobenzene	11	U	11	40	50	ug/L
65-85-0	Benzoic acid	54.7	U	54.7	80	100	ug/L
91-20-3	Naphthalene	10.2	U	10.2	40	50	ug/L
106-47-8	4-Chloroaniline	13	U	13	40	50	ug/L
87-68-3	Hexachlorobutadiene	12.7	U	12.7	40	50	ug/L

Report of Analysis

Client:		Date Collected:	4/12/2024 12:00:00 AM
Project:		Date Received:	4/12/2024 12:00:00 AM
Client Sample ID:	PB160180TB	SDG No.:	BG041624
Lab Sample ID:	PB160180TB	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060910.D	1	4/12/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160203

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	16.5	U	16.5	80	100	ug/L
59-50-7	4-Chloro-3-methylphenol	8.4	U	8.4	40	50	ug/L
91-57-6	2-Methylnaphthalene	11.3	U	11.3	40	50	ug/L
77-47-4	Hexachlorocyclopentadiene	50.2	U	50.2	80	100	ug/L
88-06-2	2,4,6-Trichlorophenol	8.9	U	8.9	40	50	ug/L
95-95-4	2,4,5-Trichlorophenol	10.1	U	10.1	40	50	ug/L
92-52-4	1,1-Biphenyl	9.1	U	9.1	40	50	ug/L
91-58-7	2-Chloronaphthalene	9.7	U	9.7	40	50	ug/L
88-74-4	2-Nitroaniline	14.1	U	14.1	40	50	ug/L
131-11-3	Dimethylphthalate	9.3	U	9.3	40	50	ug/L
208-96-8	Acenaphthylene	10.4	U	10.4	40	50	ug/L
606-20-2	2,6-Dinitrotoluene	12.4	U	12.4	40	50	ug/L
99-09-2	3-Nitroaniline	13.7	U	13.7	40	50	ug/L
83-32-9	Acenaphthene	8.1	U	8.1	40	50	ug/L
Total PAH	Total PAH	172.9	U	172.9	40	800	ug/L
51-28-5	2,4-Dinitrophenol	64.2	U	64.2	80	100	ug/L
100-02-7	4-Nitrophenol	20	U	20	80	100	ug/L
132-64-9	Dibenzofuran	9.3	U	9.3	40	50	ug/L
121-14-2	2,4-Dinitrotoluene	15.2	U	15.2	40	50	ug/L
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	27.6	U	27.6	40	100	ug/L
84-66-2	Diethylphthalate	10.4	U	10.4	40	50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	9.8	U	9.8	40	50	ug/L
86-73-7	Fluorene	9.6	U	9.6	40	50	ug/L
100-01-6	4-Nitroaniline	20.4	U	20.4	40	50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	30.7	U	30.7	80	100	ug/L
86-30-6	n-Nitrosodiphenylamine	8.9	U	8.9	40	50	ug/L
103-33-3	Azobenzene	12	U	12	40	50	ug/L
101-55-3	4-Bromophenyl-phenylether	9.5	U	9.5	40	50	ug/L
118-74-1	Hexachlorobenzene	11.4	U	11.4	40	50	ug/L
1912-24-9	Atrazine	12.6	U	12.6	40	50	ug/L

Report of Analysis

Client:		Date Collected:	4/12/2024 12:00:00 AM
Project:		Date Received:	4/12/2024 12:00:00 AM
Client Sample ID:	PB160180TB	SDG No.:	BG041624
Lab Sample ID:	PB160180TB	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100	Units:	mL
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060910.D	1	4/12/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160203

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	18.5	U	18.5	80	100	ug/L
85-01-8	Phenanthrene	8.9	U	8.9	40	50	ug/L
120-12-7	Anthracene	10.7	U	10.7	40	50	ug/L
86-74-8	Carbazole	11.5	U	11.5	40	50	ug/L
84-74-2	Di-n-butylphthalate	14.7	U	14.7	40	50	ug/L
206-44-0	Fluoranthene	12.9	U	12.9	40	50	ug/L
92-87-5	Benzidine	40.7	U	40.7	80	100	ug/L
129-00-0	Pyrene	10.6	U	10.6	40	50	ug/L
85-68-7	Butylbenzylphthalate	21	U	21	40	50	ug/L
91-94-1	3,3-Dichlorobenzidine	12.8	U	12.8	80	100	ug/L
56-55-3	Benzo(a)anthracene	9.4	U	9.4	40	50	ug/L
218-01-9	Chrysene	8.6	U	8.6	40	50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	18.9	U	18.9	40	50	ug/L
117-84-0	Di-n-octyl phthalate	25	U	25	80	100	ug/L
205-99-2	Benzo(b)fluoranthene	11.4	U	11.4	40	50	ug/L
207-08-9	Benzo(k)fluoranthene	11.9	U	11.9	40	50	ug/L
50-32-8	Benzo(a)pyrene	16.7	U	16.7	40	50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	10.2	U	10.2	40	50	ug/L
53-70-3	Dibenzo(a,h)anthracene	11.5	U	11.5	40	50	ug/L
191-24-2	Benzo(g,h,i)perylene	11.8	U	11.8	40	50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	10.9	U	10.9	40	50	ug/L
123-91-1	1,4-Dioxane	12.5	U	12.5	40	50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	7.9	U	7.9	40	50	ug/L
90-12-0	1-Methylnaphthalene	8.6	U	8.6	40	50	ug/L
SURROGATES							
367-12-4	2-Fluorophenol	125.864		10-139		84	SPK: -150
13127-88-3	Phenol-d6	109.588		10-134		73	SPK: -150
4165-60-0	Nitrobenzene-d5	94.035		49-133		94	SPK: -100
321-60-8	2-Fluorobiphenyl	77.471		52-132		77	SPK: -100

Report of Analysis

Client:		Date Collected:	4/12/2024 12:00:00 AM
Project:		Date Received:	4/12/2024 12:00:00 AM
Client Sample ID:	PB160180TB	SDG No.:	BG041624
Lab Sample ID:	PB160180TB	Matrix:	water
Analytical Method:	8270E	% Moisture:	100
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060910.D	1	4/12/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160203

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	178.592		32-145		119	SPK: -150
1718-51-0	Terphenyl-d14	80.461		36-145		80	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	46318	7.944				
1146-65-2	Naphthalene-d8	181618	10.741				
15067-26-2	Acenaphthene-d10	144480	14.577				
1517-22-2	Phenanthrene-d10	382278	17.327				
1719-03-5	Chrysene-d12	361165	21.587				
1520-96-3	Perylene-d12	417510	24.718				

Report of Analysis

Client:		Date Collected:	4/15/2024 12:00:00 AM
Project:		Date Received:	4/15/2024 12:00:00 AM
Client Sample ID:	PB160246BS	SDG No.:	BG041624
Lab Sample ID:	PB160246BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060911.D	1	4/15/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	1300		86.8	270	330	ug/Kg
110-86-1	Pyridine	770		79.9	130	170	ug/Kg
100-52-7	Benzaldehyde	260	J	180	270	330	ug/Kg
62-53-3	Aniline	800		96.9	130	170	ug/Kg
108-95-2	Phenol	1400		82.9	130	170	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	1200		83.7	130	170	ug/Kg
95-57-8	2-Chlorophenol	1400		83.5	130	170	ug/Kg
95-50-1	1,2-Dichlorobenzene	1400		84.6	130	170	ug/Kg
541-73-1	1,3-Dichlorobenzene	1300		86	130	170	ug/Kg
106-46-7	1,4-Dichlorobenzene	1400		86.5	130	170	ug/Kg
100-51-6	Benzyl Alcohol	1400		83	270	330	ug/Kg
95-48-7	2-Methylphenol	1300		80.6	130	170	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	1300		90.9	130	170	ug/Kg
98-86-2	Acetophenone	1300		86.9	130	170	ug/Kg
65794-96-9	3+4-Methylphenols	1300		79.8	270	330	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	1200		40.3	80	80	ug/Kg
67-72-1	Hexachloroethane	1400		83	130	170	ug/Kg
98-95-3	Nitrobenzene	1400		90.8	130	170	ug/Kg
78-59-1	Isophorone	1200		84.6	130	170	ug/Kg
88-75-5	2-Nitrophenol	1800		94.5	130	170	ug/Kg
105-67-9	2,4-Dimethylphenol	1100		93.2	130	170	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	1200		85.8	130	170	ug/Kg
120-83-2	2,4-Dichlorophenol	1600		75.5	130	170	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	1500		85.4	130	170	ug/Kg
65-85-0	Benzoic acid	1600		240	270	330	ug/Kg
91-20-3	Naphthalene	1300		82.6	130	170	ug/Kg
106-47-8	4-Chloroaniline	580		82.6	130	170	ug/Kg
87-68-3	Hexachlorobutadiene	1600		83.3	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/15/2024 12:00:00 AM
Project:		Date Received:	4/15/2024 12:00:00 AM
Client Sample ID:	PB160246BS	SDG No.:	BG041624
Lab Sample ID:	PB160246BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060911.D	1	4/15/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	1300		86.8	270	330	ug/Kg
59-50-7	4-Chloro-3-methylphenol	1400		77.5	130	170	ug/Kg
91-57-6	2-Methylnaphthalene	1300		82.5	130	170	ug/Kg
77-47-4	Hexachlorocyclopentadiene	3000	E	160	270	330	ug/Kg
88-06-2	2,4,6-Trichlorophenol	1600		71.4	130	170	ug/Kg
95-95-4	2,4,5-Trichlorophenol	1500		74	130	170	ug/Kg
92-52-4	1,1-Biphenyl	1300		87.4	130	170	ug/Kg
91-58-7	2-Chloronaphthalene	1300		83.3	130	170	ug/Kg
88-74-4	2-Nitroaniline	1500		95	130	170	ug/Kg
131-11-3	Dimethylphthalate	1300		81.7	130	170	ug/Kg
208-96-8	Acenaphthylene	1300		86.5	130	170	ug/Kg
606-20-2	2,6-Dinitrotoluene	1500		83.2	130	170	ug/Kg
99-09-2	3-Nitroaniline	1000		89.2	130	170	ug/Kg
83-32-9	Acenaphthene	1500		81.1	130	170	ug/Kg
Total PAH	Total PAH	21800		1325.1	130	2720	ug/Kg
51-28-5	2,4-Dinitrophenol	4800	E	240	270	330	ug/Kg
100-02-7	4-Nitrophenol	3100	E	120	270	330	ug/Kg
132-64-9	Dibenzofuran	1300		84.4	130	170	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	3200		169.4	130	340	ug/Kg
121-14-2	2,4-Dinitrotoluene	1700		86.2	130	170	ug/Kg
84-66-2	Diethylphthalate	1200		80.1	130	170	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	1400		85.6	130	170	ug/Kg
86-73-7	Fluorene	1300		85.5	130	170	ug/Kg
100-01-6	4-Nitroaniline	1500		110	130	170	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	1900		120	270	330	ug/Kg
86-30-6	n-Nitrosodiphenylamine	1300		81.6	130	170	ug/Kg
103-33-3	Azobenzene	1200		79.6	130	170	ug/Kg
101-55-3	4-Bromophenyl-phenylether	1500		78.9	130	170	ug/Kg
118-74-1	Hexachlorobenzene	1500		85	130	170	ug/Kg
1912-24-9	Atrazine	1400		91.4	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/15/2024 12:00:00 AM
Project:		Date Received:	4/15/2024 12:00:00 AM
Client Sample ID:	PB160246BS	SDG No.:	BG041624
Lab Sample ID:	PB160246BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060911.D	1	4/15/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	3100	E	77.3	270	330	ug/Kg
85-01-8	Phenanthrene	1300		84	130	170	ug/Kg
120-12-7	Anthracene	1300		84.4	130	170	ug/Kg
86-74-8	Carbazole	1300		80.3	130	170	ug/Kg
84-74-2	Di-n-butylphthalate	1200		84.3	130	170	ug/Kg
206-44-0	Fluoranthene	1300		81.7	130	170	ug/Kg
92-87-5	Benzidine	700		160	270	330	ug/Kg
129-00-0	Pyrene	1300		83	130	170	ug/Kg
85-68-7	Butylbenzylphthalate	1300		96.8	130	170	ug/Kg
91-94-1	3,3-Dichlorobenzidine	1100		98.6	270	330	ug/Kg
56-55-3	Benzo(a)anthracene	1300		80.7	130	170	ug/Kg
218-01-9	Chrysene	1300		79.5	130	170	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	1200		91	130	170	ug/Kg
117-84-0	Di-n-octyl phthalate	1300		110	270	330	ug/Kg
205-99-2	Benzo(b)fluoranthene	1500		81.1	130	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	1400		82.6	130	170	ug/Kg
50-32-8	Benzo(a)pyrene	1400		93	130	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1500		78.1	130	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1400		81.2	130	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1400		80.1	130	170	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	1500		86.8	130	170	ug/Kg
123-91-1	1,4-Dioxane	930		110	130	170	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	1600		74.7	130	170	ug/Kg
90-12-0	1-Methylnaphthalene	1200		83.2	170	170	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	122.006		18-112		81	SPK: -150
13127-88-3	Phenol-d6	116.391		15-107		78	SPK: -150
4165-60-0	Nitrobenzene-d5	91.516		18-107		92	SPK: -100
321-60-8	2-Fluorobiphenyl	76.161		20-109		76	SPK: -100

Report of Analysis

Client:		Date Collected:	4/15/2024 12:00:00 AM
Project:		Date Received:	4/15/2024 12:00:00 AM
Client Sample ID:	PB160246BS	SDG No.:	BG041624
Lab Sample ID:	PB160246BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060911.D	1	4/15/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160246

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	163.998		10-110		109	SPK: -150
1718-51-0	Terphenyl-d14	78.982		14-112		79	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	57254	7.947				
1146-65-2	Naphthalene-d8	244157	10.743				
15067-26-2	Acenaphthene-d10	192896	14.58				
1517-22-2	Phenanthrene-d10	462252	17.324				
1719-03-5	Chrysene-d12	415335	21.589				
1520-96-3	Perylene-d12	451610	24.721				



Client:				Date Collected:	4/10/2024 12:00:00 AM	
Project:				Date Received:	4/10/2024 12:00:00 AM	
Client Sample ID:	PB160120BL			SDG No.:	BG041624	
Lab Sample ID:	PB160120BL			Matrix:	Solid	
Analytical Method:	8270E			% Moisture:	0	
Sample Wt/Vol:	30.02	Units:	g	Final Vol:	1000	uL
Soil Aliquot Vol:			uL	Test:		
Extraction Type :	Decanted :			Level :	LOW	
Injection Volume :	GPC Factor :			GPC Cleanup :	PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060912.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	86.7	U	86.7	270	330	ug/Kg
110-86-1	Pyridine	79.8	U	79.8	130	170	ug/Kg
100-52-7	Benzaldehyde	180	U	180	270	330	ug/Kg
62-53-3	Aniline	96.8	U	96.8	130	170	ug/Kg
108-95-2	Phenol	82.8	U	82.8	130	170	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	83.6	U	83.6	130	170	ug/Kg
95-57-8	2-Chlorophenol	83.4	U	83.4	130	170	ug/Kg
95-50-1	1,2-Dichlorobenzene	84.5	U	84.5	130	170	ug/Kg
541-73-1	1,3-Dichlorobenzene	85.9	U	85.9	130	170	ug/Kg
106-46-7	1,4-Dichlorobenzene	86.4	U	86.4	130	170	ug/Kg
100-51-6	Benzyl Alcohol	82.9	U	82.9	270	330	ug/Kg
95-48-7	2-Methylphenol	80.5	U	80.5	130	170	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	90.8	U	90.8	130	170	ug/Kg
98-86-2	Acetophenone	86.8	U	86.8	130	170	ug/Kg
65794-96-9	3+4-Methylphenols	79.7	U	79.7	270	330	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	40.3	U	40.3	79.9	79.9	ug/Kg
67-72-1	Hexachloroethane	82.9	U	82.9	130	170	ug/Kg
98-95-3	Nitrobenzene	90.7	U	90.7	130	170	ug/Kg
78-59-1	Isophorone	84.5	U	84.5	130	170	ug/Kg
88-75-5	2-Nitrophenol	94.4	U	94.4	130	170	ug/Kg
105-67-9	2,4-Dimethylphenol	93.1	U	93.1	130	170	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	85.7	U	85.7	130	170	ug/Kg
120-83-2	2,4-Dichlorophenol	75.4	U	75.4	130	170	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	85.3	U	85.3	130	170	ug/Kg
65-85-0	Benzoic acid	240	U	240	270	330	ug/Kg
91-20-3	Naphthalene	82.5	U	82.5	130	170	ug/Kg
106-47-8	4-Chloroaniline	82.5	U	82.5	130	170	ug/Kg
87-68-3	Hexachlorobutadiene	83.2	U	83.2	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BL	SDG No.:	BG041624
Lab Sample ID:	PB160120BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060912.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	86.7	U	86.7	270	330	ug/Kg
59-50-7	4-Chloro-3-methylphenol	77.4	U	77.4	130	170	ug/Kg
91-57-6	2-Methylnaphthalene	82.4	U	82.4	130	170	ug/Kg
77-47-4	Hexachlorocyclopentadiene	160	U	160	270	330	ug/Kg
88-06-2	2,4,6-Trichlorophenol	71.4	U	71.4	130	170	ug/Kg
95-95-4	2,4,5-Trichlorophenol	74	U	74	130	170	ug/Kg
92-52-4	1,1-Biphenyl	87.3	U	87.3	130	170	ug/Kg
91-58-7	2-Chloronaphthalene	83.2	U	83.2	130	170	ug/Kg
88-74-4	2-Nitroaniline	94.9	U	94.9	130	170	ug/Kg
131-11-3	Dimethylphthalate	81.6	U	81.6	130	170	ug/Kg
208-96-8	Acenaphthylene	86.4	U	86.4	130	170	ug/Kg
606-20-2	2,6-Dinitrotoluene	83.1	U	83.1	130	170	ug/Kg
99-09-2	3-Nitroaniline	89.1	U	89.1	130	170	ug/Kg
83-32-9	Acenaphthene	81	U	81	130	170	ug/Kg
Total PAH	Total PAH	1323.5	U	1323.5	130	2720	ug/Kg
51-28-5	2,4-Dinitrophenol	240	U	240	270	330	ug/Kg
100-02-7	4-Nitrophenol	120	U	120	270	330	ug/Kg
132-64-9	Dibenzofuran	84.3	U	84.3	130	170	ug/Kg
121-14-2	2,4-Dinitrotoluene	86.1	U	86.1	130	170	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	169.2	U	169.2	130	340	ug/Kg
84-66-2	Diethylphthalate	80	U	80	130	170	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	85.5	U	85.5	130	170	ug/Kg
86-73-7	Fluorene	85.4	U	85.4	130	170	ug/Kg
100-01-6	4-Nitroaniline	110	U	110	130	170	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	120	U	120	270	330	ug/Kg
86-30-6	n-Nitrosodiphenylamine	81.5	U	81.5	130	170	ug/Kg
103-33-3	Azobenzene	79.5	U	79.5	130	170	ug/Kg
101-55-3	4-Bromophenyl-phenylether	78.8	U	78.8	130	170	ug/Kg
118-74-1	Hexachlorobenzene	84.9	U	84.9	130	170	ug/Kg
1912-24-9	Atrazine	91.3	U	91.3	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BL	SDG No.:	BG041624
Lab Sample ID:	PB160120BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060912.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	77.2	U	77.2	270	330	ug/Kg
85-01-8	Phenanthrene	83.9	U	83.9	130	170	ug/Kg
120-12-7	Anthracene	84.3	U	84.3	130	170	ug/Kg
86-74-8	Carbazole	80.2	U	80.2	130	170	ug/Kg
84-74-2	Di-n-butylphthalate	84.2	U	84.2	130	170	ug/Kg
206-44-0	Fluoranthene	81.6	U	81.6	130	170	ug/Kg
92-87-5	Benzidine	160	U	160	270	330	ug/Kg
129-00-0	Pyrene	82.9	U	82.9	130	170	ug/Kg
85-68-7	Butylbenzylphthalate	96.7	U	96.7	130	170	ug/Kg
91-94-1	3,3-Dichlorobenzidine	98.5	U	98.5	270	330	ug/Kg
56-55-3	Benzo(a)anthracene	80.6	U	80.6	130	170	ug/Kg
218-01-9	Chrysene	79.4	U	79.4	130	170	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	90.9	U	90.9	130	170	ug/Kg
117-84-0	Di-n-octyl phthalate	110	U	110	270	330	ug/Kg
205-99-2	Benzo(b)fluoranthene	81	U	81	130	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	82.5	U	82.5	130	170	ug/Kg
50-32-8	Benzo(a)pyrene	92.9	U	92.9	130	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	78	U	78	130	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	81.1	U	81.1	130	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	80	U	80	130	170	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	86.7	U	86.7	130	170	ug/Kg
123-91-1	1,4-Dioxane	110	U	110	130	170	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	74.7	U	74.7	130	170	ug/Kg
90-12-0	1-Methylnaphthalene	83.1	U	83.1	170	170	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	118.095		18-112		79	SPK: -150
13127-88-3	Phenol-d6	107.068		15-107		71	SPK: -150
4165-60-0	Nitrobenzene-d5	95.633		18-107		96	SPK: -100
321-60-8	2-Fluorobiphenyl	75.073		20-109		75	SPK: -100

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BL	SDG No.:	BG041624
Lab Sample ID:	PB160120BL	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.02 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060912.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	168.577	*	10-110		112	SPK: -150
1718-51-0	Terphenyl-d14	76.691		14-112		77	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	51189	7.946				
1146-65-2	Naphthalene-d8	199793	10.743				
15067-26-2	Acenaphthene-d10	160907	14.574				
1517-22-2	Phenanthrene-d10	439391	17.324				
1719-03-5	Chrysene-d12	424861	21.583				
1520-96-3	Perylene-d12	484252	24.715				
TENTATIVE IDENTIFIED COMPOUNDS							
000629-59-4	Tetradecane	83.6	J			24.827	
000629-62-9	Pentadecane	75.9	J			25.937	
000629-92-5	Nonadecane	70.3	J			27.277	

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BS	SDG No.:	BG041624
Lab Sample ID:	PB160120BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	
	Decanted :	Level :	LOW
Injection Volume :		GPC Factor :	
		GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060913.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
62-75-9	n-Nitrosodimethylamine	1200		86.8	270	330	ug/Kg
110-86-1	Pyridine	750		79.9	130	170	ug/Kg
100-52-7	Benzaldehyde	260	J	180	270	330	ug/Kg
62-53-3	Aniline	780		96.9	130	170	ug/Kg
108-95-2	Phenol	1400		82.9	130	170	ug/Kg
111-44-4	bis(2-Chloroethyl)ether	1200		83.7	130	170	ug/Kg
95-57-8	2-Chlorophenol	1400		83.5	130	170	ug/Kg
95-50-1	1,2-Dichlorobenzene	1300		84.6	130	170	ug/Kg
541-73-1	1,3-Dichlorobenzene	1300		86	130	170	ug/Kg
106-46-7	1,4-Dichlorobenzene	1300		86.5	130	170	ug/Kg
100-51-6	Benzyl Alcohol	1300		83	270	330	ug/Kg
95-48-7	2-Methylphenol	1300		80.6	130	170	ug/Kg
108-60-1	2,2-oxybis(1-Chloropropane)	1300		90.9	130	170	ug/Kg
98-86-2	Acetophenone	1300		86.9	130	170	ug/Kg
65794-96-9	3+4-Methylphenols	1300		79.8	270	330	ug/Kg
621-64-7	n-Nitroso-di-n-propylamine	1200		40.3	80	80	ug/Kg
67-72-1	Hexachloroethane	1400		83	130	170	ug/Kg
98-95-3	Nitrobenzene	1400		90.8	130	170	ug/Kg
78-59-1	Isophorone	1300		84.6	130	170	ug/Kg
88-75-5	2-Nitrophenol	1800		94.5	130	170	ug/Kg
105-67-9	2,4-Dimethylphenol	1100		93.2	130	170	ug/Kg
111-91-1	bis(2-Chloroethoxy)methane	1200		85.8	130	170	ug/Kg
120-83-2	2,4-Dichlorophenol	1600		75.5	130	170	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	1500		85.4	130	170	ug/Kg
65-85-0	Benzoic acid	1600		240	270	330	ug/Kg
91-20-3	Naphthalene	1300		82.6	130	170	ug/Kg
106-47-8	4-Chloroaniline	570		82.6	130	170	ug/Kg
87-68-3	Hexachlorobutadiene	1600		83.3	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BS	SDG No.:	BG041624
Lab Sample ID:	PB160120BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060913.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
105-60-2	Caprolactam	1300		86.8	270	330	ug/Kg
59-50-7	4-Chloro-3-methylphenol	1400		77.5	130	170	ug/Kg
91-57-6	2-Methylnaphthalene	1300		82.5	130	170	ug/Kg
77-47-4	Hexachlorocyclopentadiene	3000	E	160	270	330	ug/Kg
88-06-2	2,4,6-Trichlorophenol	1600		71.4	130	170	ug/Kg
95-95-4	2,4,5-Trichlorophenol	1500		74	130	170	ug/Kg
92-52-4	1,1-Biphenyl	1300		87.4	130	170	ug/Kg
91-58-7	2-Chloronaphthalene	1300		83.3	130	170	ug/Kg
88-74-4	2-Nitroaniline	1600		95	130	170	ug/Kg
131-11-3	Dimethylphthalate	1300		81.7	130	170	ug/Kg
208-96-8	Acenaphthylene	1300		86.5	130	170	ug/Kg
606-20-2	2,6-Dinitrotoluene	1500		83.2	130	170	ug/Kg
99-09-2	3-Nitroaniline	1000		89.2	130	170	ug/Kg
83-32-9	Acenaphthene	1500		81.1	130	170	ug/Kg
Total PAH	Total PAH	22000		1325.1	130	2720	ug/Kg
51-28-5	2,4-Dinitrophenol	4800	E	240	270	330	ug/Kg
100-02-7	4-Nitrophenol	3000	E	120	270	330	ug/Kg
132-64-9	Dibenzofuran	1300		84.4	130	170	ug/Kg
121-14-2	2,4-Dinitrotoluene	1600		86.2	130	170	ug/Kg
25321-14-6	2,4-Dinitrotoluene/2,6-Dinitrotoluene (mixture)	3100		169.4	130	340	ug/Kg
84-66-2	Diethylphthalate	1200		80.1	130	170	ug/Kg
7005-72-3	4-Chlorophenyl-phenylether	1300		85.6	130	170	ug/Kg
86-73-7	Fluorene	1300		85.5	130	170	ug/Kg
100-01-6	4-Nitroaniline	1500		110	130	170	ug/Kg
534-52-1	4,6-Dinitro-2-methylphenol	1900		120	270	330	ug/Kg
86-30-6	n-Nitrosodiphenylamine	1300		81.6	130	170	ug/Kg
103-33-3	Azobenzene	1200		79.6	130	170	ug/Kg
101-55-3	4-Bromophenyl-phenylether	1500		78.9	130	170	ug/Kg
118-74-1	Hexachlorobenzene	1500		85	130	170	ug/Kg
1912-24-9	Atrazine	1400		91.4	130	170	ug/Kg

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BS	SDG No.:	BG041624
Lab Sample ID:	PB160120BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060913.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
87-86-5	Pentachlorophenol	3100	E	77.3	270	330	ug/Kg
85-01-8	Phenanthrene	1300		84	130	170	ug/Kg
120-12-7	Anthracene	1300		84.4	130	170	ug/Kg
86-74-8	Carbazole	1300		80.3	130	170	ug/Kg
84-74-2	Di-n-butylphthalate	1200		84.3	130	170	ug/Kg
206-44-0	Fluoranthene	1300		81.7	130	170	ug/Kg
92-87-5	Benzidine	680		160	270	330	ug/Kg
129-00-0	Pyrene	1300		83	130	170	ug/Kg
85-68-7	Butylbenzylphthalate	1300		96.8	130	170	ug/Kg
91-94-1	3,3-Dichlorobenzidine	1100		98.6	270	330	ug/Kg
56-55-3	Benzo(a)anthracene	1300		80.7	130	170	ug/Kg
218-01-9	Chrysene	1300		79.5	130	170	ug/Kg
117-81-7	Bis(2-ethylhexyl)phthalate	1200		91	130	170	ug/Kg
117-84-0	Di-n-octyl phthalate	1300		110	270	330	ug/Kg
205-99-2	Benzo(b)fluoranthene	1500		81.1	130	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	1500		82.6	130	170	ug/Kg
50-32-8	Benzo(a)pyrene	1400		93	130	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1500		78.1	130	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1500		81.2	130	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1400		80.1	130	170	ug/Kg
95-94-3	1,2,4,5-Tetrachlorobenzene	1500		86.8	130	170	ug/Kg
123-91-1	1,4-Dioxane	870		110	130	170	ug/Kg
58-90-2	2,3,4,6-Tetrachlorophenol	1600		74.7	130	170	ug/Kg
90-12-0	1-Methylnaphthalene	1200		83.2	170	170	ug/Kg
SURROGATES							
367-12-4	2-Fluorophenol	128.399		18-112		86	SPK: -150
13127-88-3	Phenol-d6	117.342		15-107		78	SPK: -150
4165-60-0	Nitrobenzene-d5	90.217		18-107		90	SPK: -100
321-60-8	2-Fluorobiphenyl	76.506		20-109		77	SPK: -100

Report of Analysis

Client:		Date Collected:	4/10/2024 12:00:00 AM
Project:		Date Received:	4/10/2024 12:00:00 AM
Client Sample ID:	PB160120BS	SDG No.:	BG041624
Lab Sample ID:	PB160120BS	Matrix:	Solid
Analytical Method:	8270E	% Moisture:	0
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	
Extraction Type :	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	GPC Cleanup :	PH :

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG060913.D	1	4/10/2024 12:00:00 AM	4/16/2024 12:00:00 AM	PB160120

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
118-79-6	2,4,6-Tribromophenol	158.897		10-110		106	SPK: -150
1718-51-0	Terphenyl-d14	76.379		14-112		76	SPK: -100
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	71474	7.948				
1146-65-2	Naphthalene-d8	308787	10.744				
15067-26-2	Acenaphthene-d10	239703	14.581				
1517-22-2	Phenanthrene-d10	563421	17.325				
1719-03-5	Chrysene-d12	513117	21.591				
1520-96-3	Perylene-d12	545770	24.722				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: BG041624

Client:

Sample ID	Client ID	Parameter	Concentration	C	MDL	RDL	Units
Client ID :							

Total Concentration:



6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____
Lab Code: CHEM Case No.: BG032924 SAS No.: BG032924 SDG No.: BG032924
Instrument ID: _____ Calibration Date(s): 3/29/2024 : _____
Calibration Time(s): 10:48 _____

LAB FILE ID:		RRF2.5 = BG060699.D			RRF5.0 = BG060700.D		RRF010 = BG060701.D	
		RRF020 = BG060702.D			RRF040 = BG060703.D		RRF050 = BG060704.D	
COMPOUND	RRF2.5	RRF5.0	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
n-Nitrosodimethylamine		0.894	0.864	0.865	0.929	0.915	0.887	5.2
Pyridine		1.490	1.522	1.496	1.602	1.565	1.485	8.1
2-Fluorophenol		1.171	1.161	1.269	1.204	1.164	1.201	3.5
Benzaldehyde			1.121	1.113	0.947	0.875	0.983	12.9
Aniline		2.123	2.225	2.424	2.291	2.196	2.250	4.2
Phenol-d6		1.718	1.774	1.859	1.813	1.738	1.782	2.8
Phenol		1.712	1.798	1.902	1.875	1.764	1.819	3.8
bis(2-Chloroethyl)ether		1.477	1.480	1.546	1.467	1.414	1.468	2.8
2-Chlorophenol		1.265	1.350	1.428	1.376	1.348	1.361	4.0
1,2-Dichlorobenzene		1.455	1.406	1.516	1.396	1.323	1.415	4.2
1,3-Dichlorobenzene		1.527	1.396	1.447	1.391	1.357	1.415	4.0
1,4-Dichlorobenzene		1.474	1.476	1.484	1.421	1.390	1.444	2.7
Benzyl Alcohol		1.345	1.411	1.500	1.481	1.421	1.444	3.8
2-Methylphenol		1.285	1.361	1.457	1.386	1.325	1.373	4.2
2,2-oxybis(1-Chloropropane)		3.269	3.356	3.489	3.291	3.139	3.314	3.2
Acetophenone		0.539	0.534	0.597	0.540	0.545	0.551	4.4
3+4-Methylphenols		1.683	1.831	2.015	1.911	1.849	1.880	5.8
n-Nitroso-di-n-propylamine	1.345	1.260	1.367	1.452	1.377	1.318	1.359	4.2
Nitrobenzene-d5			0.270	0.331	0.341	0.371	0.350	13.6
Hexachloroethane		0.495	0.533	0.533	0.493	0.493	0.513	3.8
Nitrobenzene		0.253	0.304	0.348	0.363	0.386	0.352	16.1
Isophorone		0.765	0.787	0.853	0.804	0.813	0.807	3.8
2-Nitrophenol		0.079	0.083	0.111	0.124	0.137	0.120	25.1
2,4-Dimethylphenol		0.310	0.321	0.347	0.318	0.323	0.324	4.1
bis(2-Chloroethoxy)methane		0.469	0.467	0.496	0.469	0.482	0.477	2.5
2,4-Dichlorophenol		0.253	0.277	0.311	0.306	0.313	0.298	8.0
1,2,4-Trichlorobenzene		0.325	0.326	0.351	0.331	0.339	0.337	3.6
Benzoic acid			0.115	0.163	0.199	0.212	0.191	23.5
Naphthalene		1.074	1.091	1.136	1.077	1.065	1.082	2.9
4-Chloroaniline		0.487	0.493	0.528	0.508	0.502	0.503	2.9
Hexachlorobutadiene		0.213	0.223	0.234	0.218	0.226	0.225	3.6
Caprolactam		0.112	0.119	0.127	0.121	0.115	0.118	4.6
4-Chloro-3-methylphenol		0.370	0.365	0.394	0.392	0.377	0.380	3.1
2-Methylnaphthalene		0.808	0.811	0.828	0.783	0.772	0.795	3.5
Hexachlorocyclopentadiene		0.256	0.247	0.295	0.286	0.314	0.292	10.9
2,4,6-Trichlorophenol		0.302	0.338	0.381	0.372	0.398	0.370	10.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1



6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____
Lab Code: CHEM Case No.: BG032924 SAS No.: BG032924 SDG No.: BG032924
Instrument ID: _____ Calibration Date(s): 3/29/2024 : _____
Calibration Time(s): 10:48 _____

LAB FILE ID:		RRF2.5 = BG060699.D			RRF5.0 = BG060700.D		RRF010 = BG060701.D	
		RRF020 = BG060702.D			RRF040 = BG060703.D		RRF050 = BG060704.D	
COMPOUND	RRF2.5	RRF5.0	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorobiphenyl		1.383	1.377	1.463	1.316	1.359	1.363	4.1
2,4,5-Trichlorophenol		0.380	0.407	0.461	0.445	0.465	0.442	7.9
1,1-Biphenyl		1.417	1.415	1.496	1.364	1.405	1.413	3.0
2-Chloronaphthalene		1.036	1.062	1.133	1.045	1.083	1.073	3.0
2-Nitroaniline		0.201	0.231	0.289	0.323	0.356	0.307	22.9
Dimethylphthalate		1.570	1.576	1.696	1.565	1.570	1.581	3.6
Acenaphthylene		1.760	1.812	1.900	1.759	1.818	1.803	2.9
2,6-Dinitrotoluene		0.154	0.188	0.251	0.264	0.284	0.247	22.3
3-Nitroaniline		0.185	0.226	0.283	0.300	0.309	0.277	18.7
Acenaphthene		1.093	1.150	1.191	1.108	1.137	1.132	3.1
2,4-Dinitrophenol			0.092	0.099	0.108	0.118	0.110	14.8
4-Nitrophenol			0.180	0.218	0.236	0.244	0.231	12.4
Dibenzofuran		1.813	1.854	1.867	1.732	1.780	1.786	3.6
2,4-Dinitrotoluene		0.192	0.242	0.314	0.346	0.374	0.320	24.0
Diethylphthalate		1.622	1.665	1.690	1.557	1.558	1.590	4.5
4-Chlorophenyl-phenylether		0.794	0.789	0.798	0.738	0.742	0.758	4.7
Fluorene		1.482	1.514	1.521	1.412	1.398	1.433	5.3
4-Nitroaniline		0.214	0.255	0.299	0.303	0.312	0.289	13.8
4,6-Dinitro-2-methylphenol			0.048	0.063	0.072	0.081	0.073	21.8
n-Nitrosodiphenylamine		0.539	0.564	0.586	0.568	0.587	0.572	3.1
Azobenzene		1.516	1.599	1.667	1.536	1.532	1.550	4.1
2,4,6-Tribromophenol		0.187	0.214	0.243	0.226	0.236	0.227	9.2
4-Bromophenyl-phenylether		0.187	0.189	0.215	0.202	0.205	0.203	5.7
Hexachlorobenzene		0.222	0.219	0.239	0.225	0.233	0.229	3.4
Atrazine		0.195	0.209	0.213	0.200	0.197	0.200	4.4
Pentachlorophenol		0.108	0.133	0.155	0.152	0.164	0.150	14.9
Phenanthrene		1.052	1.057	1.109	1.010	1.030	1.040	3.9
Anthracene		1.060	1.060	1.128	1.030	1.050	1.056	3.7
Carbazole		0.957	0.949	0.999	0.906	0.919	0.931	4.5
Di-n-butylphthalate		1.134	1.187	1.258	1.153	1.154	1.161	4.5
Fluoranthene		1.343	1.329	1.337	1.216	1.228	1.262	5.8
Benzidine		0.498	0.492	0.525	0.596	0.534	0.527	6.5
Pyrene		1.411	1.408	1.525	1.368	1.366	1.399	4.5
Terphenyl-d14		1.221	1.230	1.271	1.104	1.073	1.136	9.3
Butylbenzylphthalate		0.422	0.460	0.539	0.520	0.542	0.510	9.8
3,3-Dichlorobenzidine		0.430	0.445	0.510	0.493	0.488	0.476	5.9

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1



6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: _____
Lab Code: CHEM Case No.: BG032924 SAS No.: BG032924 SDG No.: BG032924
Instrument ID: _____ Calibration Date(s): 3/29/2024 : _____
Calibration Time(s): 10:48 _____

LAB FILE ID:		RRF2.5 = BG060699.D			RRF5.0 = BG060700.D		RRF010 = BG060701.D	
		RRF020 = BG060702.D			RRF040 = BG060703.D		RRF050 = BG060704.D	
COMPOUND	RRF2.5	RRF5.0	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Benzo(a)anthracene		1.385	1.406	1.500	1.401	1.400	1.410	3.1
Chrysene		1.350	1.348	1.440	1.345	1.349	1.359	2.8
Bis(2-ethylhexyl)phthalate		0.705	0.769	0.866	0.836	0.847	0.813	7.0
Di-n-octyl phthalate		1.077	1.191	1.382	1.389	1.413	1.325	10.2
Benzo(b)fluoranthene		1.132	1.130	1.198	1.141	1.143	1.149	2.2
Benzo(k)fluoranthene		1.155	1.142	1.236	1.181	1.172	1.176	3.1
Benzo(a)pyrene		1.073	1.092	1.156	1.122	1.119	1.117	2.8
Indeno(1,2,3-cd)pyrene		1.303	1.361	1.463	1.393	1.461	1.408	4.3
Dibenzo(a,h)anthracene		1.092	1.144	1.234	1.173	1.235	1.187	4.7
Benzo(g,h,i)perylene		1.091	1.120	1.190	1.139	1.189	1.155	3.5
1,2,4,5-Tetrachlorobenzene		0.553	0.550	0.595	0.555	0.589	0.575	3.8
1,4-Dioxane		0.553	0.531	0.496	0.476	0.492	0.492	8.9
2,3,4,6-Tetrachlorophenol		0.325	0.373	0.396	0.400	0.404	0.385	7.4
1-Methylnaphthalene		0.770	0.762	0.806	0.727	0.732	0.751	4.1

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG NO.: BG041624

EPA Sample No.: SSTDICCC040 Date Analyzed: Apr 16 2024 12:00AM

Lab File ID: BG060703.D Time Analyzed: 12:29

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	51170	7.961	236385	10.76	178236	14.59
UPPER LIMIT	102340	8.461	472770	11.258	356472	15.089
LOWER LIMIT	25585	7.461	118192.5	10.258	89118	14.089
EPA SAMPLE NO.						
01 PB160180TB	46318	7.94	181618	10.74	144480	14.58
02 PB160246BS	57254	7.95	244157	10.74	192896	14.58
03 PB160120BL	51189	7.95	199793	10.74	160907	14.57
04 PB160120BS	71474	7.95	308787	10.74	239703	14.58

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG NO.: BG041624

EPA Sample No.: SSTDCCC040 Date Analyzed: Apr 16 2024 12:00AM

Lab File ID: BG060909.D Time Analyzed: 12:29

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	87907	7.947	346268	10.75	266577	14.58
UPPER LIMIT	175814	8.447	692536	11.249	533154	15.08
LOWER LIMIT	43953.5	7.447	173134	10.249	133288.5	14.08
EPA SAMPLE NO.						
01 PB160180TB	46318	7.94	181618	10.74	144480	14.58
02 PB160246BS	57254	7.95	244157	10.74	192896	14.58
03 PB160120BL	51189	7.95	199793	10.74	160907	14.57
04 PB160120BS	71474	7.95	308787	10.74	239703	14.58

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG NO.: BG041624

EPA Sample No.: SSTDICCC040 Date Analyzed: Apr 16 2024 12:00AM

Lab File ID: BG060703.D Time Analyzed: 12:29

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	416987	17.339	381473	21.604	450367	24.748
UPPER LIMIT	833974	17.839	762946	22.104	900734	25.248
LOWER LIMIT	208493.5	16.839	190736.5	21.104	225183.5	24.248
EPA SAMPLE NO.						
01 PB160180TB	382278	17.33	361165	21.59	417510	24.72
02 PB160246BS	462252	17.32	415335	21.59	451610	24.72
03 PB160120BL	439391	17.32	424861	21.58	484252	24.72
04 PB160120BS	563421	17.33	513117	21.59	545770	24.72

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: BG041624 SAS No.: BG041624 SDG NO.: BG041624

EPA Sample No.: SSTDCCC040 Date Analyzed: Apr 16 2024 12:00AM

Lab File ID: BG060909.D Time Analyzed: 12:29

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	624499	17.33	575816	21.595	674847	24.727
UPPER LIMIT	1248998	17.83	1151632	22.095	1349694	25.227
LOWER LIMIT	312249.5	16.83	287908	21.095	337423.5	24.227
EPA SAMPLE NO.						
01 PB160180TB	382278	17.33	361165	21.59	417510	24.72
02 PB160246BS	462252	17.32	415335	21.59	451610	24.72
03 PB160120BL	439391	17.32	424861	21.58	484252	24.72
04 PB160120BS	563421	17.33	513117	21.59	545770	24.72

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

DataFile: _____

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB160120BS	n-Nitrosodimethylamine	1700	1200	ug/Kg	71				57	103	
	Pyridine	1700	750	ug/Kg	44				28	98	
	Benzaldehyde	1700	260	ug/Kg	15				10	147	
	Aniline	1700	780	ug/Kg	46				38	100	
	Phenol	1700	1400	ug/Kg	82				62	112	
	bis(2-Chloroethyl)ether	1700	1200	ug/Kg	71				60	101	
	2-Chlorophenol	1700	1400	ug/Kg	82				65	112	
	1,2-Dichlorobenzene	1700	1300	ug/Kg	76				51	102	
	1,3-Dichlorobenzene	1700	1300	ug/Kg	76				52	99	
	1,4-Dichlorobenzene	1700	1300	ug/Kg	76				52	100	
	Benzyl Alcohol	1700	1300	ug/Kg	76				42	116	
	2-Methylphenol	1700	1300	ug/Kg	76				61	108	
	2,2-oxybis(1-Chloropropane)	1700	1300	ug/Kg	76				51	100	
	Acetophenone	1700	1300	ug/Kg	76				55	104	
	3+4-Methylphenols	1700	1300	ug/Kg	76				58	111	
	N-Nitroso-di-n-propylamine	1700	1200	ug/Kg	71				63	95	
	Hexachloroethane	1700	1400	ug/Kg	82				72	108	
	Nitrobenzene	1700	1400	ug/Kg	82				57	101	
	Isophorone	1700	1300	ug/Kg	76				59	99	
	2-Nitrophenol	1700	1800	ug/Kg	106				61	111	
	2,4-Dimethylphenol	1700	1100	ug/Kg	65		*		67	119	
	bis(2-Chloroethoxy)methane	1700	1200	ug/Kg	71				53	105	
	2,4-Dichlorophenol	1700	1600	ug/Kg	94				62	107	
	1,2,4-Trichlorobenzene	1700	1500	ug/Kg	88				44	110	
	Benzoic acid	1700	1600	ug/Kg	94				10	140	
	Naphthalene	1700	1300	ug/Kg	76				62	100	
	4-Chloroaniline	1700	570	ug/Kg	34				16	100	
	Hexachlorobutadiene	1700	1600	ug/Kg	94				53	98	
	Caprolactam	1700	1300	ug/Kg	76				67	110	
	4-Chloro-3-methylphenol	1700	1400	ug/Kg	82				58	112	
	2-Methylnaphthalene	1700	1300	ug/Kg	76				60	104	
	Hexachlorocyclopentadiene	3300	3000	ug/Kg	91				45	134	
	2,4,6-Trichlorophenol	1700	1600	ug/Kg	94				59	102	
	2,4,5-Trichlorophenol	1700	1500	ug/Kg	88				61	98	
	1,1-Biphenyl	1700	1300	ug/Kg	76				57	103	
	2-Chloronaphthalene	1700	1300	ug/Kg	76				58	99	
	2-Nitroaniline	1700	1600	ug/Kg	94				66	101	
	Dimethylphthalate	1700	1300	ug/Kg	76				61	99	
	Acenaphthylene	1700	1300	ug/Kg	76				63	101	
	2,6-Dinitrotoluene	1700	1500	ug/Kg	88				61	104	
3-Nitroaniline	1700	1000	ug/Kg	59				28	100		
Acenaphthene	1700	1500	ug/Kg	88				57	104		
Total PAH		27200	22000	ug/Kg	81				57	104	
2,4-Dinitrophenol		3300	4800	ug/Kg	145		*		37	128	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **BG041624**

Client:

Analytical Method: **8270E**

DataFile:

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB160120BS	4-Nitrophenol	3300	3000	ug/Kg	91				48	119	
	Dibenzofuran	1700	1300	ug/Kg	76				63	99	
	2,4-Dinitrotoluene/2,6-Dinitrotol	3400	3100	ug/Kg	91				60	106	
	2,4-Dinitrotoluene	1700	1600	ug/Kg	94				60	106	
	Diethylphthalate	1700	1200	ug/Kg	71				60	101	
	4-Chlorophenyl-phenylether	1700	1300	ug/Kg	76				58	98	
	Fluorene	1700	1300	ug/Kg	76				61	101	
	4-Nitroaniline	1700	1500	ug/Kg	88				64	103	
	4,6-Dinitro-2-methylphenol	1700	1900	ug/Kg	112				76	113	
	N-Nitrosodiphenylamine	1700	1300	ug/Kg	76				56	107	
	Azobenzene	1700	1200	ug/Kg	71				54	103	
	4-Bromophenyl-phenylether	1700	1500	ug/Kg	88				55	99	
	Hexachlorobenzene	1700	1500	ug/Kg	88				64	98	
	Atrazine	1700	1400	ug/Kg	82				67	113	
	Pentachlorophenol	3300	3100	ug/Kg	94				67	105	
	Phenanthrene	1700	1300	ug/Kg	76				59	103	
	Anthracene	1700	1300	ug/Kg	76				61	105	
	Carbazole	1700	1300	ug/Kg	76				61	99	
	Di-n-butylphthalate	1700	1200	ug/Kg	71				58	104	
	Fluoranthene	1700	1300	ug/Kg	76				57	107	
	Benzidine	3300	680	ug/Kg	21				10	98	
	Pyrene	1700	1300	ug/Kg	76				59	103	
	Butylbenzylphthalate	1700	1300	ug/Kg	76				55	103	
	3,3-Dichlorobenzidine	1700	1100	ug/Kg	65				42	91	
	Benzo(a)anthracene	1700	1300	ug/Kg	76				60	102	
	Chrysene	1700	1300	ug/Kg	76				59	101	
	bis(2-Ethylhexyl)phthalate	1700	1200	ug/Kg	71				63	111	
	Di-n-octyl phthalate	1700	1300	ug/Kg	76				70	108	
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(k)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(a)pyrene	1700	1400	ug/Kg	82				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1500	ug/Kg	88				63	101	
	Dibenz(a,h)anthracene	1700	1500	ug/Kg	88				61	112	
	Benzo(g,h,i)perylene	1700	1400	ug/Kg	82				70	108	
	1,2,4,5-Tetrachlorobenzene	1700	1500	ug/Kg	88				53	101	
	1,4-Dioxane	1700	870	ug/Kg	51				50	96	
	2,3,4,6-Tetrachlorophenol	1700	1600	ug/Kg	94				59	108	
	1-Methylnaphthalene	1700	1200	ug/Kg	71				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

DataFile: _____

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB160246BS	n-Nitrosodimethylamine	1700	1300	ug/Kg	76				57	103	
	Pyridine	1700	770	ug/Kg	45				28	98	
	Benzaldehyde	1700	260	ug/Kg	15				10	147	
	Aniline	1700	800	ug/Kg	47				38	100	
	Phenol	1700	1400	ug/Kg	82				62	112	
	bis(2-Chloroethyl)ether	1700	1200	ug/Kg	71				60	101	
	2-Chlorophenol	1700	1400	ug/Kg	82				65	112	
	1,2-Dichlorobenzene	1700	1400	ug/Kg	82				51	102	
	1,3-Dichlorobenzene	1700	1300	ug/Kg	76				52	99	
	1,4-Dichlorobenzene	1700	1400	ug/Kg	82				52	100	
	Benzyl Alcohol	1700	1400	ug/Kg	82				42	116	
	2-Methylphenol	1700	1300	ug/Kg	76				61	108	
	2,2-oxybis(1-Chloropropane)	1700	1300	ug/Kg	76				51	100	
	Acetophenone	1700	1300	ug/Kg	76				55	104	
	3+4-Methylphenols	1700	1300	ug/Kg	76				58	111	
	N-Nitroso-di-n-propylamine	1700	1200	ug/Kg	71				63	95	
	Hexachloroethane	1700	1400	ug/Kg	82				72	108	
	Nitrobenzene	1700	1400	ug/Kg	82				57	101	
	Isophorone	1700	1200	ug/Kg	71				59	99	
	2-Nitrophenol	1700	1800	ug/Kg	106				61	111	
	2,4-Dimethylphenol	1700	1100	ug/Kg	65		*		67	119	
	bis(2-Chloroethoxy)methane	1700	1200	ug/Kg	71				53	105	
	2,4-Dichlorophenol	1700	1600	ug/Kg	94				62	107	
	1,2,4-Trichlorobenzene	1700	1500	ug/Kg	88				44	110	
	Benzoic acid	1700	1600	ug/Kg	94				10	140	
	Naphthalene	1700	1300	ug/Kg	76				62	100	
	4-Chloroaniline	1700	580	ug/Kg	34				16	100	
	Hexachlorobutadiene	1700	1600	ug/Kg	94				53	98	
	Caprolactam	1700	1300	ug/Kg	76				67	110	
	4-Chloro-3-methylphenol	1700	1400	ug/Kg	82				58	112	
	2-Methylnaphthalene	1700	1300	ug/Kg	76				60	104	
	Hexachlorocyclopentadiene	3300	3000	ug/Kg	91				45	134	
	2,4,6-Trichlorophenol	1700	1600	ug/Kg	94				59	102	
	2,4,5-Trichlorophenol	1700	1500	ug/Kg	88				61	98	
	1,1-Biphenyl	1700	1300	ug/Kg	76				57	103	
	2-Chloronaphthalene	1700	1300	ug/Kg	76				58	99	
	2-Nitroaniline	1700	1500	ug/Kg	88				66	101	
	Dimethylphthalate	1700	1300	ug/Kg	76				61	99	
	Acenaphthylene	1700	1300	ug/Kg	76				63	101	
	2,6-Dinitrotoluene	1700	1500	ug/Kg	88				61	104	
	3-Nitroaniline	1700	1000	ug/Kg	59				28	100	
	Acenaphthene	1700	1500	ug/Kg	88				57	104	
	Total PAH	27200	21800	ug/Kg	80				57	104	
	2,4-Dinitrophenol	3300	4800	ug/Kg	145		*		37	128	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

DataFile: _____

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB160246BS	4-Nitrophenol	3300	3100	ug/Kg	94				48	119		
	Dibenzofuran	1700	1300	ug/Kg	76				63	99		
	2,4-Dinitrotoluene	1700	1700	ug/Kg	100				60	106		
	2,4-Dinitrotoluene/2,6-Dinitro	3400	3200	ug/Kg	94				60	106		
	Diethylphthalate	1700	1200	ug/Kg	71				60	101		
	4-Chlorophenyl-phenylether	1700	1400	ug/Kg	82				58	98		
	Fluorene	1700	1300	ug/Kg	76				61	101		
	4-Nitroaniline	1700	1500	ug/Kg	88				64	103		
	4,6-Dinitro-2-methylphenol	1700	1900	ug/Kg	112				76	113		
	N-Nitrosodiphenylamine	1700	1300	ug/Kg	76				56	107		
	Azobenzene	1700	1200	ug/Kg	71				54	103		
	4-Bromophenyl-phenylether	1700	1500	ug/Kg	88				55	99		
	Hexachlorobenzene	1700	1500	ug/Kg	88				64	98		
	Atrazine	1700	1400	ug/Kg	82				67	113		
	Pentachlorophenol	3300	3100	ug/Kg	94				67	105		
	Phenanthrene	1700	1300	ug/Kg	76				59	103		
	Anthracene	1700	1300	ug/Kg	76				61	105		
	Carbazole	1700	1300	ug/Kg	76				61	99		
	Di-n-butylphthalate	1700	1200	ug/Kg	71				58	104		
	Fluoranthene	1700	1300	ug/Kg	76				57	107		
	Benzidine	3300	700	ug/Kg	21				10	98		
	Pyrene	1700	1300	ug/Kg	76				59	103		
	Butylbenzylphthalate	1700	1300	ug/Kg	76				55	103		
	3,3-Dichlorobenzidine	1700	1100	ug/Kg	65				42	91		
	Benzo(a)anthracene	1700	1300	ug/Kg	76				60	102		
	Chrysene	1700	1300	ug/Kg	76				59	101		
	bis(2-Ethylhexyl)phthalate	1700	1200	ug/Kg	71				63	111		
	Di-n-octyl phthalate	1700	1300	ug/Kg	76				70	108		
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109		
	Benzo(k)fluoranthene	1700	1400	ug/Kg	82				62	109		
	Benzo(a)pyrene	1700	1400	ug/Kg	82				63	103		
	Indeno(1,2,3-cd)pyrene	1700	1500	ug/Kg	88				63	101		
	Dibenz(a,h)anthracene	1700	1400	ug/Kg	82				61	112		
	Benzo(g,h,i)perylene	1700	1400	ug/Kg	82				70	108		
	1,2,4,5-Tetrachlorobenzene	1700	1500	ug/Kg	88				53	101		
	1,4-Dioxane	1700	930	ug/Kg	55				50	96		
	2,3,4,6-Tetrachlorophenol	1700	1600	ug/Kg	94				59	108		
	1-Methylnaphthalene	1700	1200	ug/Kg	71				70	130		

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB160120BL

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BG041624

SAS No.: BG041624 SDG NO.: BG041624

Lab File ID: BG060912.D

Lab Sample ID: PB160120BL

Instrument ID: BNA_G

Date Extracted: 04/10/2024

Matrix: (soil/water) Solid

Date Analyzed: 04/16/2024

Level: (low/med) LOW

Time Analyzed: 13:51

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB160120BS	PB160120BS	BG060913.D	04/16/2024

COMMENTS: _____

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB160180TB

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BG041624

SAS No.: BG041624 SDG NO.: BG041624

Lab File ID: BG060910.D

Lab Sample ID: PB160180TB

Instrument ID: BNA_G

Date Extracted: 04/12/2024

Matrix: (soil/water) water

Date Analyzed: 04/16/2024

Level: (low/med) LOW

Time Analyzed: 12:29

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB160180TB	PB160180TB	BG060910.D	04/16/2024

COMMENTS: _____

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB160246BS

Lab Name: CHEMTECH

Contract: _____

Lab Code: CHEM Case No.: BG041624

SAS No.: BG041624 SDG NO.: BG041624

Lab File ID: BG060911.D

Lab Sample ID: PB160246BS

Instrument ID: BNA_G

Date Extracted: 04/15/2024

Matrix: (soil/water) Solid

Date Analyzed: 04/16/2024

Level: (low/med) LOW

Time Analyzed: 13:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB160246BS	PB160246BS	BG060911.D	04/16/2024

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
PB160120BL	PB160120BL	BG060912.D	2-Fluorophenol	150	118.10	79		18	112
			Phenol-d6	150	107.07	71		15	107
			Nitrobenzene-d5	100	95.63	96		18	107
			2-Fluorobiphenyl	100	75.07	75		20	109
			2,4,6-Tribromophenol	150	168.58	112	*	10	110
			Terphenyl-d14	100	76.69	77		14	112
PB160120BS	PB160120BS	BG060913.D	2-Fluorophenol	150	128.40	86		18	112
			Phenol-d6	150	117.34	78		15	107
			Nitrobenzene-d5	100	90.22	90		18	107
			2-Fluorobiphenyl	100	76.51	77		20	109
			2,4,6-Tribromophenol	150	158.90	106		10	110
			Terphenyl-d14	100	76.38	76		14	112



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
PB160180TB	PB160180TB	BG060910.D	2-Fluorophenol	150	125.86	84		10	139
			Phenol-d6	150	109.59	73		10	134
			Nitrobenzene-d5	100	94.04	94		49	133
			2-Fluorobiphenyl	100	77.47	77		52	132
			2,4,6-Tribromophenol	150	178.59	119		32	145
			Terphenyl-d14	100	80.46	80		36	145



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SW-846

SDG No.: BG041624

Client: _____

Analytical Method: 8270E

Lab Sample ID	Client ID	Datafile	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
								Low	High
PB160246BS	PB160246BS	BG060911.D	2-Fluorophenol	150	122.01	81		18	112
			Phenol-d6	150	116.39	78		15	107
			Nitrobenzene-d5	100	91.52	92		18	107
			2-Fluorobiphenyl	100	76.16	76		20	109
			2,4,6-Tribromophenol	150	164.00	109		10	110
			Terphenyl-d14	100	78.98	79		14	112

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: BG041624

Lab Code: CHEM

SAS No.: BG041624

SDG NO.: BG041624

Lab File ID: BG060908.D

DFTPP Injection Date: 04/16/2024

Instrument ID: BNA_G

DFTPP Injection Time: 10:50

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36.4
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	27.7
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	35.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.1
275	10.0 - 60.0% of mass 198	29.7
365	Greater than 1% of mass 198	4.7
441	Present, but less than mass 443	16
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG060909.D	04/16/2024	11:36
PB160180TB	PB160180TB	BG060910.D	04/16/2024	12:29
PB160246BS	PB160246BS	BG060911.D	04/16/2024	13:10
PB160120BL	PB160120BL	BG060912.D	04/16/2024	13:51
PB160120BS	PB160120BS	BG060913.D	04/16/2024	14:32