

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041724\
 Data File : BG060924.D
 Acq On : 17 Apr 2024 18:10
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG041724

Quant Time: Apr 18 02:06:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	41.630	-4.1	105	0.00
3	Pyridine	40.000	39.728	0.7	96	0.00
4	n-Nitrosodimethylamine	40.000	37.556	6.1	91	0.00
5 S	2-Fluorophenol	80.000	77.275	3.4	95	0.00
6	Aniline	40.000	39.454	1.4	95	0.00
7 S	Phenol-d6	80.000	76.930	3.8	90	0.00
8	2-Chlorophenol	40.000	39.241	1.9	94	0.00
9	Benzaldehyde	40.000	37.432	6.4	92	0.00
10 C	Phenol	40.000	39.161	2.1	91	0.00
11	bis(2-Chloroethyl)ether	40.000	38.858	2.9	91	0.00
12	1,3-Dichlorobenzene	40.000	39.147	2.1	96	0.00
13 C	1,4-Dichlorobenzene	40.000	38.894	2.8	94	0.00
14	1,2-Dichlorobenzene	40.000	39.225	1.9	94	0.00
15	Benzyl Alcohol	40.000	39.286	1.8	92	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.505	3.7	93	0.00
17	2-Methylphenol	40.000	39.385	1.5	93	0.00
18	Hexachloroethane	40.000	38.346	4.1	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.740	3.1	91	0.00
20	3+4-Methylphenols	40.000	38.362	4.1	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	91	0.00
22	Acetophenone	40.000	38.621	3.4	93	0.00
23 S	Nitrobenzene-d5	80.000	77.138	3.6	91	0.00
24	Nitrobenzene	40.000	38.848	2.9	92	0.00
25	Isophorone	40.000	37.747	5.6	86	0.00
26 C	2-Nitrophenol	40.000	40.722	-1.8	93	0.00
27	2,4-Dimethylphenol	40.000	38.489	3.8	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.602	3.5	92	0.00
29 C	2,4-Dichlorophenol	40.000	37.527	6.2	89	0.00
30	1,2,4-Trichlorobenzene	40.000	37.312	6.7	89	0.00
31	Naphthalene	40.000	38.060	4.8	91	0.00
32	Benzoic acid	40.000	39.636	0.9	93	-0.01
33	4-Chloroaniline	40.000	38.950	2.6	89	0.00
34 C	Hexachlorobutadiene	40.000	37.970	5.1	92	0.00
35	Caprolactam	40.000	38.524	3.7	87	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.521	1.2	93	0.00
37	2-Methylnaphthalene	40.000	37.805	5.5	88	0.00
38	1-Methylnaphthalene	40.000	38.437	3.9	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	92	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.094	7.3	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.889	2.8	90	0.00
42 S	2,4,6-Tribromophenol	80.000	77.343	3.3	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.524	6.2	89	0.00
44	2,4,5-Trichlorophenol	40.000	37.915	5.2	90	0.00
45 S	2-Fluorobiphenyl	80.000	73.989	7.5	89	0.00
46	1,1'-Biphenyl	40.000	37.226	6.9	89	0.00
47	2-Chloronaphthalene	40.000	37.474	6.3	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041724\
 Data File : BG060924.D
 Acq On : 17 Apr 2024 18:10
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG041724

Quant Time: Apr 18 02:06:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.406	1.5	92	0.00
49	Acenaphthylene	40.000	38.501	3.7	92	0.00
50	Dimethylphthalate	40.000	38.348	4.1	92	0.00
51	2,6-Dinitrotoluene	40.000	39.482	1.3	92	0.00
52 C	Acenaphthene	40.000	38.439	3.9	92	0.00
53	3-Nitroaniline	40.000	39.936	0.2	94	0.00
54 P	2,4-Dinitrophenol	40.000	38.853	2.9	95	0.00
55	Dibenzofuran	40.000	37.697	5.8	90	0.00
56 P	4-Nitrophenol	40.000	40.939	-2.3	93	0.00
57	2,4-Dinitrotoluene	40.000	40.728	-1.8	94	0.00
58	Fluorene	40.000	38.417	4.0	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.267	1.8	92	0.00
60	Diethylphthalate	40.000	38.732	3.2	91	0.00
61	4-Chlorophenyl-phenylether	40.000	38.015	5.0	90	0.00
62	4-Nitroaniline	40.000	38.677	3.3	94	0.00
63	Azobenzene	40.000	38.920	2.7	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.165	-0.4	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.778	3.1	94	0.00
67	4-Bromophenyl-phenylether	40.000	38.660	3.4	92	0.00
68	Hexachlorobenzene	40.000	39.203	2.0	93	0.00
69	Atrazine	40.000	40.143	-0.4	93	0.00
70 C	Pentachlorophenol	40.000	40.603	-1.5	91	0.00
71	Phenanthrene	40.000	38.513	3.7	93	0.00
72	Anthracene	40.000	38.784	3.0	93	0.00
73	Carbazole	40.000	38.635	3.4	93	0.00
74	Di-n-butylphthalate	40.000	39.096	2.3	94	0.00
75 C	Fluoranthene	40.000	38.444	3.9	94	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	37.528	6.2	83	0.00
78	Pyrene	40.000	39.010	2.5	95	0.00
79 S	Terphenyl-d14	80.000	77.404	3.2	94	0.00
80	Butylbenzylphthalate	40.000	39.485	1.3	97	0.00
81	Benzo(a)anthracene	40.000	38.925	2.7	96	0.00
82	3,3'-Dichlorobenzidine	40.000	37.768	5.6	92	0.00
83	Chrysene	40.000	39.280	1.8	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.237	1.9	97	0.00
85 c	Di-n-octyl phthalate	40.000	39.101	2.2	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.138	4.7	95	0.00
88	Benzo(b)fluoranthene	40.000	38.938	2.7	97	0.00
89	Benzo(k)fluoranthene	40.000	37.881	5.3	95	0.00
90 C	Benzo(a)pyrene	40.000	38.181	4.5	96	0.00
91	Dibenzo(a,h)anthracene	40.000	39.089	2.3	96	-0.01
92	Benzo(g,h,i)perylene	40.000	38.687	3.3	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041724\
Data File : BG060924.D
Acq On : 17 Apr 2024 18:10
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG041724

Quant Time: Apr 18 02:06:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 02:02:19 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0