




Data Path : Z:\HPCHEM1\BNA_G\DATA\BG041816\
Data File : BG021719.D
Acq On : 18 Apr 2016 1:13
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 18 08:03:40 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG041316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 15:48:13 2016
Response via : Initial Calibration



TIC: BG021719.D

(69) Pentachlorophenol (C)




17.204min (+0.005) 0.00ng

response 40784

Ion	Exp%	Act%
265.70	100	100
268.00	68.10	63.03
264.00	67.20	66.09
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG041816\
Data File : BG021719.D
Acq On : 18 Apr 2016 1:13
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 18 08:03:40 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG041316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 13 15:48:13 2016
Response via : Initial Calibration



TIC: BG021719.D

(76) Benzidine

19.766min (-0.001) 0.00ng

response 285404

Ion	Exp%	Act%
184.00	100	100
185.00	14.40	14.59
183.00	12.80	12.72
0.00	0.00	0.00