

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
 Data File : BG060929.D
 Acq On : 18 Apr 2024 9:34
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 10:17:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	95	0.00
2	1,4-Dioxane	40.000	38.878	2.8	103	0.00
3	Pyridine	40.000	38.197	4.5	97	0.00
4	n-Nitrosodimethylamine	40.000	36.180	9.6	93	0.00
5 S	2-Fluorophenol	80.000	77.331	3.3	101	0.00
6	Aniline	40.000	40.433	-1.1	102	0.00
7 S	Phenol-d6	80.000	82.010	-2.5	101	0.00
8	2-Chlorophenol	40.000	38.462	3.8	98	0.00
9	Benzaldehyde	40.000	39.133	2.2	102	0.00
10 C	Phenol	40.000	41.047	-2.6	101	0.00
11	bis(2-Chloroethyl)ether	40.000	38.973	2.6	97	0.00
12	1,3-Dichlorobenzene	40.000	37.832	5.4	98	0.00
13 C	1,4-Dichlorobenzene	40.000	38.258	4.4	98	0.00
14	1,2-Dichlorobenzene	40.000	38.866	2.8	98	0.00
15	Benzyl Alcohol	40.000	38.528	3.7	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.926	10.2	92	-0.01
17	2-Methylphenol	40.000	37.792	5.5	95	0.00
18	Hexachloroethane	40.000	37.712	5.7	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.175	7.1	93	0.00
20	3+4-Methylphenols	40.000	37.725	5.7	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	36.616	8.5	93	0.00
23 S	Nitrobenzene-d5	80.000	74.361	7.0	93	0.00
24	Nitrobenzene	40.000	37.752	5.6	94	0.00
25	Isophorone	40.000	37.236	6.9	90	0.00
26 C	2-Nitrophenol	40.000	39.069	2.3	94	0.00
27	2,4-Dimethylphenol	40.000	38.301	4.2	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.925	5.2	96	0.00
29 C	2,4-Dichlorophenol	40.000	38.125	4.7	96	0.00
30	1,2,4-Trichlorobenzene	40.000	37.728	5.7	95	0.00
31	Naphthalene	40.000	38.219	4.5	96	0.00
32	Benzoic acid	40.000	37.820	5.4	94	-0.01
33	4-Chloroaniline	40.000	38.699	3.3	94	0.00
34 C	Hexachlorobutadiene	40.000	38.253	4.4	98	0.00
35	Caprolactam	40.000	39.554	1.1	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.984	0.0	99	0.00
37	2-Methylnaphthalene	40.000	39.563	1.1	98	0.00
38	1-Methylnaphthalene	40.000	39.427	1.4	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.456	3.9	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.550	3.6	94	0.00
42 S	2,4,6-Tribromophenol	80.000	76.846	3.9	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.534	3.7	97	0.00
44	2,4,5-Trichlorophenol	40.000	39.384	1.5	99	0.00
45 S	2-Fluorobiphenyl	80.000	76.831	4.0	98	0.00
46	1,1'-Biphenyl	40.000	38.767	3.1	98	0.00
47	2-Chloronaphthalene	40.000	38.658	3.4	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
 Data File : BG060929.D
 Acq On : 18 Apr 2024 9:34
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 10:17:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.495	1.3	98	0.00
49	Acenaphthylene	40.000	38.892	2.8	98	0.00
50	Dimethylphthalate	40.000	38.792	3.0	98	0.00
51	2,6-Dinitrotoluene	40.000	39.935	0.2	98	0.00
52 C	Acenaphthene	40.000	38.606	3.5	98	0.00
53	3-Nitroaniline	40.000	38.920	2.7	96	0.00
54 P	2,4-Dinitrophenol	40.000	36.620	8.5	94	0.00
55	Dibenzofuran	40.000	38.090	4.8	96	0.00
56 P	4-Nitrophenol	40.000	40.533	-1.3	98	0.00
57	2,4-Dinitrotoluene	40.000	40.536	-1.3	99	0.00
58	Fluorene	40.000	38.511	3.7	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.742	3.1	96	0.00
60	Diethylphthalate	40.000	38.277	4.3	96	0.00
61	4-Chlorophenyl-phenylether	40.000	38.127	4.7	96	0.00
62	4-Nitroaniline	40.000	39.237	1.9	100	0.00
63	Azobenzene	40.000	38.173	4.6	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.243	4.4	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.419	4.0	98	0.00
67	4-Bromophenyl-phenylether	40.000	38.138	4.7	96	0.00
68	Hexachlorobenzene	40.000	37.683	5.8	94	0.00
69	Atrazine	40.000	39.027	2.4	96	0.00
70 C	Pentachlorophenol	40.000	40.374	-0.9	96	0.00
71	Phenanthrene	40.000	38.393	4.0	98	0.00
72	Anthracene	40.000	38.742	3.1	97	0.00
73	Carbazole	40.000	38.491	3.8	98	0.00
74	Di-n-butylphthalate	40.000	38.509	3.7	97	0.00
75 C	Fluoranthene	40.000	38.155	4.6	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	45.366	-13.4	106	0.00
78	Pyrene	40.000	38.543	3.6	98	0.00
79 S	Terphenyl-d14	80.000	76.349	4.6	98	0.00
80	Butylbenzylphthalate	40.000	38.278	4.3	98	0.00
81	Benzo(a)anthracene	40.000	38.107	4.7	98	0.00
82	3,3'-Dichlorobenzidine	40.000	38.087	4.8	97	0.00
83	Chrysene	40.000	38.358	4.1	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.788	5.5	98	0.00
85 c	Di-n-octyl phthalate	40.000	37.938	5.2	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.119	4.7	98	0.00
88	Benzo(b)fluoranthene	40.000	38.461	3.8	98	0.00
89	Benzo(k)fluoranthene	40.000	37.649	5.9	97	0.00
90 C	Benzo(a)pyrene	40.000	37.943	5.1	98	0.00
91	Dibenzo(a,h)anthracene	40.000	38.631	3.4	98	-0.01
92	Benzo(g,h,i)perylene	40.000	38.357	4.1	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
Data File : BG060929.D
Acq On : 18 Apr 2024 9:34
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 18 10:17:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 02:02:19 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0