

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
 Data File : BG060938.D
 Acq On : 18 Apr 2024 16:26
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 16:59:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	115	0.00
2	1,4-Dioxane	40.000	39.249	1.9	126	0.00
3	Pyridine	40.000	39.326	1.7	121	0.00
4	n-Nitrosodimethylamine	40.000	38.131	4.7	118	0.00
5 S	2-Fluorophenol	80.000	75.984	5.0	120	0.00
6	Aniline	40.000	40.396	-1.0	124	0.00
7 S	Phenol-d6	80.000	80.707	-0.9	121	0.00
8	2-Chlorophenol	40.000	39.552	1.1	122	0.00
9	Benzaldehyde	40.000	38.826	2.9	122	0.00
10 C	Phenol	40.000	40.830	-2.1	122	0.00
11	bis(2-Chloroethyl)ether	40.000	40.071	-0.2	121	0.00
12	1,3-Dichlorobenzene	40.000	39.290	1.8	123	0.00
13 C	1,4-Dichlorobenzene	40.000	38.906	2.7	120	0.00
14	1,2-Dichlorobenzene	40.000	39.576	1.1	121	0.00
15	Benzyl Alcohol	40.000	37.992	5.0	114	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.968	2.6	120	0.00
17	2-Methylphenol	40.000	39.036	2.4	119	0.00
18	Hexachloroethane	40.000	38.896	2.8	121	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.060	4.8	115	0.00
20	3+4-Methylphenols	40.000	39.011	2.5	116	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	38.852	2.9	119	0.00
23 S	Nitrobenzene-d5	80.000	75.837	5.2	114	0.00
24	Nitrobenzene	40.000	37.912	5.2	114	0.00
25	Isophorone	40.000	37.286	6.8	109	0.00
26 C	2-Nitrophenol	40.000	40.633	-1.6	118	0.00
27	2,4-Dimethylphenol	40.000	37.827	5.4	114	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.676	3.3	118	0.00
29 C	2,4-Dichlorophenol	40.000	36.904	7.7	111	0.00
30	1,2,4-Trichlorobenzene	40.000	38.406	4.0	117	0.00
31	Naphthalene	40.000	38.384	4.0	116	0.00
32	Benzoic acid	40.000	36.454	8.9	109	0.00
33	4-Chloroaniline	40.000	37.473	6.3	109	0.00
34 C	Hexachlorobutadiene	40.000	39.384	1.5	121	0.00
35	Caprolactam	40.000	35.178	12.1	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.360	9.1	108	0.00
37	2-Methylnaphthalene	40.000	37.732	5.7	112	0.00
38	1-Methylnaphthalene	40.000	37.806	5.5	112	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.639	0.9	113	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.734	-6.8	117	0.00
42 S	2,4,6-Tribromophenol	80.000	71.777	10.3	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.414	1.5	111	0.00
44	2,4,5-Trichlorophenol	40.000	39.004	2.5	110	0.00
45 S	2-Fluorobiphenyl	80.000	79.285	0.9	114	0.00
46	1,1'-Biphenyl	40.000	39.351	1.6	112	0.00
47	2-Chloronaphthalene	40.000	39.865	0.3	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
 Data File : BG060938.D
 Acq On : 18 Apr 2024 16:26
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 18 16:59:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 02:02:19 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.926	2.7	108	0.00
49	Acenaphthylene	40.000	39.297	1.8	111	0.00
50	Dimethylphthalate	40.000	37.693	5.8	107	0.00
51	2,6-Dinitrotoluene	40.000	38.953	2.6	107	0.00
52 C	Acenaphthene	40.000	38.970	2.6	111	0.00
53	3-Nitroaniline	40.000	37.647	5.9	105	0.00
54 P	2,4-Dinitrophenol	40.000	37.248	6.9	107	0.00
55	Dibenzofuran	40.000	37.896	5.3	107	0.00
56 P	4-Nitrophenol	40.000	37.750	5.6	102	0.00
57	2,4-Dinitrotoluene	40.000	38.586	3.5	106	0.00
58	Fluorene	40.000	36.726	8.2	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.254	6.9	103	0.00
60	Diethylphthalate	40.000	36.384	9.0	102	0.00
61	4-Chlorophenyl-phenylether	40.000	36.937	7.7	104	0.00
62	4-Nitroaniline	40.000	35.469	11.3	102	0.00
63	Azobenzene	40.000	36.781	8.0	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.447	-1.1	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.629	0.9	104	0.00
67	4-Bromophenyl-phenylether	40.000	39.604	1.0	103	0.00
68	Hexachlorobenzene	40.000	38.911	2.7	101	0.00
69	Atrazine	40.000	39.023	2.4	99	0.00
70 C	Pentachlorophenol	40.000	41.072	-2.7	100	0.00
71	Phenanthrene	40.000	38.674	3.3	101	0.00
72	Anthracene	40.000	38.745	3.1	100	0.00
73	Carbazole	40.000	37.723	5.7	99	0.00
74	Di-n-butylphthalate	40.000	38.124	4.7	99	0.00
75 C	Fluoranthene	40.000	37.447	6.4	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	45.567	-13.9	108	0.00
78	Pyrene	40.000	38.594	3.5	100	0.00
79 S	Terphenyl-d14	80.000	76.373	4.5	99	0.00
80	Butylbenzylphthalate	40.000	38.927	2.7	101	0.00
81	Benzo(a)anthracene	40.000	38.775	3.1	101	0.00
82	3,3'-Dichlorobenzidine	40.000	37.953	5.1	98	0.00
83	Chrysene	40.000	38.386	4.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.852	2.9	102	0.00
85 c	Di-n-octyl phthalate	40.000	38.229	4.4	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.072	4.8	100	0.00
88	Benzo(b)fluoranthene	40.000	37.910	5.2	99	0.00
89	Benzo(k)fluoranthene	40.000	38.745	3.1	103	0.00
90 C	Benzo(a)pyrene	40.000	38.133	4.7	101	0.00
91	Dibenzo(a,h)anthracene	40.000	38.680	3.3	100	0.00
92	Benzo(g,h,i)perylene	40.000	38.762	3.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG041824\
Data File : BG060938.D
Acq On : 18 Apr 2024 16:26
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 18 16:59:49 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG041724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 18 02:02:19 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0