


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG042015\
 Data File : BG016565.D
 Acq On : 21 Apr 2015 1:13
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG040615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Apr 21 01:34:55 2015



AutoFind: Scans 2516, 2517, 2518; Background Corrected with Scan 2510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.2	21474	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	44.4	24906	PASS
70	69	0.00	2	0.7	180	PASS
127	198	10	80	59.9	33648	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	56154	PASS
199	198	5	9	6.8	3799	PASS
275	198	10	60	21.0	11769	PASS
365	198	1	100	1.9	1083	PASS
441	442	0.01	24	20.1	6804	PASS
442	198	50	100	60.4	33928	PASS
443	442	15	24	20.1	6819	PASS