

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042015\
 Data File : BG016576.D
 Acq On : 21 Apr 2015 8:11
 Operator : TP/IZ
 Sample : G1932-04
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-31

Manual Integrations
 APPROVED

apatel
 4/21/2015 5:06:04 PM

Quant Time: Apr 21 15:39:56 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG040615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 14 14:00:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	82078	20.00	ng	-0.02
21) Naphthalene-d8	10.44	136	377767	20.00	ng	-0.03
38) Acenaphthene-d10	14.29	164	224009	20.00	ng	-0.03
63) Phenanthrene-d10	17.04	188	435052	20.00	ng	-0.03
75) Chrysene-d12	21.21	240	368162	20.00	ng	-0.03
86) Perylene-d12	23.43	264	352375	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	242916	46.71	ng	-0.02
7) Phenol-d6	6.84	99	238507	30.55	ng	-0.02
23) Nitrobenzene-d5	8.81	82	423084	64.88	ng	-0.03
41) 2,4,6-Tribromophenol	15.79	330	161042	106.30	ng	-0.02
44) 2-Fluorobiphenyl	12.91	172	960710	73.02	ng	-0.03
78) Terphenyl-d14	19.66	244	937332	59.59	ng	-0.03
Target Compounds						
31) Naphthalene	10.48	128	59101	3.00	ng	Qvalue # 95
32) Benzoic acid	9.95	122	831322m	180.85	ng	
37) 2-Methylnaphthalene	12.10	142	84547	5.97	ng	99
83) Bis(2-ethylhexyl)phthalate	21.12	149	45494	2.67	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042015\
Data File : BG016576.D
Acq On : 21 Apr 2015 8:11
Operator : TP/IZ
Sample : G1932-04
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-31

Manual Integrations
APPROVED

apatel
4/21/2015 5:06:04 PM

Quant Time: Apr 21 15:39:56 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG040615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 14 14:00:56 2015
Response via : Initial Calibration

