


Data Path : Z:\HPCHEM1\BNA G\DATA\BG042016\  
 Data File : BG021761.D  
 Acq On : 19 Apr 2016 18:57  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG041316.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Apr 13 15:48:13 2016



AutoFind: Scans 2573, 2574, 2575; Background Corrected with Scan 2567

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 37.5      | 9669    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 35.3      | 9106    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 45.9      | 11820   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 25768   | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 1766    | PASS             |
| 275         | 198          | 10           | 60           | 27.2      | 7015    | PASS             |
| 365         | 198          | 1            | 100          | 3.6       | 937     | PASS             |
| 441         | 442          | 0.01         | 24           | 15.9      | 2780    | PASS             |
| 442         | 198          | 50           | 100          | 67.9      | 17506   | PASS             |
| 443         | 442          | 15           | 24           | 19.7      | 3447    | PASS             |