

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042024\
 Data File : BG060979.D
 Acq On : 20 Apr 2024 23:06
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042024

Quant Time: Apr 22 04:24:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 22 04:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	0.00
2	1,4-Dioxane	40.000	38.167	4.6	97	0.00
3	Pyridine	40.000	38.372	4.1	100	0.00
4	n-Nitrosodimethylamine	40.000	37.876	5.3	96	0.00
5 S	2-Fluorophenol	80.000	76.648	4.2	100	0.00
6	Aniline	40.000	38.324	4.2	98	0.00
7 S	Phenol-d6	80.000	77.904	2.6	98	0.00
8	2-Chlorophenol	40.000	37.971	5.1	100	0.00
9	Benzaldehyde	40.000	39.283	1.8	106	0.00
10 C	Phenol	40.000	38.788	3.0	98	0.00
11	bis(2-Chloroethyl)ether	40.000	38.523	3.7	100	0.00
12	1,3-Dichlorobenzene	40.000	38.032	4.9	99	0.00
13 C	1,4-Dichlorobenzene	40.000	37.420	6.4	97	0.00
14	1,2-Dichlorobenzene	40.000	37.835	5.4	98	0.00
15	Benzyl Alcohol	40.000	39.063	2.3	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.491	6.3	97	0.00
17	2-Methylphenol	40.000	38.728	3.2	98	0.00
18	Hexachloroethane	40.000	36.382	9.0	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.286	1.8	103	0.00
20	3+4-Methylphenols	40.000	38.613	3.5	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	39.065	2.3	102	0.00
23 S	Nitrobenzene-d5	80.000	78.111	2.4	100	0.00
24	Nitrobenzene	40.000	38.421	3.9	98	0.00
25	Isophorone	40.000	38.769	3.1	100	0.00
26 C	2-Nitrophenol	40.000	37.844	5.4	96	0.00
27	2,4-Dimethylphenol	40.000	38.356	4.1	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.503	3.7	102	0.00
29 C	2,4-Dichlorophenol	40.000	38.520	3.7	100	0.00
30	1,2,4-Trichlorobenzene	40.000	38.365	4.1	99	0.00
31	Naphthalene	40.000	38.135	4.7	100	0.00
32	Benzoic acid	40.000	42.403	-6.0	105	0.00
33	4-Chloroaniline	40.000	39.422	1.4	102	0.00
34 C	Hexachlorobutadiene	40.000	38.515	3.7	100	0.00
35	Caprolactam	40.000	40.113	-0.3	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.064	4.8	98	0.00
37	2-Methylnaphthalene	40.000	39.894	0.3	102	0.00
38	1-Methylnaphthalene	40.000	39.663	0.8	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.488	3.8	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.064	7.3	95	0.00
42 S	2,4,6-Tribromophenol	80.000	77.445	3.2	103	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.268	1.8	104	0.00
44	2,4,5-Trichlorophenol	40.000	39.279	1.8	105	0.00
45 S	2-Fluorobiphenyl	80.000	77.427	3.2	102	0.00
46	1,1'-Biphenyl	40.000	39.006	2.5	102	0.00
47	2-Chloronaphthalene	40.000	38.505	3.7	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042024\
 Data File : BG060979.D
 Acq On : 20 Apr 2024 23:06
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG042024

Quant Time: Apr 22 04:24:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 22 04:22:00 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.098	2.3	105	0.00
49	Acenaphthylene	40.000	38.432	3.9	100	0.00
50	Dimethylphthalate	40.000	38.162	4.6	101	0.00
51	2,6-Dinitrotoluene	40.000	38.085	4.8	100	0.00
52 C	Acenaphthene	40.000	38.266	4.3	98	0.00
53	3-Nitroaniline	40.000	37.618	6.0	103	0.00
54 P	2,4-Dinitrophenol	40.000	38.625	3.4	106	0.00
55	Dibenzofuran	40.000	38.323	4.2	101	0.00
56 P	4-Nitrophenol	40.000	38.625	3.4	99	0.00
57	2,4-Dinitrotoluene	40.000	39.544	1.1	103	0.00
58	Fluorene	40.000	38.201	4.5	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.442	6.4	99	0.00
60	Diethylphthalate	40.000	38.597	3.5	103	0.00
61	4-Chlorophenyl-phenylether	40.000	38.085	4.8	102	0.00
62	4-Nitroaniline	40.000	37.664	5.8	100	0.00
63	Azobenzene	40.000	38.311	4.2	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.358	4.1	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.595	3.5	102	0.00
67	4-Bromophenyl-phenylether	40.000	38.839	2.9	104	0.00
68	Hexachlorobenzene	40.000	38.133	4.7	100	0.00
69	Atrazine	40.000	39.427	1.4	101	0.00
70 C	Pentachlorophenol	40.000	39.588	1.0	100	0.00
71	Phenanthrene	40.000	38.278	4.3	102	0.00
72	Anthracene	40.000	37.960	5.1	100	0.00
73	Carbazole	40.000	37.611	6.0	100	0.00
74	Di-n-butylphthalate	40.000	38.135	4.7	101	0.00
75 C	Fluoranthene	40.000	37.811	5.5	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	40.398	-1.0	96	0.00
78	Pyrene	40.000	38.065	4.8	103	0.00
79 S	Terphenyl-d14	80.000	76.257	4.7	103	0.00
80	Butylbenzylphthalate	40.000	38.864	2.8	104	0.00
81	Benzo(a)anthracene	40.000	38.763	3.1	105	0.00
82	3,3'-Dichlorobenzidine	40.000	38.806	3.0	104	0.00
83	Chrysene	40.000	38.207	4.5	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.972	5.1	103	0.00
85 c	Di-n-octyl phthalate	40.000	38.516	3.7	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.594	3.5	105	0.00
88	Benzo(b)fluoranthene	40.000	38.059	4.9	104	0.00
89	Benzo(k)fluoranthene	40.000	37.541	6.1	103	0.00
90 C	Benzo(a)pyrene	40.000	38.154	4.6	105	0.00
91	Dibenzo(a,h)anthracene	40.000	38.847	2.9	103	0.00
92	Benzo(g,h,i)perylene	40.000	38.331	4.2	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042024\
Data File : BG060979.D
Acq On : 20 Apr 2024 23:06
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG042024

Quant Time: Apr 22 04:24:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 22 04:22:00 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0