

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042215\
 Data File : BG016634.D
 Acq On : 23 Apr 2015 1:28
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 23 14:04:06 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 01:45:23 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.561	0.563	-0.4	96	0.00
3	Pyridine	1.552	1.601	-3.2	98	0.00
4	n-Nitrosodimethylamine	0.465	0.484	-4.1	100	0.00
5 S	2-Fluorophenol	1.248	1.312	-5.1	102	0.00
6	Aniline	2.408	2.460	-2.2	97	0.00
7 S	Phenol-d6	1.743	1.836	-5.3	99	0.00
8	2-Chlorophenol	1.490	1.561	-4.8	102	0.00
9	Benzaldehyde	0.793	0.722	9.0	101	0.00
10 C	Phenol	1.844	1.927	-4.5	98	0.00
11	bis(2-Chloroethyl)ether	1.444	1.422	1.5	97	0.00
12	1,3-Dichlorobenzene	1.555	1.593	-2.4	102	0.00
13 C	1,4-Dichlorobenzene	1.583	1.629	-2.9	101	0.00
14	1,2-Dichlorobenzene	1.506	1.552	-3.1	102	0.00
15	Benzyl Alcohol	1.109	1.135	-2.3	94	0.00
16	2,2'-oxybis(1-Chloropropane	1.398	1.348	3.6	96	0.00
17	2-Methylphenol	1.243	1.289	-3.7	99	0.00
18	Hexachloroethane	0.569	0.580	-1.9	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.127	1.081	4.1	93	0.00
20	3+4-Methylphenols	1.712	1.766	-3.2	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.445	0.451	-1.3	95	0.00
23 S	Nitrobenzene-d5	0.325	0.335	-3.1	96	0.00
24	Nitrobenzene	0.338	0.348	-3.0	96	0.00
25	Isophorone	0.665	0.654	1.7	90	0.00
26 C	2-Nitrophenol	0.187	0.199	-6.4	95	0.00
27	2,4-Dimethylphenol	0.317	0.331	-4.4	95	0.00
28	bis(2-Chloroethoxy)methane	0.399	0.393	1.5	94	0.00
29 C	2,4-Dichlorophenol	0.262	0.284	-8.4	97	0.00
30	1,2,4-Trichlorobenzene	0.279	0.290	-3.9	98	0.00
31	Naphthalene	1.044	1.069	-2.4	97	0.00
32	Benzoic acid	0.103	0.128	-24.3	112	0.02
33	4-Chloroaniline	0.477	0.507	-6.3	95	0.00
34 C	Hexachlorobutadiene	0.124	0.126	-1.6	101	0.00
35	Caprolactam	0.152	0.172	-13.2	92	0.00
36 C	4-Chloro-3-methylphenol	0.317	0.345	-8.8	95	0.00
37	2-Methylnaphthalene	0.747	0.752	-0.7	94	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
39	1,2,4,5-Tetrachlorobenzene	0.460	0.478	-3.9	97	0.00
40 P	Hexachlorocyclopentadiene	0.143	0.126	11.9	81	0.00
41 S	2,4,6-Tribromophenol	0.136	0.166	-22.1	97	0.00
42 C	2,4,6-Trichlorophenol	0.326	0.363	-11.3	95	0.00
43	2,4,5-Trichlorophenol	0.347	0.396	-14.1	96	0.00
44 S	2-Fluorobiphenyl	1.208	1.242	-2.8	94	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042215\
 Data File : BG016634.D
 Acq On : 23 Apr 2015 1:28
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 23 14:04:06 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 01:45:23 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.540	1.571	-2.0	92	0.00
46	2-Chloronaphthalene	1.180	1.233	-4.5	94	0.00
47	2-Nitroaniline	0.344	0.393	-14.2	92	0.00
48	Acenaphthylene	2.094	2.200	-5.1	91	0.00
49	Dimethylphthalate	1.479	1.516	-2.5	87	0.00
50	2,6-Dinitrotoluene	0.363	0.393	-8.3	90	0.00
51 C	Acenaphthene	1.250	1.297	-3.8	92	0.00
52	3-Nitroaniline	0.449	0.533	-18.7	94	0.00
53 P	2,4-Dinitrophenol	0.133	0.104	21.8	60	0.00
54	Dibenzofuran	1.759	1.844	-4.8	90	0.00
55 P	4-Nitrophenol	0.340	0.403	-18.5	93	0.00
56	2,4-Dinitrotoluene	0.500	0.587	-17.4	91	0.00
57	Fluorene	1.550	1.664	-7.4	91	0.00
58	2,3,4,6-Tetrachlorophenol	0.262	0.313	-19.5	95	0.00
59	Diethylphthalate	1.576	1.642	-4.2	86	0.00
60	4-Chlorophenyl-phenylether	0.636	0.657	-3.3	90	0.00
61	4-Nitroaniline	0.514	0.648	-26.1#	96	0.00
62	Azobenzene	1.462	1.550	-6.0	88	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
64	4,6-Dinitro-2-methylphenol	0.129	0.126	2.3	83	0.00
65 c	n-Nitrosodiphenylamine	0.671	0.666	0.7	91	0.00
66	4-Bromophenyl-phenylether	0.157	0.156	0.6	92	0.00
67	Hexachlorobenzene	0.164	0.163	0.6	91	0.00
68	Atrazine	0.187	0.196	-4.8	91	0.00
69 C	Pentachlorophenol	0.073	0.084	-15.1	97	0.00
70	Phenanthrene	1.139	1.165	-2.3	93	0.00
71	Anthracene	1.131	1.181	-4.4	93	0.00
72	Carbazole	1.154	1.262	-9.4	96	0.00
73	Di-n-butylphthalate	1.405	1.492	-6.2	93	0.00
74 C	Fluoranthene	1.226	1.342	-9.5	98	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
76	Benzidine	0.740	0.737	0.4	97	0.00
77	Pyrene	1.343	1.390	-3.5	96	0.00
78 S	Terphenyl-d14	0.830	0.855	-3.0	99	0.00
79	Butylbenzylphthalate	0.703	0.747	-6.3	99	0.00
80	Benzo(a)anthracene	1.206	1.232	-2.2	99	0.00
81	3,3'-Dichlorobenzidine	0.399	0.409	-2.5	99	0.00
82	Chrysene	1.158	1.168	-0.9	98	0.00
83	Bis(2-ethylhexyl)phthalate	0.965	1.036	-7.4	99	0.00
84 c	Di-n-octyl phthalate	1.603	1.721	-7.4	99	0.00
85	Indeno(1,2,3-cd)pyrene	1.271	1.334	-5.0	101	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Benzo(b)fluoranthene	1.193	1.267	-6.2	104	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042215\
Data File : BG016634.D
Acq On : 23 Apr 2015 1:28
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 23 14:04:06 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 23 01:45:23 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.171	1.170	0.1	95	0.00
89 C	Benzo(a)pyrene	1.133	1.171	-3.4	100	0.00
90	Dibenzo(a,h)anthracene	1.101	1.158	-5.2	101	0.00
91	Benzo(a,h,i)perylene	1.111	1.150	-3.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0