

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042222\
 Data File : BG053255.D
 Acq On : 22 Apr 2022 11:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

 Reviewed By : Christian
 Giraldo

Quant Time: Apr 24 01:14:14 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Apr 24 01:13:30 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

Internal Standards

1) 1,4-Dichlorobenzene-d4	8.113	152	29898	20.000	ng	0.00
21) Naphthalene-d8	10.921	136	117323	20.000	ng	# 0.00
39) Acenaphthene-d10	14.734	164	84322	20.000	ng	0.00
64) Phenanthrene-d10	17.483	188	195413	20.000	ng	# 0.00
76) Chrysene-d12	21.766	240	184209	20.000	ng	# 0.00
86) Perylene-d12	25.061	264	195179	20.000	ng	0.00

04/25/2022

Supervised By : Jagrut
Opadhyay

System Monitoring Compounds

5) 2-Fluorophenol	5.676	112	148556	85.174	ng	0.00
7) Phenol-d6	7.273	99	226667	84.273	ng	0.00
23) Nitrobenzene-d5	9.277	82	237766	82.287	ng	0.00
42) 2,4,6-Tribromophenol	16.226	330	110388	82.285	ng	0.00
45) 2-Fluorobiphenyl	13.359	172	521174	85.227	ng	0.00
79) Terphenyl-d14	20.086	244	881364	80.807	ng	0.00

04/25/2022

Target Compounds

Compound	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.555	88	32134	38.579	ng	95
3) Pyridine	3.960	79	89725	40.843	ng	# 92
4) n-Nitrosodimethylamine	3.866	42	44138	40.573	ng	89
6) Aniline	7.432	93	128854	41.337	ng	97
8) 2-Chlorophenol	7.679	128	78980	41.945	ng	91
9) Benzaldehyde	7.244	77	64467m	40.098	ng	
10) Phenol	7.297	94	114083	42.611	ng	88
11) bis(2-Chloroethyl)ether	7.526	93	80298	41.607	ng	98
12) 1,3-Dichlorobenzene	8.002	146	93631m	42.792	ng	
13) 1,4-Dichlorobenzene	8.149	146	96100	43.081	ng	95
14) 1,2-Dichlorobenzene	8.472	146	88622	41.199	ng	99
15) Benzyl Alcohol	8.343	79	95144	39.797	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.636	45	131909	40.939	ng	99
17) 2-Methylphenol	8.554	107	73805	40.737	ng	93
18) Hexachloroethane	9.200	117	35800	42.798	ng	92
19) n-Nitroso-di-n-propyla...	8.912	70	79962	40.608	ng	# 95
20) 3+4-Methylphenols	8.883	107	106781	41.751	ng	94
22) Acetophenone	8.930	105	136928	42.069	ng	98
24) Nitrobenzene	9.318	77	113013	40.211	ng	93
25) Isophorone	9.841	82	195579	39.822	ng	99
26) 2-Nitrophenol	10.029	139	48279	41.080	ng	91
27) 2,4-Dimethylphenol	10.087	122	69317	41.183	ng	92
28) bis(2-Chloroethoxy)met...	10.316	93	109369	39.638	ng	97
29) 2,4-Dichlorophenol	10.575	162	87790	41.489	ng	98
30) 1,2,4-Trichlorobenzene	10.786	180	101315	42.305	ng	97
31) Naphthalene	10.974	128	261460	41.771	ng	99
32) Benzoic acid	10.234	122	44552m	41.225	ng	
33) 4-Chloroaniline	11.074	127	109169	40.727	ng	96
34) Hexachlorobutadiene	11.262	225	75060	42.193	ng	98
35) Caprolactam	11.850	113	27648	37.069	ng	# 82
36) 4-Chloro-3-methylphenol	12.196	107	100602	40.838	ng	95
37) 2-Methylnaphthalene	12.572	142	187803	40.548	ng	# 91
38) 1-Methylnaphthalene	12.790	142	183945	40.687	ng	# 95
40) 1,2,4,5-Tetrachloroben...	12.942	216	123029	42.959	ng	99
41) Hexachlorocyclopentadiene	12.919	237	55650	37.642	ng	91
43) 2,4,6-Trichlorophenol	13.177	196	83734	42.407	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042222\
 Data File : BG053255.D
 Acq On : 22 Apr 2022 11:06
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo

Quant Time: Apr 24 01:14:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	13.254	196	87462	41.567	ng	97	04/25/2022
46) 1,1'-Biphenyl	13.571	154	256260	42.201	ng	98	Supervised By :Jagrut Upadhyay
47) 2-Chloronaphthalene	13.618	162	201453	41.583	ng	98	
48) 2-Nitroaniline	13.812	65	74950	40.766	ng	91	
49) Acenaphthylene	14.458	152	322223	42.488	ng	99	
50) Dimethylphthalate	14.182	163	272606	40.420	ng	99	
51) 2,6-Dinitrotoluene	14.305	165	57205	40.657	ng	90	04/25/2022
52) Acenaphthene	14.799	154	211648m	41.153	ng		
53) 3-Nitroaniline	14.634	138	61591	41.398	ng	93	
54) 2,4-Dinitrophenol	14.846	184	30870	34.482	ng	#	79
55) Dibenzofuran	15.133	168	336776	41.820	ng		99
56) 4-Nitrophenol	14.945	139	46093	40.622	ng		85
57) 2,4-Dinitrotoluene	15.092	165	79631	39.753	ng	#	93
58) Fluorene	15.780	166	266470	40.969	ng		99
59) 2,3,4,6-Tetrachlorophenol	15.362	232	81405	39.839	ng		99
60) Diethylphthalate	15.539	149	273444	38.696	ng		99
61) 4-Chlorophenyl-phenyle...	15.768	204	149955	40.671	ng		98
62) 4-Nitroaniline	15.797	138	64968	41.383	ng	#	83
63) Azobenzene	16.062	77	285242	40.037	ng		98
65) 4,6-Dinitro-2-methylph...	15.862	198	53738	39.416	ng		95
66) n-Nitrosodiphenylamine	15.985	169	240076	42.722	ng		99
67) 4-Bromophenyl-phenylether	16.661	248	103082	41.168	ng		95
68) Hexachlorobenzene	16.790	284	111205	41.010	ng		93
69) Atrazine	16.925	200	84611	38.546	ng		98
70) Pentachlorophenol	17.137	266	60591	40.885	ng		89
71) Phenanthrene	17.524	178	439072	41.135	ng		97
72) Anthracene	17.612	178	431552	40.811	ng		99
73) Carbazole	17.883	167	425038	41.350	ng		99
74) Di-n-butylphthalate	18.429	149	461302	39.570	ng		98
75) Fluoranthene	19.527	202	532415	38.913	ng		99
77) Benzidine	19.704	184	229302	43.629	ng		99
78) Pyrene	19.892	202	532755	40.608	ng		99
80) Butylbenzylphthalate	20.767	149	195234	39.916	ng		97
81) Benzo(a)anthracene	21.748	228	513582	40.777	ng		99
82) 3,3'-Dichlorobenzidine	21.654	252	189333	40.875	ng		97
83) Chrysene	21.813	228	480626	41.118	ng		99
84) Bis(2-ethylhexyl)phtha...	21.636	149	270657	40.938	ng		100
85) Di-n-octyl phthalate	22.870	149	459132	41.524	ng	#	95
87) Indeno(1,2,3-cd)pyrene	28.850	276	583116	40.213	ng	#	97
88) Benzo(b)fluoranthene	24.021	252	491253	39.892	ng		99
89) Benzo(k)fluoranthene	24.086	252	488651	40.374	ng		96
90) Benzo(a)pyrene	24.914	252	428080	40.805	ng		96
91) Dibenzo(a,h)anthracene	28.903	278	485650	40.692	ng		96
92) Benzo(g,h,i)perylene	30.037	276	473282	40.233	ng		97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042222\
Data File : BG053255.D
Acq On : 22 Apr 2022 11:06
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Quant Time: Apr 24 01:14:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.113 min Scan# 80
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:152 Resp: 2989
Ion Ratio Lower Upper
152 100
150 159.2 114.6 172.0
115 60.4 45.9 68.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#2
1,4-Dioxane
Concen: 38.579 ng
RT: 3.555 min Scan# 28
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion: 88 Resp: 32134
Ion Ratio Lower Upper
88 100
58 75.7 58.2 87.2
43 34.3 23.4 35.0





#3

Pyridine

Concen: 40.843 ng

RT: 3.960 min Scan# 91

Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040

Tgt Ion: 79 Resp: 8972

Ion	Ratio	Lower	Upper
79	100		
52	68.3	50.9	76.3
51	37.0	24.4	36.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



04/25/2022
Supervised By :Jagrut
Upadhyay

04/25/2022



#4

n-Nitrosodimethylamine

Concen: 40.573 ng

RT: 3.866 min Scan# 81




Delta R.T. 0.000 min


Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Tgt Ion: 42 Resp: 44138

Ion	Ratio	Lower	Upper
42	100		
74	122.8	108.9	163.3
44	6.8	5.7	8.5





#5
2-Fluorophenol
Concen: 85.174 ng
RT: 5.676 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:112 Resp: 148550
Ion Ratio Lower Upper
112 100
64 70.6 48.0 72.0
63 39.3 26.5 39.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#6
Aniline
Concen: 41.337 ng
RT: 7.432 min Scan# 688
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion: 93 Resp: 128854
Ion Ratio Lower Upper
93 100
66 41.0 31.3 46.9
65 21.9 16.7 25.1





#7
Phenol-d6
Concen: 84.273 ng
RT: 7.273 min Scan# 60
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion: 99 Resp: 22666
Ion Ratio Lower Upper
99 100
42 20.1 14.7 22.1
71 43.0 29.2 43.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#8
2-Chlorophenol
Concen: 41.945 ng
RT: 7.679 min Scan# 730
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:128 Resp: 78980
Ion Ratio Lower Upper
128 100
130 34.3 13.4 53.4
64 60.3 30.9 70.9





#9
Benzaldehyde
Concen: 40.098 ng m
RT: 7.244 min Scan# 61
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion: 77 Resp: 6446
Ion Ratio Lower Upper
77 100
105 86.1 70.6 110.6
106 81.0 67.6 107.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#10
Phenol
Concen: 42.611 ng
RT: 7.297 min Scan# 665
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion: 94 Resp: 114083
Ion Ratio Lower Upper
94 100
65 33.9 11.1 51.1
66 50.4 20.0 60.0





#11
bis(2-Chloroethyl)ether
Concen: 41.607 ng
RT: 7.526 min Scan# 704
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 93 Resp: 80298
Ion Ratio Lower Upper
93 100
63 77.8 56.9 96.9
95 31.8 14.5 54.5



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#12
1,3-Dichlorobenzene
Concen: 42.792 ng m
RT: 8.002 min Scan# 785
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:146 Resp: 93631
Ion Ratio Lower Upper
146 100
148 64.2 50.6 76.0
75 40.6 25.9 38.9#





#13

1,4-Dichlorobenzene

Concen: 43.081 ng

RT: 8.149 min Scan# 81

Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 96100
 Ion Ratio Lower Upper
 146 100
 148 64.3 50.0 75.0
 111 43.6 31.0 46.4



Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo



04/25/2022
 Supervised By :Jagrut
 Upadhyay

04/25/2022



#14

1,2-Dichlorobenzene

Concen: 41.199 ng




RT: 8.472 min Scan# 865


Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Tgt Ion:146 Resp: 88622
 Ion Ratio Lower Upper
 146 100
 148 64.1 51.0 76.6
 111 42.8 34.7 52.1







#15
Benzyl Alcohol
Concen: 39.797 ng
RT: 8.343 min Scan# 846
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 95144
Ion Ratio Lower Upper
79 100
108 64.5 55.1 82.7
77 65.0 52.4 78.6



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 40.939 ng
RT: 8.636 min Scan# 893
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion: 45 Resp: 131909
Ion Ratio Lower Upper
45 100
77 15.3 0.0 34.9
79 10.9 0.0 30.7





#17
2-Methylphenol
Concen: 40.737 ng
RT: 8.554 min Scan# 811
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion	Ratio	Lower	Upper
107	100		
108	118.8	91.5	137.3
77	59.0	42.5	63.7
79	59.7	41.0	61.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#18
Hexachloroethane
Concen: 42.798 ng
RT: 9.200 min Scan# 989
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion	Ratio	Lower	Upper
117	100		
119	88.8	75.1	112.7
201	108.9	96.6	145.0





#19
n-Nitroso-di-n-propylamine
Concen: 40.608 ng
RT: 8.912 min Scan# 943
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion: 70 Resp: 79962
Ion Ratio Lower Upper
70 100
42 59.8 46.0 69.0
101 8.2 6.2 9.4
130 16.8 17.0 25.4#


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#20
3+4-Methylphenols
Concen: 41.751 ng
RT: 8.883 min Scan# 935
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:107 Resp: 106781
Ion Ratio Lower Upper
107 100
108 87.6 64.8 104.8
77 41.7 15.9 55.9
79 35.9 11.5 51.5






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.921 min Scan# 11732
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:136 Resp: 11732
Ion Ratio Lower Upper
136 100
137 11.9 9.4 14.2
54 10.5 6.4 9.6
68 5.7 5.0 7.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#22
Acetophenone
Concen: 42.069 ng
RT: 8.930 min Scan# 943
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:105 Resp: 136928
Ion Ratio Lower Upper
105 100
71 2.6 2.1 3.1
51 28.7 21.9 32.9
120 20.1 15.3 22.9





#23

Nitrobenzene-d5

Concen: 82.287 ng

RT: 9.277 min Scan# 1002

Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 237760
 Ion Ratio Lower Upper
 82 100
 128 37.7 31.3 46.9
 54 51.0 37.8 56.8



Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo



04/25/2022
 Supervised By :Jagrut
 Upadhyay

04/25/2022



#24

Nitrobenzene

Concen: 40.211 ng




RT: 9.318 min Scan# 1009


Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Tgt Ion: 77 Resp: 113013
 Ion Ratio Lower Upper
 77 100
 123 37.2 34.4 51.6
 65 14.1 11.2 16.8






#25
Isophorone
Concen: 39.822 ng
RT: 9.841 min Scan# 1100
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion: 82 Resp: 195579
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.6
138 16.2 12.6 18.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#26
2-Nitrophenol
Concen: 41.080 ng
RT: 10.029 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:139 Resp: 48279
Ion Ratio Lower Upper
139 100
109 40.8 29.4 44.0
65 68.3 48.2 72.2





#27
2,4-Dimethylphenol
Concen: 41.183 ng
RT: 10.087 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:122 Resp: 6931
Ion Ratio Lower Upper
122 100
107 132.5 96.7 145.1
121 55.5 46.3 69.5



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#28
bis(2-Chloroethoxy)methane
Concen: 39.638 ng
RT: 10.316 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion: 93 Resp: 109369
Ion Ratio Lower Upper
93 100
95 34.7 26.5 39.7
123 8.9 8.6 12.8






#29
2,4-Dichlorophenol
Concen: 41.489 ng
RT: 10.575 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:162 Resp: 87790
Ion Ratio Lower Upper
162 100
164 64.9 44.0 84.0
98 35.0 17.3 57.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#30
1,2,4-Trichlorobenzene
Concen: 42.305 ng
RT: 10.786 min Scan# 1259
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:180 Resp: 101315
Ion Ratio Lower Upper
180 100
182 93.7 77.8 116.6
145 29.0 24.4 36.6





#31
Naphthalene
Concen: 41.771 ng
RT: 10.974 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 261460
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 13.1 10.5 15.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#32
Benzoic acid
Concen: 41.225 ng m
RT: 10.234 min Scan# 1165
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:122 Resp: 44552
Ion Ratio Lower Upper
122 100
105 150.2 112.8 152.8
77 114.1 78.4 118.4






#33
4-Chloroaniline
Concen: 40.727 ng
RT: 11.074 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:127 Resp: 109169
Ion Ratio Lower Upper
127 100
129 32.6 27.6 41.4
65 37.7 28.2 42.4
92 23.2 17.0 25.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#34
Hexachlorobutadiene
Concen: 42.193 ng
RT: 11.262 min Scan# 1340
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:225 Resp: 75060
Ion Ratio Lower Upper
225 100
223 60.9 50.2 75.2
227 63.2 52.0 78.0





#35
Caprolactam
Concen: 37.069 ng
RT: 11.850 min Scan# 1441
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:113 Resp: 27648
Ion Ratio Lower Upper
113 100
55 208.2 151.3 191.3
56 158.1 128.5 168.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#36
4-Chloro-3-methylphenol
Concen: 40.838 ng
RT: 12.196 min Scan# 1499
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:107 Resp: 100602
Ion Ratio Lower Upper
107 100
144 23.6 21.4 32.0
142 72.8 61.7 92.5





#37
2-Methylnaphthalene
Concen: 40.548 ng
RT: 12.572 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 18780
Ion Ratio Lower Upper
142 100
141 95.3 71.3 106.9
115 44.8 29.5 44.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#38
1-Methylnaphthalene
Concen: 40.687 ng
RT: 12.790 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:142 Resp: 183945
Ion Ratio Lower Upper
142 100
141 93.3 70.5 105.7
116 4.7 3.1 4.7#






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.734 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G

ClientSampleId :



SSTDCCC040



Tgt Ion:164 Resp: 8432
Ion Ratio Lower Upper
164 100
162 103.1 80.0 120.0
160 44.3 38.8 58.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.959 ng
RT: 12.942 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:216 Resp: 123029
Ion Ratio Lower Upper
216 100
214 79.1 63.8 95.6
179 19.4 15.7 23.5
108 15.8 13.2 19.8





#41
Hexachlorocyclopentadiene
Concen: 37.642 ng
RT: 12.919 min Scan# 1622
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:237 Resp: 55650
Ion Ratio Lower Upper
237 100
235 65.4 37.7 77.7
272 12.3 0.0 30.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#42
2,4,6-Tribromophenol
Concen: 82.285 ng
RT: 16.226 min Scan# 2185
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:330 Resp: 110388
Ion Ratio Lower Upper
330 100
332 96.2 75.8 113.6
141 32.4 24.6 36.8





#43
2,4,6-Trichlorophenol
Concen: 42.407 ng
RT: 13.177 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:196 Resp: 83734
Ion Ratio Lower Upper
196 100
198 96.4 78.8 118.2
200 30.5 25.1 37.7



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#44
2,4,5-Trichlorophenol
Concen: 41.567 ng
RT: 13.254 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:196 Resp: 87462
Ion Ratio Lower Upper
196 100
198 96.7 75.8 113.6
97 49.6 36.5 54.7
132 21.9 18.1 27.1





#45
2-Fluorobiphenyl
Concen: 85.227 ng
RT: 13.359 min Scan# 1697
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:172 Resp: 521174
Ion Ratio Lower Upper
172 100
171 35.6 28.3 42.5
170 23.0 18.0 27.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#46
1,1'-Biphenyl
Concen: 42.201 ng
RT: 13.571 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:154 Resp: 256260
Ion Ratio Lower Upper
154 100
153 42.6 21.6 61.6
76 12.3 0.0 31.5





#47
2-Chloronaphthalene
Concen: 41.583 ng
RT: 13.618 min Scan# 1741
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:162 Resp: 20145
Ion Ratio Lower Upper
162 100
127 37.2 28.1 42.1
164 32.0 26.0 39.0



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#48
2-Nitroaniline
Concen: 40.766 ng
RT: 13.812 min Scan# 1774
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion: 65 Resp: 74950
Ion Ratio Lower Upper
65 100
92 62.2 52.1 78.1
138 88.0 80.1 120.1





#49
Acenaphthylene
Concen: 42.488 ng
RT: 14.458 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 32222
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.6
153 13.0 10.6 15.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#50
Dimethylphthalate
Concen: 40.420 ng
RT: 14.182 min Scan# 1837
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:163 Resp: 272606
Ion Ratio Lower Upper
163 100
194 3.7 3.3 4.9
164 10.4 8.5 12.7





#51
2,6-Dinitrotoluene
Concen: 40.657 ng
RT: 14.305 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 5720
Ion Ratio Lower Upper
165 100
63 62.4 43.7 65.5
89 62.7 45.0 67.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#52
Acenaphthene
Concen: 41.153 ng m
RT: 14.799 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:154 Resp: 211648
Ion Ratio Lower Upper
154 100
153 114.7 89.0 133.4
152 53.0 40.5 60.7





#53
3-Nitroaniline
Concen: 41.398 ng
RT: 14.634 min Scan# 1916
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 6159
Ion Ratio Lower Upper
138 100
108 10.4 8.0 12.0
92 128.6 96.2 144.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#54
2,4-Dinitrophenol
Concen: 34.482 ng
RT: 14.846 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:184 Resp: 30870
Ion Ratio Lower Upper
184 100
63 74.1 51.8 77.6
154 83.7 47.8 71.6#





#55
Dibenzofuran
Concen: 41.820 ng
RT: 15.133 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 336770
Ion Ratio Lower Upper
168 100
139 38.6 30.2 45.4
169 13.5 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#56
4-Nitrophenol
Concen: 40.622 ng
RT: 14.945 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:139 Resp: 46093
Ion Ratio Lower Upper
139 100
109 103.2 67.1 107.1
65 121.4 88.2 128.2





#57
2,4-Dinitrotoluene
Concen: 39.753 ng
RT: 15.092 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 7963
Ion Ratio Lower Upper
165 100
63 48.9 34.8 52.2
89 77.0 57.4 86.0
182 2.1 2.1 3.1#


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#58
Fluorene
Concen: 40.969 ng
RT: 15.780 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:166 Resp: 266470
Ion Ratio Lower Upper
166 100
165 96.9 78.6 117.8
167 13.5 10.9 16.3





#59
2,3,4,6-Tetrachlorophenol
Concen: 39.839 ng
RT: 15.362 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:232 Resp: 8140
Ion Ratio Lower Upper
232 100
131 37.8 30.1 45.1
130 2.6 1.8 2.8
166 26.2 22.2 33.2



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#60
Diethylphthalate
Concen: 38.696 ng
RT: 15.539 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:149 Resp: 273444
Ion Ratio Lower Upper
149 100
177 22.9 18.2 27.2
150 13.4 9.9 14.9





#61
4-Chlorophenyl-phenylether
Concen: 40.671 ng
RT: 15.768 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 14995
Ion Ratio Lower Upper
204 100
206 33.1 28.3 42.5
141 57.9 45.8 68.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#62
4-Nitroaniline
Concen: 41.383 ng
RT: 15.797 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:138 Resp: 64968
Ion Ratio Lower Upper
138 100
92 54.6 33.3 73.3
108 119.0 74.2 114.2#





#63
Azobenzene
Concen: 40.037 ng
RT: 16.062 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion: 77 Resp: 28524
Ion Ratio Lower Upper
77 100
182 24.1 5.6 45.6
105 18.5 0.0 38.8
51 29.2 8.4 48.4



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.483 min Scan# 2399
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:188 Resp: 195413
Ion Ratio Lower Upper
188 100
94 6.5 5.7 8.5
80 6.7 7.2 10.8







#65
4,6-Dinitro-2-methylphenol
Concen: 39.416 ng
RT: 15.862 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 53738
Ion Ratio Lower Upper
198 100
51 39.0 21.7 61.7
105 37.6 20.8 60.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#66
n-Nitrosodiphenylamine
Concen: 42.722 ng
RT: 15.985 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:169 Resp: 240076
Ion Ratio Lower Upper
169 100
168 67.2 54.5 81.7
167 38.2 29.6 44.4





#67
4-Bromophenyl-phenylether
Concen: 41.168 ng
RT: 16.661 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:248 Resp: 10308
Ion Ratio Lower Upper
248 100
250 100.5 77.9 116.9
141 67.9 49.2 73.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#68
Hexachlorobenzene
Concen: 41.010 ng
RT: 16.790 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:284 Resp: 111205
Ion Ratio Lower Upper
284 100
142 28.9 21.0 31.4
249 33.0 22.7 34.1





#69
 Atrazine
 Concen: 38.546 ng
 RT: 16.925 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06



Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040



Tgt Ion:200 Resp: 8461
 Ion Ratio Lower Upper
 200 100
 173 23.2 3.8 43.8
 215 48.3 26.3 66.3


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo






04/25/2022
 Supervised By :Jagrut
 Upadhyay


04/25/2022



#70
 Pentachlorophenol
 Concen: 40.885 ng
 RT: 17.137 min Scan# 2340
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06


Tgt Ion:266 Resp: 60591
 Ion Ratio Lower Upper
 266 100
 268 65.5 45.9 68.9
 264 64.5 45.3 67.9





#71
Phenanthrene
Concen: 41.135 ng
RT: 17.524 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 439072
Ion Ratio Lower Upper
178 100
176 18.9 15.9 23.9
179 15.3 13.4 20.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#72
Anthracene
Concen: 40.811 ng
RT: 17.612 min Scan# 2421
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:178 Resp: 431552
Ion Ratio Lower Upper
178 100
176 18.4 14.7 22.1
179 15.3 13.2 19.8





#73
 Carbazole
 Concen: 41.350 ng
 RT: 17.883 min Scan# 2469
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06



Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040



Tgt Ion:167 Resp: 425038
 Ion Ratio Lower Upper
 167 100
 166 22.6 18.5 27.7
 139 12.6 10.0 15.0


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




04/25/2022
 Supervised By :Jagrut
 Upadhyay



04/25/2022




#74
 Di-n-butylphthalate
 Concen: 39.570 ng
 RT: 18.429 min Scan# 2560
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06




Tgt Ion:149 Resp: 461302
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.0 10.6
 104 6.4 4.6 7.0





#75
Fluoranthene
Concen: 38.913 ng
RT: 19.527 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 532415
Ion Ratio Lower Upper
202 100
101 7.1 0.0 27.3
203 17.3 0.0 37.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.766 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06




Tgt Ion:240 Resp: 184209
Ion Ratio Lower Upper
240 100
120 5.2 5.4 8.0#
236 24.5 19.8 29.6





#77
Benzidine
Concen: 43.629 ng
RT: 19.704 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 229302
Ion Ratio Lower Upper
184 100
185 14.1 11.5 17.3
183 11.7 9.4 14.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#78
Pyrene
Concen: 40.608 ng
RT: 19.892 min Scan# 2809
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:202 Resp: 532755
Ion Ratio Lower Upper
202 100
200 20.6 17.0 25.6
203 18.2 15.0 22.6





#79
 Terphenyl-d14
 Concen: 80.807 ng
 RT: 20.086 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06

Instrument :

BNA_G



ClientSampleId :

SSTDCCC040

Tgt Ion:244 Resp: 881364
 Ion Ratio Lower Upper
 244 100
 212 7.7 5.4 8.2
 122 6.0 5.8 8.6



Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo






04/25/2022
 Supervised By :Jagrut
 Upadhyay


04/25/2022



#80
 Butylbenzylphthalate
 Concen: 39.916 ng
 RT: 20.767 min Scan# 2958
 Delta R.T. 0.000 min
 Lab File: BG053255.D
 Acq: 22 Apr 2022 11:06


Tgt Ion:149 Resp: 195234
 Ion Ratio Lower Upper
 149 100
 91 78.1 60.2 90.4
 206 22.9 19.4 29.2





#81
Benzo(a)anthracene
Concen: 40.777 ng
RT: 21.748 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 51358
Ion Ratio Lower Upper
228 100
226 26.5 21.4 32.0
229 20.4 16.7 25.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#82
3,3'-Dichlorobenzidine
Concen: 40.875 ng
RT: 21.654 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:252 Resp: 189333
Ion Ratio Lower Upper
252 100
254 66.4 51.2 76.8
126 7.2 5.8 8.8





#83
Chrysene
Concen: 41.118 ng
RT: 21.813 min Scan# 3139
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 480620
Ion Ratio Lower Upper
228 100
226 30.6 24.2 36.4
229 19.5 16.3 24.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.938 ng
RT: 21.636 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:149 Resp: 270657
Ion Ratio Lower Upper
149 100
167 27.0 21.4 32.2
279 4.6 3.8 5.8





#85
Di-n-octyl phthalate
Concen: 41.524 ng
RT: 22.870 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 45913
Ion Ratio Lower Upper
149 100
167 1.8 1.2 1.8
43 11.8 7.6 11.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/25/2022
Supervised By :Jagrut
Upadhyay



04/25/2022




#86
Perylene-d12
Concen: 20.000 ng
RT: 25.061 min Scan# 3689
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Tgt Ion:264 Resp: 195179
Ion Ratio Lower Upper
264 100
260 24.3 18.2 27.2
265 21.0 15.1 22.7





#87

Indeno(1,2,3-cd)pyrene

Concen: 40.213 ng

RT: 28.850 min Scan# 41

Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06

Instrument :

BNL_G

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 583110

Ion Ratio Lower Upper

276 100



138 7.4 8.7 13.1

277 25.5 20.2 30.4

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/25/2022
Supervised By :Jagrut
Upadhyay

04/25/2022



#88

Benzo(b)fluoranthene

Concen: 39.892 ng

RT: 24.021 min Scan# 3512

Delta R.T. 0.000 min

Lab File: BG053255.D

Acq: 22 Apr 2022 11:06




Tgt Ion:252 Resp: 491253


Ion Ratio Lower Upper

252 100

253 21.8 17.2 25.8


125 5.9 5.2 7.8





#89
Benzo(k)fluoranthene
Concen: 40.374 ng
RT: 24.086 min Scan# 31
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:252 Resp: 48865
Ion Ratio Lower Upper
252 100
253 24.0 17.4 26.0
125 5.2 5.0 7.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/25/2022
Supervised By :Jagrut
Upadhyay


04/25/2022



#90
Benzo(a)pyrene
Concen: 40.805 ng
RT: 24.914 min Scan# 3664
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06


Tgt Ion:252 Resp: 428080
Ion Ratio Lower Upper
252 100
253 20.6 18.4 27.6
125 6.6 5.6 8.4





#91
Dibenzo(a,h)anthracene
Concen: 40.692 ng
RT: 28.903 min Scan# 41
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06



Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:278 Resp: 485650
Ion Ratio Lower Upper
278 100
139 8.6 7.8 11.6
279 25.1 18.5 27.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



04/25/2022
Supervised By :Jagrut
Upadhyay

04/25/2022



#92
Benzo(g,h,i)perylene
Concen: 40.233 ng
RT: 30.037 min Scan# 4536
Delta R.T. 0.000 min
Lab File: BG053255.D
Acq: 22 Apr 2022 11:06

Tgt Ion:276 Resp: 473282
Ion Ratio Lower Upper
276 100
277 24.4 18.2 27.2
138 11.1 9.1 13.7

