

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042318\
 Data File : BG033924.D
 Acq On : 23 Apr 2018 17:11
 Operator : SJ/JU
 Sample : J2492-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 041918-JETB-E-U-M

Quant Time: Apr 24 04:01:37 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG041918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 19 13:56:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	66626	20.00	ng	0.00
21) Naphthalene-d8	10.61	136	291027	20.00	ng	-0.01
38) Acenaphthene-d10	14.46	164	207502	20.00	ng	-0.01
63) Phenanthrene-d10	17.21	188	557475	20.00	ng	-0.01
75) Chrysene-d12	21.47	240	579121	20.00	ng	-0.02
86) Perylene-d12	24.53	264	583956	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	188163	45.09	ng	0.00
7) Phenol-d6	6.99	99	156627	25.73	ng	0.00
23) Nitrobenzene-d5	8.98	82	514806	100.67	ng	0.00
41) 2,4,6-Tribromophenol	15.95	330	273679	113.05	ng	-0.01
44) 2-Fluorobiphenyl	13.09	172	1212581	85.85	ng	-0.01
78) Terphenyl-d14	19.85	244	2342945	90.89	ng	-0.01


Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042318\
Data File : BG033924.D
Acq On : 23 Apr 2018 17:11
Operator : SJ/JU
Sample : J2492-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
041918-JETB-E-U-M

Quant Time: Apr 24 04:01:37 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG041918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 19 13:56:22 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.83 min Scan# 773
Delta R.T. -0.01 min
Lab File: BG033924.D
Acq: 23 Apr 2018 17:11

Instrument :
BNA_G
ClientSampleId :
041918-JETB-E-U-M

Tot Ion:152 Resp: 66626
Ion Ratio Lower Upper
152 100
150 156.0 126.6 189.8
115 64.3 53.0 79.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 45.089 ng
RT: 5.44 min Scan# 366
Delta R.T. -0.01 min
Lab File: BG033924.D
Acq: 23 Apr 2018 17:11

Tgt Ion:112 Resp: 188163
Ion Ratio Lower Upper
112 100
64 66.9 56.3 84.5
63 35.8 30.1 45.1



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BG
Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 25.732 ng

RT: 6.99 min Scan# 631

Delta R.T. -0.01 min

Lab File: BG033924.D

Acq: 23 Apr 2018 17:11

Instrument :

BNA_G

ClientSampleId :

041918-JETB-E-U-M


Tot Ion: 99 Resp: 156627

Ion Ratio Lower Upper




99 100

42 20.1 14.1 21.1

71 35.8 26.1 39.1



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.61 min Scan# 1247

Delta R.T. -0.01 min

Lab File: BG033924.D

Acq: 23 Apr 2018 17:11

Tgt Ion: 136 Resp: 291027


Ion Ratio Lower Upper

136 100



137 12.3 9.2 13.8


54 8.5 10.3 15.5#

68 6.2 4.5 6.7



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 100.669 ng
 RT: 8.98 min Scan# 969
 Delta R.T. -0.01 min
 Lab File: BG033924.D
 Acq: 23 Apr 2018 17:11




Instrument :
 BNA_G
 ClientSampleId :
 041918-JETB-E-U-M


Tot Ion	Ratio	Lower	Upper
82	100		
128	42.2	29.7	44.5
54	52.3	43.1	64.7



#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.46 min Scan# 1902
 Delta R.T. -0.01 min
 Lab File: BG033924.D
 Acq: 23 Apr 2018 17:11

Tgt Ion	Ratio	Lower	Upper
164	100		
162	100.9	78.8	118.2
160	44.4	35.4	53.0









#41
 2,4,6-Tribromophenol
 Concen: 113.051 ng
 RT: 15.95 min Scan# 2156
 Delta R.T. -0.01 min
 Lab File: BG033924.D
 Acq: 23 Apr 2018 17:11

Instrument :
 BNA_G
 ClientSampleId :
 041918-JETB-E-U-M

Tot Ion	Ratio	Lower	Upper
330	100		
332	96.4	77.0	115.6
141	39.9	25.2	37.8#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 85.845 ng
 RT: 13.09 min Scan# 1668
 Delta R.T. -0.01 min
 Lab File: BG033924.D
 Acq: 23 Apr 2018 17:11

Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.8	30.0	45.0
170	25.3	19.5	29.3



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.21 min Scan# 2370
Delta R.T. -0.01 min
Lab File: BG033924.D
Acq: 23 Apr 2018 17:11

Instrument :
BNA_G
ClientSampleId :
041918-JETB-E-U-M

Tot Ion:188 Resp: 557475
Ion Ratio Lower Upper
188 100
94 7.7 6.9 10.3
80 9.2 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#75
Chrysene-d12
Concen: 20.000 ng
RT: 21.47 min Scan# 3095
Delta R.T. -0.02 min
Lab File: BG033924.D
Acq: 23 Apr 2018 17:11

Tgt Ion:240 Resp: 579121
Ion Ratio Lower Upper
240 100
120 9.0 6.8 10.2
236 25.8 20.9 31.3



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 90.891 ng

RT: 19.85 min Scan# 2819

Delta R.T. -0.01 min

Lab File: BG033924.D

Acq: 23 Apr 2018 17:11

Instrument :

BNA_G

ClientSampleId :

041918-JETB-E-U-M


Tot Ion:244 Resp: 2342945

Ion Ratio Lower Upper

244 100

212 9.3 5.8 8.8#




122 11.7 7.0 10.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.53 min Scan# 3616

Delta R.T. -0.03 min

Lab File: BG033924.D

Acq: 23 Apr 2018 17:11


Tgt Ion:264 Resp: 583956

Ion Ratio Lower Upper

264 100

260 23.7 17.4 26.0

265 21.8 17.7 26.5



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

