


Benzaldehyde

Response Ratio




$R = -9.25e-002 A^2 + 1.12e+000 A + 2.18e-001$
Coef of Det (r^2) = 0.997 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
Calibration Table Last Updated: Wed Apr 24 05:35:33 2019

2,4-Dinitrophenol


Response Ratio



Resp Ratio = $2.02 \times 10^{-1} \times \text{Amt} - 7.08 \times 10^{-2}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
Calibration Table Last Updated: Wed Apr 24 05:35:33 2019

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.16 \times 10^{-1} \times \text{Amt} - 4.11 \times 10^{-2}$
Coef of Det (r^2) = 0.995 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
Calibration Table Last Updated: Wed Apr 24 05:35:33 2019