

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG042419\
 Data File : BG040585.D
 Acq On : 24 Apr 2019 9:41
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 13:24:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 05:35:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.516	0.546	-5.8	98	0.00
3	Pyridine	1.496	1.624	-8.6	95	0.00
4	n-Nitrosodimethylamine	0.830	0.910	-9.6	102	0.00
5 S	2-Fluorophenol	1.135	1.204	-6.1	97	0.00
6	Aniline	2.315	2.463	-6.4	96	0.00
7 S	Phenol-d6	1.747	1.829	-4.7	95	0.00
8	2-Chlorophenol	1.385	1.432	-3.4	93	0.00
9	Benzaldehyde	1.083	1.164	-7.5	100	0.00
10 C	Phenol	1.878	1.975	-5.2	96	0.00
11	bis(2-Chloroethyl)ether	1.408	1.473	-4.6	95	0.00
12	1,3-Dichlorobenzene	1.512	1.525	-0.9	92	0.00
13 C	1,4-Dichlorobenzene	1.533	1.565	-2.1	93	0.00
14	1,2-Dichlorobenzene	1.478	1.501	-1.6	93	0.00
15	Benzyl Alcohol	1.818	1.943	-6.9	96	0.00
16	2,2'-oxybis(1-Chloropropane	2.804	2.948	-5.1	96	0.00
17	2-Methylphenol	1.329	1.363	-2.6	94	0.00
18	Hexachloroethane	0.629	0.648	-3.0	95	0.00
19 P	n-Nitroso-di-n-propylamine	1.573	1.645	-4.6	97	0.00
20	3+4-Methylphenols	1.851	1.926	-4.1	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.556	0.567	-2.0	95	0.00
23 S	Nitrobenzene-d5	0.433	0.450	-3.9	98	0.00
24	Nitrobenzene	0.462	0.480	-3.9	100	0.00
25	Isophorone	0.964	0.987	-2.4	97	0.00
26 C	2-Nitrophenol	0.166	0.172	-3.6	99	0.00
27	2,4-Dimethylphenol	0.311	0.317	-1.9	97	0.00
28	bis(2-Chloroethoxy)methane	0.484	0.493	-1.9	95	0.00
29 C	2,4-Dichlorophenol	0.353	0.350	0.8	92	0.00
30	1,2,4-Trichlorobenzene	0.392	0.388	1.0	93	0.00
31	Naphthalene	1.063	1.065	-0.2	93	0.00
32	Benzoic acid	0.213	0.232	-8.9	106	-0.01
33	4-Chloroaniline	0.471	0.479	-1.7	97	0.00
34 C	Hexachlorobutadiene	0.289	0.294	-1.7	96	0.00
35	Caprolactam	0.139	0.142	-2.2	95	0.00
36 C	4-Chloro-3-methylphenol	0.445	0.463	-4.0	101	0.00
37	2-Methylnaphthalene	0.794	0.807	-1.6	95	0.00
38	1-Methylnaphthalene	0.777	0.786	-1.2	95	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.745	0.710	4.7	94	0.00
41 P	Hexachlorocyclopentadiene	0.440	0.398	9.5	90	0.00
42 S	2,4,6-Tribromophenol	0.275	0.272	1.1	99	0.00
43 C	2,4,6-Trichlorophenol	0.450	0.439	2.4	99	0.00
44	2,4,5-Trichlorophenol	0.490	0.480	2.0	98	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG042419\
 Data File : BG040585.D
 Acq On : 24 Apr 2019 9:41
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 24 13:24:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 05:35:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.385	1.355	2.2	96	0.00
46	1,1'-Biphenyl	1.553	1.496	3.7	95	0.00
47	2-Chloronaphthalene	1.215	1.165	4.1	95	0.00
48	2-Nitroaniline	0.433	0.463	-6.9	109	0.00
49	Acenaphthylene	2.062	2.013	2.4	96	0.00
50	Dimethylphthalate	1.673	1.649	1.4	97	0.00
51	2,6-Dinitrotoluene	0.327	0.340	-4.0	104	0.00
52 C	Acenaphthene	1.201	1.172	2.4	99	0.00
53	3-Nitroaniline	0.335	0.356	-6.3	106	0.00
54 P	2,4-Dinitrophenol	0.161	0.164	-1.9	105	0.00
55	Dibenzofuran	1.967	1.938	1.5	99	0.00
56 P	4-Nitrophenol	0.290	0.303	-4.5	106	0.00
57	2,4-Dinitrotoluene	0.444	0.476	-7.2	108	0.00
58	Fluorene	1.590	1.578	0.8	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.494	0.494	0.0	99	0.00
60	Diethylphthalate	1.766	1.735	1.8	98	0.00
61	4-Chlorophenyl-phenylether	0.925	0.895	3.2	97	0.00
62	4-Nitroaniline	0.374	0.382	-2.1	102	0.00
63	Azobenzene	1.818	1.844	-1.4	100	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	0.092	0.097	-5.4	113	0.00
66 c	n-Nitrosodiphenylamine	0.588	0.571	2.9	98	0.00
67	4-Bromophenyl-phenylether	0.249	0.240	3.6	98	0.00
68	Hexachlorobenzene	0.258	0.248	3.9	98	0.00
69	Atrazine	0.233	0.229	1.7	96	0.00
70 C	Pentachlorophenol	0.151	0.157	-4.0	104	0.00
71	Phenanthrene	1.077	1.070	0.6	99	0.00
72	Anthracene	1.083	1.073	0.9	98	0.00
73	Carbazole	0.987	0.978	0.9	99	0.00
74	Di-n-butylphthalate	1.191	1.189	0.2	100	0.00
75 C	Fluoranthene	1.342	1.344	-0.1	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
77	Benzidine	0.548	0.682	-24.5	106	0.00
78	Pyrene	1.254	1.416	-12.9	100	0.00
79 S	Terphenyl-d14	0.903	1.030	-14.1	100	0.00
80	Butylbenzylphthalate	0.489	0.544	-11.2	100	0.00
81	Benzo(a)anthracene	1.264	1.388	-9.8	98	0.00
82	3,3'-Dichlorobenzidine	0.451	0.504	-11.8	99	0.00
83	Chrysene	1.202	1.331	-10.7	99	0.00
84	Bis(2-ethylhexyl)phthalate	0.705	0.780	-10.6	99	0.00
85 c	Di-n-octyl phthalate	1.188	1.320	-11.1	100	0.00
86	Indeno(1,2,3-cd)pyrene	1.453	1.577	-8.5	98	0.00
87 I	Perylene-d12	1.000	1.000	0.0	99	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG042419\
Data File : BG040585.D
Acq On : 24 Apr 2019 9:41
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 24 13:24:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG042319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 05:35:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.222	1.234	-1.0	100	0.00
89	Benzo(k)fluoranthene	1.141	1.135	0.5	99	0.00
90 C	Benzo(a)pyrene	1.157	1.165	-0.7	99	0.00
91	Dibenzo(a,h)anthracene	1.171	1.151	1.7	97	0.00
92	Benzo(g,h,i)perylene	1.165	1.154	0.9	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0