

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042522\
 Data File : BG053305.D
 Acq On : 25 Apr 2022 12:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 25 18:24:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	39.597	1.0	91	0.00
3	Pyridine	40.000	43.403	-8.5	94	-0.01
4	n-Nitrosodimethylamine	40.000	41.495	-3.7	91	0.00
5 S	2-Fluorophenol	80.000	88.683	-10.9	100	0.00
6	Aniline	40.000	45.493	-13.7	101	0.00
7 S	Phenol-d6	80.000	92.349	-15.4	102	0.00
8	2-Chlorophenol	40.000	44.173	-10.4	98	0.00
9	Benzaldehyde	40.000	41.281	-3.2	104	0.00
10 C	Phenol	40.000	45.570	-13.9	104	0.00
11	bis(2-Chloroethyl)ether	40.000	45.708	-14.3	104	0.00
12	1,3-Dichlorobenzene	40.000	43.442	-8.6	98	0.00
13 C	1,4-Dichlorobenzene	40.000	43.699	-9.2	99	0.00
14	1,2-Dichlorobenzene	40.000	44.347	-10.9	102	-0.01
15	Benzyl Alcohol	40.000	42.927	-7.3	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.326	-8.3	96	0.00
17	2-Methylphenol	40.000	45.120	-12.8	102	0.00
18	Hexachloroethane	40.000	42.591	-6.5	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.863	-12.2	100	0.00
20	3+4-Methylphenols	40.000	46.073	-15.2	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	42.907	-7.3	101	0.00
23 S	Nitrobenzene-d5	80.000	82.124	-2.7	97	0.00
24	Nitrobenzene	40.000	41.696	-4.2	101	0.00
25	Isophorone	40.000	42.921	-7.3	99	0.00
26 C	2-Nitrophenol	40.000	42.638	-6.6	99	0.00
27	2,4-Dimethylphenol	40.000	43.263	-8.2	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.392	-6.0	98	0.00
29 C	2,4-Dichlorophenol	40.000	43.577	-8.9	100	0.00
30	1,2,4-Trichlorobenzene	40.000	41.985	-5.0	98	0.00
31	Naphthalene	40.000	42.638	-6.6	100	0.00
32	Benzoic acid	40.000	40.241	-0.6	90	0.00
33	4-Chloroaniline	40.000	44.773	-11.9	103	0.00
34 C	Hexachlorobutadiene	40.000	41.189	-3.0	97	0.00
35	Caprolactam	40.000	45.092	-12.7	103	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.960	-9.9	101	0.00
37	2-Methylnaphthalene	40.000	43.098	-7.7	101	0.00
38	1-Methylnaphthalene	40.000	43.246	-8.1	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.916	2.7	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.695	13.3	85	0.00
42 S	2,4,6-Tribromophenol	80.000	81.657	-2.1	102	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.533	1.2	99	0.00
44	2,4,5-Trichlorophenol	40.000	39.055	2.4	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.591	0.5	102	0.00
46	1,1'-Biphenyl	40.000	40.213	-0.5	103	0.00
47	2-Chloronaphthalene	40.000	39.727	0.7	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042522\
 Data File : BG053305.D
 Acq On : 25 Apr 2022 12:25
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 25 18:24:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.283	-0.7	98	0.00
49	Acenaphthylene	40.000	41.218	-3.0	103	0.00
50	Dimethylphthalate	40.000	40.416	-1.0	102	0.00
51	2,6-Dinitrotoluene	40.000	40.006	-0.0	101	0.00
52 C	Acenaphthene	40.000	40.845	-2.1	102	0.00
53	3-Nitroaniline	40.000	41.174	-2.9	102	0.00
54 P	2,4-Dinitrophenol	40.000	39.804	0.5	97	0.00
55	Dibenzofuran	40.000	40.795	-2.0	105	0.00
56 P	4-Nitrophenol	40.000	40.051	-0.1	96	0.00
57	2,4-Dinitrotoluene	40.000	41.700	-4.3	103	0.00
58	Fluorene	40.000	40.804	-2.0	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.856	0.4	99	0.00
60	Diethylphthalate	40.000	40.302	-0.8	102	0.00
61	4-Chlorophenyl-phenylether	40.000	40.110	-0.3	102	0.00
62	4-Nitroaniline	40.000	41.761	-4.4	106	0.00
63	Azobenzene	40.000	40.592	-1.5	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.925	-2.3	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.881	-2.2	104	0.00
67	4-Bromophenyl-phenylether	40.000	41.151	-2.9	103	0.00
68	Hexachlorobenzene	40.000	40.591	-1.5	103	0.00
69	Atrazine	40.000	38.838	2.9	100	0.00
70 C	Pentachlorophenol	40.000	39.037	2.4	92	0.00
71	Phenanthrene	40.000	40.780	-2.0	104	0.00
72	Anthracene	40.000	40.519	-1.3	103	0.00
73	Carbazole	40.000	40.438	-1.1	103	0.00
74	Di-n-butylphthalate	40.000	40.024	-0.1	102	0.00
75 C	Fluoranthene	40.000	37.890	5.3	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	93	0.00
77	Benzydine	40.000	39.550	1.1	96	0.00
78	Pyrene	40.000	43.331	-8.3	98	0.00
79 S	Terphenyl-d14	80.000	85.217	-6.5	97	0.00
80	Butylbenzylphthalate	40.000	41.950	-4.9	93	0.00
81	Benzo(a)anthracene	40.000	40.789	-2.0	93	0.00
82	3,3'-Dichlorobenzidine	40.000	39.938	0.2	91	0.00
83	Chrysene	40.000	41.029	-2.6	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.748	-4.4	95	0.00
85 c	Di-n-octyl phthalate	40.000	41.054	-2.6	93	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.028	-2.6	93	0.00
88	Benzo(b)fluoranthene	40.000	40.682	-1.7	92	0.00
89	Benzo(k)fluoranthene	40.000	39.919	0.2	91	0.00
90 C	Benzo(a)pyrene	40.000	40.276	-0.7	91	0.00
91	Dibenzo(a,h)anthracene	40.000	41.182	-3.0	94	0.00
92	Benzo(g,h,i)perylene	40.000	40.716	-1.8	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042522\
Data File : BG053305.D
Acq On : 25 Apr 2022 12:25
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 25 18:24:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0