

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042524\
 Data File : BG061033.D
 Acq On : 25 Apr 2024 11:18
 Operator : MA/JU
 Sample : PB160476BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB160476BL

Quant Time: Apr 25 12:14:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 22 04:22:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.935	152	19696	20.000	ng	0.00
21) Naphthalene-d8	10.738	136	79321	20.000	ng	# 0.00
39) Acenaphthene-d10	14.568	164	67542	20.000	ng	0.00
64) Phenanthrene-d10	17.318	188	195480	20.000	ng	0.00
76) Chrysene-d12	21.578	240	187558	20.000	ng	0.00
86) Perylene-d12	24.709	264	213055	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.526	112	98835	101.502	ng	0.00
7) Phenol-d6	7.112	99	134128	92.839	ng	0.00
23) Nitrobenzene-d5	9.092	82	135172	76.564	ng	0.00
42) 2,4,6-Tribromophenol	16.061	330	142748	96.245	ng	0.00
45) 2-Fluorobiphenyl	13.193	172	421075	86.283	ng	0.00
79) Terphenyl-d14	19.944	244	1038521	85.948	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042524\
Data File : BG061033.D
Acq On : 25 Apr 2024 11:18
Operator : MA/JU
Sample : PB160476BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB160476BL

Quant Time: Apr 25 12:14:28 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 22 04:22:00 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.935 min Scan# 843
Delta R.T. -0.008 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18




Instrument :
BNA_G
ClientSampleId :
PB160476BL


Tgt Ion:152 Resp: 19696
Ion Ratio Lower Upper
152 100
150 138.1 118.6 178.0
115 59.8 47.4 71.0



#5
2-Fluorophenol
Concen: 101.502 ng
RT: 5.526 min Scan# 432
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18

Tgt Ion:112 Resp: 98835
Ion Ratio Lower Upper
112 100
64 69.3 53.7 80.5
63 42.6 37.0 55.6









#7
Phenol-d6
Concen: 92.839 ng
RT: 7.112 min Scan# 701
Delta R.T. 0.004 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18




Instrument :
BNA_G
ClientSampleId :
PB160476BL


Tgt Ion: 99 Resp: 134128
Ion Ratio Lower Upper
99 100
42 31.8 27.2 40.8
71 49.6 38.8 58.2



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.738 min Scan# 1319
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18

Tgt Ion: 136 Resp: 79321
Ion Ratio Lower Upper
136 100
137 10.2 8.5 12.7
54 10.7 9.4 14.0
68 3.1 3.8 5.6#









#23
Nitrobenzene-d5
Concen: 76.564 ng
RT: 9.092 min Scan# 1039
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18




Instrument :
BNA_G
ClientSampleId :
PB160476BL


Tgt Ion: 82 Resp: 135172
Ion Ratio Lower Upper
82 100
128 40.4 32.1 48.1
54 75.2 62.3 93.5



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.568 min Scan# 1971
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18


Tgt Ion: 164 Resp: 67542
Ion Ratio Lower Upper
164 100
162 98.0 81.2 121.8
160 43.3 36.2 54.2








#42
2,4,6-Tribromophenol
Concen: 96.245 ng
RT: 16.061 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18

Instrument :
BNA_G
ClientSampleId :
PB160476BL







Tgt Ion:330 Resp: 142748
Ion Ratio Lower Upper
330 100
332 96.5 77.1 115.7
141 24.9 20.6 31.0



#45
2-Fluorobiphenyl
Concen: 86.283 ng
RT: 13.193 min Scan# 1737
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18

Tgt Ion:172 Resp: 421075
Ion Ratio Lower Upper
172 100
171 36.1 29.3 43.9
170 23.8 20.1 30.1









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.318 min Scan# 2439
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18




Instrument :
BNA_G
ClientSampleId :
PB160476BL


Tgt Ion:188 Resp: 195480
Ion Ratio Lower Upper
188 100
94 4.2 4.1 6.1
80 5.7 4.0 6.0



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.578 min Scan# 3164
Delta R.T. -0.002 min
Lab File: BG061033.D
Acq: 25 Apr 2024 11:18

Tgt Ion:240 Resp: 187558
Ion Ratio Lower Upper
240 100
120 4.3 3.0 4.6
236 27.8 21.8 32.6









#79
 Terphenyl-d14
 Concen: 85.948 ng
 RT: 19.944 min Scan# 21
 Delta R.T. 0.004 min
 Lab File: BG061033.D
 Acq: 25 Apr 2024 11:18

Instrument :
 BNA_G
 ClientSampleId :
 PB160476BL

Tgt Ion:244 Resp: 1038521
 Ion Ratio Lower Upper
 244 100
 212 7.5 5.9 8.9
 122 4.8 4.2 6.2



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 24.709 min Scan# 3697
 Delta R.T. 0.010 min
 Lab File: BG061033.D
 Acq: 25 Apr 2024 11:18

Tgt Ion:264 Resp: 213055
 Ion Ratio Lower Upper
 264 100
 260 25.5 19.4 29.2
 265 21.2 17.0 25.4

